Keysight Technologies E4980A Precision LCR Meter 20 Hz to 2 MHz

E4980AL Precision LCR Meter 20 Hz to 300 kHz/500 kHz/1 MHz

Data Sheet





Table of Contents

Definitions	03
How to Use Tables	03
E4980A/E4980AL	03
Basic Specifications.	04
Measurement functions Test signal	
Measurement display ranges	
Absolute measurement accuracy	
Relative accuracy	
Basic accuracy	12
Short offset Zs	14
Open offset Yo	14
Calibration accuracy Acal	16
Measurement accuracy	17
Compensation function	18
DC bias signal	19
Measurement assistance functions	19
Options	2C
Frequency options	
Interface options	
Other options	
Power and DC bias enhancement specification	
Test signal	
DCR measurement specification	
General Specifications	27
Supplemental Information	31
Settling time	31
Measurement circuit protection	
Measurement time	
Display time	
Measurement data transfer time	
Maximum DC bias current	
Relative measurement accuracy with bias current isolation	
DC bias settling time	
Web Resources	38

Definitions

All specifications apply to the conditions of a 0 to 55°C temperature range, unless otherwise stated, and 30 minutes after the instrument has been turned on.

Specifications (spec.): Warranted performance. Specifications include guardbands to account for the expected statistical performance distribution, measurement uncertainties, and changes in performance due to environmental conditions.

Supplemental information is provided as information that is useful in operating the instrument, but is not covered by the product warranty. This information is classified as either typical or nominal.

Typical (typ.): Expected performance of an average unit without taking guardbands into account.

Nominal (nom.): A general descriptive term that does not imply a level of performance.

How to Use Tables

When measurement conditions fall under multiple categories in a table, apply the best value.

For example, basic accuracy Ab is 0.10% under the following conditions;

Measurement time mode SHORT
Test frequency 125 Hz
Test signal voltage 0.3 Vrms

E4980A/E4980AL

The E4980A is the model number of the 20 Hz to 2 MHz frequency range LCR meter. The E4980AL is the model number of the 20 Hz to 300 kHz, 500 kHz or 1 MHz frequency range LCR meter. See the E4980A/E4980AL Configuration Guide (5989-8321EN) for more details.

Frequency range	Model number and option
20 Hz to 2 MHz	E4980A
20 Hz to 1 MHz	E4980AL-102
20 Hz to 500 kHz	E4980AL-052
20 Hz to 300 kHz	E4980AL-032



Basic Specifications

Measurement functions

Measurement parameters

- Cp-D, Cp-Q, Cp-G, Cp-Rp
- Cs-D, Cs-Q, Cs-Rs
- Lp-D, Lp-Q, Lp-G, Lp-Rp, Lp-Rdc
- Ls-D, Ls-Q, Ls-Rs, Ls-Rdc
- R-X
- Z- θ d, Z- θ r
- G-B
- $Y-\theta d, Y-\theta r$
- Vdc-Idc¹

Definitions

- Cp Capacitance value measured with parallel-equivalent circuit model
- Cs Capacitance value measured with series-equivalent circuit model
- Lp Inductance value measured with parallel-equivalent circuit model
- Ls Inductance value measured with series-equivalent circuit model
- D Dissipation factor
- **Q** Quality factor (inverse of D)
- G Equivalent parallel conductance measured with parallel-equivalent circuit model
- **Rp** Equivalent parallel resistance measured with parallel-equivalent circuit model
- Rs Equivalent series resistance measured with series-equivalent circuit model
- Rdc Direct-current resistance
- R Resistance
- X Reactance
- Z Impedance
- Y Admittance
- θd Phase angle of impedance/admittance (degree)
- θr Phase angle of impedance/admittance (radian)
- B Susceptance
- Vdc Direct-current voltage
- Idc Direct-current electricity

Deviation measurement function: Deviation from reference value and percentage of deviation from reference value can be output as the result.

Equivalent circuits for measurement: Parallel, Series

Impedance range selection: Auto (auto range mode), manual (hold range mode)

Trigger mode: Internal trigger (INT), manual trigger (MAN), external trigger (EXT), GPIB trigger (BUS)



Table 1. Trigger delay time

Range	0 s - 999 s
Resolution	100 μs (0 s ≤ - ≤ 100 s)
	1 ms (100 s < - ≤ 999 s)

Table 2. Step delay time

Range	0 s - 999 s
Resolution	100 μs (0 s ≤ - ≤ 100 s)
	1 ms (100 s < - ≤ 999 s)

Measurement terminal: Four-terminal pair

Test cable length: 0 m, 1 m, 2 m, 4 m

Measurement time modes: Short mode, medium mode, long mode.

Table 3. Averaging

Range	1 - 256 measurements
Resolution	1

Test signal

Table 4. Test frequencies

Test frequencies	20 Hz - 2 MHz (E4980A) 20 Hz - 1 MHz (E4980AL-102) 20 Hz - 500 kHz (E4980AL-052) 20 Hz - 300 kHz (E4980AL-032)
Resolution	0.01 Hz (20 Hz - 99.99 Hz) 0.1 Hz (100 Hz - 999.9 Hz) 1 Hz (1 kHz - 9.999 kHz) 10 Hz (10 kHz - 99.99 kHz) 100 Hz (100 kHz - 999.9 kHz) 1 kHz (1 MHz - 2 MHz)
Measurement accuracy	±0.01%

Table 5. Test signal modes

Normal	Program selected voltage or current at the measurement terminals when	
	they are opened or short-circuited, respectively.	
Constant	Maintains selected voltage or current at the device under test (DUT)	
	independently of changes in impedance of DUT.	

Signal level

Table 6. Test signal voltage

Range		0 Vrms - 2.0 Vrms	
Resolution		100 μVrms (0 Vrms ≤ - ≤ 0.2 Vrms) 200 μVrms (0.2 Vrms < - ≤ 0.5 Vrms) 500 μVrms (0.5 Vrms < - ≤ 1 Vrms) 1 mVrms (1 Vrms < - ≤ 2 Vrms)	
Accuracy	Normal	±(10% + 1 mVrms)	Test frequency ≤ 1 MHz: spec. Test frequency > 1 MHz: typ.
	Constant ¹	±(6% + 1 mVrms)	Test frequency ≤ 1 MHz: spec. Test frequency > 1 MHz: typ.

Table 7. Test signal current

Range		0 Arms - 20 mArms	
Resolution		1 μArms (0 Arms ≤ - ≤ 2 mArms) 2 μArms (2 mArms < - ≤ 5 mArms) 5 μArms (5 mArms < - ≤ 10 mArms) 10 μArms (10 mArms < - ≤ 20 mArms)	
Accuracy	Normal	±(10% + 10 μArms)	Test frequency ≤ 1 MHz: spec. Test frequency > 1 MHz: typ.
	Constant ¹	±(6% + 10 μArms)	Test frequency < = 1 MHz: spec. Test frequency > 1 MHz: typ.

Output impedance: 100Ω (nominal)

Test signal level monitor function

- Test signal voltage and test signal current can be monitored.
- Level monitor accuracy:

Table 8. Test signal voltage monitor accuracy (Vac)

Test signal voltage ²	Test frequency	Specification
5 mVrms - 2 Vrms	≤1 MHz >1 MHz	±(3% of reading value + 0.5 mVrms) ±6% of reading value + 1 mVrms)

Table 9. Test signal current monitor accuracy (lac)

Test signal current ²	Test frequency	Specification
50 μArms - 20 mArms	≤ 1 MHz > 1 MHz	\pm (3% of reading value + 5 μ Arms) \pm (6% of reading value + 10 μ Arms)



^{1.} When auto level control function is on.

^{2.} This is not an output value but rather a displayed test signal level.

Measurement display ranges

Table 10 shows the range of measured value that can be displayed on the screen. For the effective measurement ranges, refer to Figure 1 impedance measurement accuracy example.

Table 10. Allowable display ranges for measured values

Parameter	Measurement display range
Cs, Cp	± 1.000000 aF to 999.9999 EF
Ls, Lp	±1.000000 aH to 999.9999 EH
D	±0.000001 to 9.999999
Q	±0.01 to 99999.99
R, Rs, Rp, X, Z, Rdc	±1.000000 aΩ to 999.9999 EΩ
G, B, Y	±1.000000 aS to 999.9999 ES
Vdc	±1.000000 aV to 999.9999 EV
ldc	±1.000000 aA to 999.9999 EA
θr	±1.000000 arad to 3.141593 rad
θd	±0.0001 deg to 180.0000 deg
Δ%	±0.0001% to 999.9999%

a: 1 x 10⁻¹⁸, E: 1 x 10¹⁸

Absolute measurement accuracy

The following equations are used to calculate absolute accuracy.

Absolute accuracy Aa of |Z|, |Y|, L, C, R, X, G, B (L, C, X, and B accuracies apply when Dx \leq 0.1, R and G accuracies apply when Qx \leq 0.1)

When $D_x \ge 0.1$, multiply Acal by $\sqrt{1+D_X^2}$ for L, C, X, and B accuracies

When $Q_x \ge 0.1$, multiply Acal by $\sqrt{1+Q_x^2}$ for R and G accuracies

Under an AC magnetic field, the following equation is applied to the measurement accuracy.

$$A \times (1 + B \times (2 + 0.5 / Vs))$$

where A Absolute accuracy

B Magnetic flux density [Gauss]

Vs Test signal voltage level [Volts]

Equation 1. Aa = Ae + Acal

Aa Absolute accuracy (% of reading value)

Ae Relative accuracy (% of reading value)

Acal Calibration accuracy (%)

where G accuracy is applied only to G-B measurements.

D accuracy (when $Dx \le 0.1$)

Equation 2. De $+\theta$ cal

Dx Measured D value
De Relative accuracy of D

 θ cal Calibration accuracy of θ (radian)

when $0.1 < Dx \le 1$, multiply θ cal by (1 + Dx)

Q accuracy (When $Qx \times Da < 1$)

Equation 3.
$$\pm \frac{(Qx^2 \times Da)}{(1 + Qx \times Da)}$$

Qx Measured Q value
Da Absolute accuracy of D

θ accuracy

Equation 4. $\theta e + \theta cal$

 θ e Relative accuracy of θ (degree) θ cal Calibration accuracy of θ (degree)



G accuracy (when Dx ≤ 0.1)

Equation 5. Bx + Da (S) $Bx = 2\pi fCx = \frac{1}{2\pi fLx}$

Dx Measured D value
Bx Measured B value (S)
Da Absolute accuracy of D
f Test frequency (Hz)
Cx Measured C value (F)
Lx Measured L value (H)

where the accuracy of G is applied to Cp-G measurements.

Absolute accuracy of Rp (when $Dx \le 0.1$)

Equation 6.
$$\pm \frac{Rpx \times Da}{Dx_{\mp} Da} \quad (\Omega)$$

 $\begin{array}{ll} \text{Rpx} & \text{Measured Rp value } (\Omega) \\ \text{Dx} & \text{Measured D value} \\ \text{Da} & \text{Absolute accuracy of D} \end{array}$

Absolute accuracy of Rs (when Dx ≤ 0.1)

Equation 7.
$$Xx \times Da$$
 (Ω)

$$Xx = \frac{1}{2\pi f Cx} = 2\pi f Lx$$

Dx Measured D value
 Xx Measured X value (Ω)
 Da Absolute accuracy of D
 f Test frequency (Hz)
 Cx Measured C value (F)
 Lx Measured L value (H)

Relative accuracy

Relative accuracy includes stability, temperature coefficient, linearity, repeatability, and calibration interpolation error. Relative accuracy is specified when all of the following conditions are satisfied:

- Warm-up time: 30 minutes
- Test cable length: 0 m, 1 m, 2 m, or 4 m (Keysight Technologies, Inc. 16048A/D/E)
- A "Signal Source Overload" warning does not appear.
 When the test signal current exceeds a value in table 11 below, a "Signal Source Overload" warning appears.

Table 11.

Test signal voltage	Test frequency	Condition ¹
≤ 2 Vrms	-	-
> 2 Vrms	≤1 MHz	The smaller value of either 110 mA or
		130 mA - 0.0015 × Vac × (Fm / 1 MHz) × (L_cable + 0.5)
	> 1 MHz	70 mA - 0.0015 × Vac × (Fm / 1 MHz) × (L_cable + 0.5)

Vac [V] Test signal voltage
Fm [Hz] Test frequency
L_cable [m] Cable length

- OPEN and SHORT corrections have been performed.
- Bias current isolation: Off
- The DC bias current does not exceed a set value within each range of the DC bias current
- The optimum impedance range is selected by matching the impedance of DUT to the effective measuring range.

|Z|, |Y|, L, C, R, X, G, and B accuracy (L, C, X, and B accuracies apply when Dx \leq 0.1, R and G accuracies apply Qx \leq 0.1)

When $D_x > 0.1$, multiply A_p by $\sqrt{1+D_x^2}$ for L, C, X, and B accuracies

When $Q_x > 0.1$, multiply A_a by $\sqrt{1+Q_x^2}$ for R and G accuracies

Relative accuracy Ae is given as:

Equation 8. Ae =
$$[Ab + Zs / |Zm| \times 100 + Yo \times |Zm| \times 100] \times Kt$$

Zm Impedance of DUT
Ab Basic accuracy
Zs Short offset
Yo Open offset

Kt Temperature coefficient

D accuracy

D accuracy De is given as - when Dx ≤ 0.1

Equation 9. De = $\pm Ae/100$

Dx Measured D value

Ae Relative accuracies of |Z|, |Y|, L, C, R, X, G, and B

When $0.1 < Dx \le 1$, multiply De by (1 + Dx)



^{1.} When the calculation result is a negative value, 0 A is applied.

Q accuracy (when Q x De < 1)

Q accuracy Qe is given as:

$$Qe = \pm \frac{(Qx^2 \times De)}{(1_{\mp} Qx \times De)}$$

Qx Measured Q value

De Relative D accuracy

θ accuracy

 θ accuracy θ e is given as:

$$\theta e = \frac{180 \times Ae}{\pi \times 100}$$
 (deg)

Ae Relative accuracies of |Z|, |Y|, L, C, R, X, G, and B

G accuracy (when $Dx \le 0.1$)

G accuracy Ge is given as:

Ge = Bx × De (S)

$$Bx = 2\pi fCx = \frac{1}{2\pi fLx}$$

Ge Relative G accuracy

Dx Measured D value

Bx Measured B value

De Relative D accuracy

f Test frequency (Hz)

Cx Measured C value (F)

Lx Measured L value (H)

Rp accuracy (when $Dx \le 0.1$)

Rp accuracy Rpe is given as:

$$Rpe = \pm \frac{Rpx \times De}{Dx \pm De}$$
 (Ω)

Rpe Relative Rp accuracy

Rpx Measured Rp value (Ω)

Dx Measured D value

De Relative D accuracy

Rs accuracy (when $Dx \le 0.1$)

Rs accuracy Rse is given as:

Equation 14.

Rse =
$$Xx \times De$$
 (Ω)
 $Xx = \frac{1}{2\pi f Cx} = 2\pi f Lx$

Rse Relative Rs accuracy

Dx Measured D value

Xx Measured X value (Ω)

De Relative D accuracy

f Test frequency (Hz)

Cx Measured C value (F)

Lx Measured L value (H)



Example of C-D accuracy calculation

Measurement conditions

Test frequency: 1 kHz
Measured C value: 100 nF
Test signal voltage: 1 Vrms
Measurement time mode: Medium
Measurement temperature: 23°C

Ab = 0.05%

 $|Zm| = 1 / (2\pi \times 1 \times 103 \times 100 \times 10-9) = 1590 \Omega$

 $Zs = 0.6 \text{ m}\Omega \times (1 + 0.400/1) \times (1 + \sqrt{(1000/1000)} = 1.68 \text{ m}\Omega$ $Yo = 0.5 \text{ nS} \times (1 + 0.100/1) \times (1 + \sqrt{(100/1000)} = 0.72 \text{ nS}$

C accuracy: Ae = $[0.05 + 1.68 \text{ m}/1590 \times 100 + 0.72 \text{ n} \times 1590 \times 100] \times 1 = 0.05\%$

D accuracy: De = 0.05/100 = 0.0005

Basic accuracy

Basic accuracy Ab is given from Table 12, 13, 14 and 15.

Table 12. Measurement time mode = SHORT

Test		Test signal volt	age		
frequency [Hz]	5 mVrms ≤ - < 50 mVrms	50 mVrms ≤ - < 0.3 Vrms	0.3 Vrms ≤ - ≤ 1 Vrms	1 Vrms < - ≤ 10 Vrms	10 Vrms < - ≤ 20 Vrms
20 ≤ - < 125	(0.6%) × (50 mVrms/Vs)	0.60%	0.30%	0.30%	0.30%
125 ≤ - ≤ 1 M	(0.2%) × (50 mVrms/Vs)	0.20%	0.10%	0.15%	0.15%
1 M < - ≤ 2 M	(0.4%) × (50 mVrms/Vs)	0.40%	0.20%	0.30%	0.30%

Table 13. Measurement time mode = MED, LONG

Test	Test signal voltage				
frequency [Hz]	5 mVrms ≤ - < 30 mVrms	30 mVrms ≤ - < 0.3 Vrms	0.3 Vrms ≤ - ≤ 1 Vrms	1 Vrms < - ≤ 10 Vrms	10 Vrms < - ≤ 20 Vrms
20 ≤ - < 100	(0.25%) × (30 mVrms/Vs)	0.25%	0.10%	0.15%	0.15%
100 ≤ - ≤ 1 M	(0.1%) × (30 mVrms/Vs)	0.10%	0.05%	0.10%	0.15%
1 M < - ≤ 2 M	(0.2%) × (30 mVrms/Vs)	0.20%	0.10%	0.20%	0.30%

Vs [Vrms] Test signal voltage



Effect by impedance of DUT

Table 14. For impedance of DUT below 30 Ω , the following value is added.

Test frequency [Hz]	Impedance of DUT		
	1.08 Ω ≤ Zx < 30 Ω	Zx < 1.08 Ω	
20 ≤ - ≤ 1 M	0.05%	0.10%	
1 M < - ≤ 2 M	0.10%	0.20%	

Table 15. For impedance of DUT over 9.2 k Ω , the following value is added.

Test frequency [Hz]	Impedance of DUT		
	9.2 kΩ < Zx ≤ 92 kΩ	92 kΩ < Zx	
10 k ≤ - ≤ 100 k	0%	0.05%	
100 k < - ≤ 1 M	0.05%	0.05%	
1 M < - ≤ 2 M	0.10%	0.10%	

Effect of cable extension

When the cable is extended, the following element is added per one meter.

 $0.015 \% \times (Fm/1 MHz)^2 \times (L_cable)^2$

Fm [Hz] Test frequency L_cable [m] Cable length

Short offset Zs

Table 16. Impedance of DUT > 1.08 Ω

Test	Measurement time mode	
frequency [Hz]	SHORT	MED, LONG
20 - 2 M	2.5 m $\Omega \times (1 + 0.400/Vs) \times (1 + \sqrt{(1000/Fm)})$	$0.6 \text{ m}\Omega \times (1 + 0.400/\text{Vs}) \times (1 + \sqrt{(1000/\text{Fm})})$

Table 17. Impedance of DUT \leq 1.08 Ω

Test	Measurement time mode		
frequency [Hz]	SHORT	MED, LONG	
20 - 2 M	1 m Ω × (1 + 1/Vs) × (1 + $\sqrt{(1000/Fm)}$)	$0.2 \text{ m}\Omega \times (1 + 1/\text{Vs}) \times (1 + \sqrt{(1000/\text{Fm})})$	

Vs [Vrms] Test signal voltage Fm [Hz] Test frequency

Effect of cable extension (Short offset)

Table 18. When the cable is extended, the following value is added to Zs (independent of the measurement time mode).

Test	Cable length			
frequency [Hz]	0 m	1 m	2 m	4 m
20 ≤ - ≤ 1 M	0	$0.25~\text{m}\Omega$	$0.5~\text{m}\Omega$	$1~\text{m}\Omega$
1 M < - ≤ 2 M	0	1 mΩ	2 mΩ	4 mΩ

Open offset Yo

Table 19. Test signal voltage ≤ 2.0 Vrms

Test	Measurement time mode		
frequency [Hz]	SHORT	MED, LONG	
20 ≤ - ≤ 100 k	$2 \text{ nS} \times (1 + 0.100/\text{Vs}) \times (1 + \sqrt{(100/\text{Fm}))}$	$0.5 \text{ nS} \times (1 + 0.100/\text{Vs}) \times (1 + \sqrt{(100/\text{Fm})})$	
100 k < - ≤ 1 M	20 nS × (1 + 0.100/Vs)	5 nS × (1 + 0.100/Vs)	
1 M < - ≤ 2 M	40 nS × (1 + 0.100/Vs)	10 nS × (1 + 0.100/Vs)	

Table 20. Test signal voltage > 2.0 Vrms

Test	Measurement time mode		
frequency [Hz]	SHORT	MED, LONG	
20 ≤ - ≤ 100 k	$2 \text{ nS} \times (1 + 2/\text{Vs}) \times (1 + \sqrt{(100/\text{Fm}))}$	$0.5 \text{ nS} \times (1 + 2/\text{Vs}) \times (1 + \sqrt{(100/\text{Fm}))}$	
100 k < - ≤ 1 M	20 nS × (1 + 2/Vs)	5 nS × (1 + 2/Vs)	
1 M < - ≤ 2 M	40 nS × (1 + 2/Vs)	10 nS × (1 + 2/Vs)	

Vs [Vrms] Test signal voltage Fm [Hz] Test frequency



The Open Offset may become three times greater in the ranges of 40 to 70 kHz and 80 to 100 kHz due to residual response.



Effect of cable length

Table 21. When the cable is extended, multiply Yo by the following factor.

y y y y y y y y					
Test	Cable	Cable length			
frequency [Hz]	0 m	1 m	2 m	4 m	
100 ≤ - < 100 k	1	$1 + 5 \times Fm/1 MHz$	$1 + 10 \times Fm/1 MHz$	$1 + 20 \times Fm/1 MHz$	
100 k ≤ - ≤ 1 M	1	1 + 0.5 × Fm/1 MHz	1 + 1 × Fm/1 MHz	1 + 2 × Fm/1 MHz	
1 M < - ≤ 2 M	1	1 + 1 × Fm/1 MHz	1 + 2 × Fm/1 MHz	1 + 4 × Fm/1 MHz	

Fm [Hz] Test frequency

Temperature factor Kt

Table 22. The temperature factor Kt is given below.

Temperature [°C]	Kt
0 < - < 18	4
18 ≤ - ≤ 28	1
28 < - ≤ 55	4

Calibration accuracy Acal

Calibration accuracy Acal is given below.

For impedance of DUT or test frequency on the boundary line, apply the smaller value.

Table 23. Impedance range = 0.1, 1, 10 Ω

	Test frequency [Hz]					
	20 - 1 k	1 k - 10 k	10 k -100 k	100 k - 300 k	300 k - 1 M	1 M - 2 M
Z [%]	0.03	0.05	0.05	0.05 + 5 × 10 ⁻⁵ Fm	0.05 + 5 × 10 ⁻⁵ Fm	0.1 + 1 × 10 ⁻⁴ Fm
θ [radian]	1 × 10 ⁻⁴	2 × 10 ⁻⁴	3 × 10 ⁻⁴	3 × 10 ⁻⁴ + 2 × 10-7 Fm	3 × 10 ⁻⁴ + 2 × 10-7 Fm	6 × 10 ⁻⁴ + 4 × 10-7 Fm

Table 24. Impedance range = 100Ω

	Test frequency [Hz]					
	20 - 1 k	1 k - 10 k	10 k -100 k	100 k - 300 k	300 k - 1 M	1 M - 2 M
Z [%]	0.03	0.05	0.05	0.05 + 5 × 10 ⁻⁵ Fm	0.05 + 5 × 10 ⁻⁵ Fm	0.1 + 1 × 10 ⁻⁴ Fm
θ [radian]	1 × 10 ⁻⁴	2 × 10 ⁻⁴	3 × 10 ⁻⁴	3×10^{-4}	3 × 10 ⁻⁴	6 × 10 ⁻⁴

Table 25. Impedance range = 300, 1 $k\Omega$

	Test frequency [Hz]					
	20 - 1 k	1 k - 10 k	10 k -100 k	100 k - 300 k	300 k - 1 M	1 M - 2 M
Z [%]	0.03	0.03	0.05	0.05	0.05	0.1
θ [radian]	1 × 10 ⁻⁴	1 × 10 ⁻⁴	3 × 10 ⁻⁴	3 × 10 ⁻⁴	3 × 10 ⁻⁴	6 × 10 ⁻⁴

Table 26. Impedance range = 3 k, 10 k Ω

	Test frequency [Hz]					
	20 - 1 k	1 k - 10 k	10 k -100 k	100 k - 300 k	300 k - 1 M	1 M - 2 M
Z [%]	0.03 + 1 × 10 ⁻⁴ Fm	0.03 + 1 × 10 ⁻⁴ Fm	0.00	0.00	0.03 + 1 × 10 ⁻⁴ Fm	0.06 + 2 × 10 ⁻⁴ Fm
θ [radian]		(100 + 2.5 Fm) × 10 ⁻⁶				(200 + 5 Fm) × 10 ⁻⁶

Table 27. Impedance range = 30 k, 100 k Ω

	Test frequency [Hz]					
	20 - 1 k	1 k - 10 k	10 k -100 k	100 k - 300 k	300 k - 1 M	1 M - 2 M
Z [%]				0.00	0.03 + 1 × 10 ⁻⁴ Fm	0.06 + 2 × 10 ⁻⁴ Fm
θ [radian]		(100 + 20 Fm) × 10 ⁻⁶		(100 + 20 Fm) × 10 ⁻⁶	(100 + 2.5 Fm) × 10 ⁻⁶	(200 + 5 Fm) × 10 ⁻⁶

Fm[kHz] Test frequency



Measurement accuracy

The impedance measurement calculation example below is the result of absolute measurement accuracy.




Figure 1. Impedance measurement accuracy (Test signal voltage = 1 Vrms, cable length = 0 m, measurement time mode = MED)

Compensation function

Table 28. The E4980A provides three types of compensation functions: OPEN compensation, SHORT compensation, and LOAD compensation.

Type of compensation	Description
OPEN compensation	Compensates errors caused by the stray admittance (C, G) of the test fixture.
SHORT compensation	Compensates errors caused by the residual impedance (L, R) of the test fixture.
LOAD compensation	Compensates errors between the actual measured value and a known standard value under the measurement conditions desired by the user.

List sweep

Points: There is a maximum of 201 points.

First sweep parameter (primary parameter): Test frequency, test signal voltage, test signal current, test signal voltage of DC bias signal, test signal current of DC bias signal, DC source voltage.

Second sweep parameter (secondary parameter): None, impedance range, test frequency, test signal voltage, test signal current, test signal voltage of DC bias signal, test signal current of DC bias signal, DC source voltage

Trigger mode

Sequential mode: When the E4980A is triggered once, the device is measured at all sweep points. /EOM/INDEX is output only once.

Step mode: The sweep point is incremented each time the E4980A is triggered. /EOM/INDEX is output at each point, but the result of the comparator function of the list sweep is available only after the last /EOM is output.

Note

A parameter selected for one of the two parameters cannot be selected for the other parameter. It is not possible to set up a combination of test signal voltage and test signal current or one of test signal voltage of DC bias signal and test signal current of DC bias.

The secondary parameter can be set only with SCPI commands.

Comparator function of list sweep: The comparator function enables setting one pair of lower and upper limits for each measurement point.

You can select from: Judge with the first sweep parameter/Judge with the second parameter/Not used for each pair of limits.

Time stamp function: In the sequential mode, it is possible to record the measurement starting time at each measurement point by defining the time when FW detects a trigger as 0 and obtain it later with the SCPI command.

Comparator function

Bin sort: The primary parameter can be sorted into 9 BINs, OUT_OF_BINS, AUX_BIN, and LOW_C_REJECT. The secondary parameter can be sorted into HIGH, IN, and LOW. The sequential mode and tolerance mode can be selected as the sorting mode.

Limit setup: Absolute value, deviation value, and % deviation value can be used for setup.

BIN count: Countable from 0 to 999999.

DC bias signal

Table 29. Test signal voltage

Range	0 V to +2 V
Resolution	0 V / 1.5 V / 2 V only
Accuracy	0.1% + 2 mV (23°C ± 5°C) (0.1% + 2 mV) × 4
	(0 to 18°C or 28 to 55°C)

Output impedance: 100Ω (nominal)

Measurement assistance functions

Data buffer function: Up to 201 measurement results can be read out in a batch.

Save/Recall function:

- Up to 10 setup conditions can be written to/read from the built-in non-volatile memory.
- Up to 10 setup conditions can be written to/read from the USB memory.
- Auto recall function can be performed when the setting conditions are written to Register 10 of the USB memory.

Key lock function: The front panel keys can be locked.

GPIB: 24-pin D-Sub (Type D-24), female; complies with IEEE488.1, 2 and SCPI

USB host port: Universal serial bus jack, type-A (4 contact positions, contact 1 is on your left), female (for connection to USB memory only).

USB interface port: Universal serial bus jack, type mini-B (4 contact positions); complies with USBTMC-USB488 and USB 2.0; female; for connection to the external controller.

USBTMC: Abbreviation for USB Test & Measurement Class

LAN: 10/100 BaseT Ethernet, 8 pins (two speed options)

LXI Compliance: Class C (only applies to units with firmware revision A.02.00 or later)

Note

This note is for the E4980A/E4980AL units with firmware A.xx.xx.

The following USB memory can be used. Complies with USB 1.1; mass storage class, FAT16/FAT32 format; maximum consumption current is below 500 mA.

Recommended USB memory: 4 GB USB Flash memory (Keysight PN 1819-0637) and 16GB USB Flash memory (Keysight PN 1819-1235).

Use the recommended USB memory device exclusively for the E4980A/E4980AL, otherwise, previously saved data may be cleared. If you use a USB memory other than the recommended device, data may not be saved or recalled normally.

Keysight will NOT be responsible for data loss in the USB memory caused by using the E4980A/ E4980AL.



Options

Vote

Option xxx is described as E4980A-xxx in the order information

Frequency options

E4980A	20 Hz to 2 MHz	
E4980AL-032	20 Hz to 300 kHz	
E4980AL-052	20 Hz to 500 kHz	
E4980AL-102	20 Hz to 1 MHz	

Table 30. Installable options

Options	E4980A	E4980AL
Power and DC bias enhancement (001)	Installable	Not installable
DCR measurement (200)	Installable ¹	Not installable ²
Handler interface (201)	Installable	Installable
Scanner interface (301)	Installable	Installable

Interface options

Option 201 (Handler interface)

Adds handler interface.

Option 301 (Scanner interface)

Adds scanner interface.

Option 710 (No interface)

An option with no interface.

Up to 2 interface options can be installed in the interface connector on the rear panel.

When no interface is installed, two of the option 710 are installed. When one interface is installed, the option number of its interface and one option 710 are installed.

Other options

Option 001 (Power and DC Bias enhancement)

Increases test signal voltage and adds the variable DC bias voltage.

Option 007 (Standard model)

Upgrades the entry model to the standard (only available in E4980AU).

Option 200 (DCR measurement)

Adds DCR measurement

Note Option 007 can be installed only in the E4980A with option 005.

Note

E4980A-200/001 and E4980AL-032/052/102 supports DCR measurement function.

- 1. Mandatory option
- 2. DCR measurement function is equipped by default.



Power and DC bias enhancement specification

Increases test signal voltage and adds the variable DC bias voltage function. The Vdc-Idc measurement function is available when the option 001 is installed.

Measurement parameters

The following parameters can be used.

- Lp-Rdc
- Ls-Rdc
- Vdc-ldc

where

Rdc Direct-current resistance (DCR)

Vdc Direct-current voltage Idc Direct-current electricity

Test signal

Signal level

Table 31. Test signal voltage

Range		0 Vrms to 20 Vrms (test frequency ≤ 1 MHz) 0 Vrms to 15 Vrms (test frequency > 1 MHz)
Resolution		100 μVrms (0 Vrms < - ≤ 0.2 Vrms) 200 μVrms (0.2 Vrms < - ≤ 0.5 Vrms) 500 μVrms (0.5 Vrms < - ≤ 1 Vrms) 1 mVrms (1 Vrms < - ≤ 2 Vrms) 2 mVrms (2 Vrms < - ≤ 5 Vrms) 5 mVrms (5 Vrms < - ≤ 10 Vrms) 10 mVrms (10 Vrms < - ≤ 20 Vrms)
Setup accuracy	normal	±(10% + 1 mVrms) (test signal voltage ≤ 2 Vrms) (test frequency ≤ 1 MHz : spec., test frequency > 1 MHz : typ.) ±(10% + 10 mVrms) (Test frequency ≤ 300 kHz, test signal voltage > 2 Vrms) (spec.) ±(15% + 20 mVrms) (test frequency > 300 kHz, test signal voltage > 2 Vrms) (test frequency ≤ 1 MHz : spec., test frequency > 1 MHz : typ.)
	Constant ¹	±(6% + 1 mVrms) (test signal voltage ≤ 2 Vrms) (test frequency ≤ 1 MHz : spec., test frequency > 1 MHz : typ.) ±(6% + 10 mVrms) (test frequency ≤ 300 kHz, test signal voltage > 2 Vrms) (spec.) ±(12% + 20 mVrms) (test frequency > 300 kHz, test signal voltage > 2 Vrms) (test frequency ≤ 1 MHz : spec., test frequency > 1 MHz : typ.)



Table 32. Test signal current

Range		0 Arms - 100 mArms
Resolution		1 μArms (0 Arms ≤ - ≤ 2 mArms)
		2 μArms (2 mArms < - ≤ 5 mArms)
		5 μArms (5 mArms < - ≤ 10 mArms)
		10 μArms (10 mArms < - ≤ 20 mArms)
		20 μArms (20 mArms < - ≤ 50 mArms)
		$50 \mu\text{Arms}$ ($50 \text{mArms} < - \le 100 \text{mArms}$)
Setup accuracy	normal	±(10% + 10 µArms) (test signal voltage ≤ 20 mArms) (test frequency ≤ 1 MHz : spec., test frequency > 1 MHz : typ.)
		±(10% + 100 µArms) (test frequency ≤ 300 kHz, test signal current > 20 mArms) (spec.)
		±(15% + 200 µArms) (test frequency > 300 kHz, test signal voltage > 20 mArms) (test frequency ≤ 1 MHz: spec., test frequency > 1 MHz: typ.)
	Constant ¹	±(6% + 10 µArms) (test signal voltage ≤ 20 mArms) (test frequency ≤ 1 MHz : spec., test frequency > 1 MHz : typ.)
		±(6% + 100 µArms) (test frequency ≤ 300 kHz, test signal voltage > 20 mArms) (spec.)
		±(12% + 200 µArms) (test frequency > 300 kHz, test signal voltage > 20 mArms) (test frequency ≤ 1 MHz: spec., test frequency > 1 MHz: typ.)

Test signal level monitor function

- Test signal voltage and test signal current can be monitored.
- Level monitor accuracy:

Table 33. Test signal voltage monitor accuracy (Vac)

Test signal voltage ²	Test frequency	Specification
5 mVrms to 2 Vrms	≤1 MHz	±(3% of reading value + 0.5 mVrms)
	> 1MHz	±(6% of reading value + 1 mVrms)
> 2 Vrms	≤ 300 kHz	±(3% of reading value + 5 mVrms)
	> 300 kHz	±(6% of reading value + 10 mVrms) ³

Table 34. Test signal current monitor accuracy (Iac)

Test signal current ²	Test frequency	Specification
$50\mu\text{Arms}$ to 20mArms	≤1 MHz	\pm (3% of reading value + 5 μ Arms)
	> 1MHz	±(6% of reading value + 10 μArms)
> 20 mArms	≤ 300 kHz	±(3% of reading value + 50 μArms)
	> 300 kHz	±(6% of reading value + 100 μArms)



^{1.} When auto level control function is on.

This is not an output value but a displayed test signal level

^{3.} Typ. when test frequency is > 1 MHz with test signal voltage > 10 Vrms.

DC bias signal

Table 35. Test signal voltage

Range		-40 V to +40 V	
Resolution		Setup resolution: 100 μ V, effective resolution: 330 μ V ±(0 V ≤ - ≤ 5 V)	
		$1 \text{ mV} \pm (5 \text{ V} < - \le 10 \text{ V})$	
		$2 \text{ mV} \pm (10 \text{ V} < - \le 20 \text{ V})$	
		$5 \text{ mV} \pm (20 \text{ V} < - \le 40 \text{ V})$	
Accuracy	test signal voltage ≤ 2 Vrms	0.1% + 2 mV (23°C ±5°C)	
		(0.1% + 2 mV) x 4	
		(0 to 18°C or 28 to 55°C)	
	test signal voltage > 2 Vrms	0.1 % + 4 mV (23°C ±5°C)	
		(0.1% + 4 mV) x 4	
		(0 to 18°C or 28 to 55°C)	
Table 36. Te	est signal current		
Range		–100 mA - 100 mA	
Resolution		Setup resolution: 1 μ A, effective resolution: 3.3 μ A \pm (0 A \leq - \leq 50 mA)	
		$10 \mu A \pm (50 mA < - \le 100 mA)$	

DC bias voltage level monitor Vdc

 $(0.5\% \text{ of reading value} + 60 \text{ mV}) \times \text{Kt}$

When using Vdc-ldc measurement: (spec.)

When using level monitor: (typ.)

Kt Temperature coefficient

DC bias current level monitor Idc

(A [%] of the measurement value + B [A]) \times Kt

When using Vdc-Idc measurement: (spec.)

When using level monitor: (typ.)

A [%] When the measurement time mode is SHORT: 2% When the measurement time mode is MED or LONG: 1%

B [A] given below

Kt Temperature coefficient

When the measurement mode is SHORT, double the following value.



Table 37. Test signal voltage ≤ 0.2 Vrms (measurement time mode = MED, LONG)

DC bias	Impedance range $[\Omega]$				
current range	< 100	100	300, 1 k	3 k, 10 k	30k, 100 k
20 μΑ	150 μΑ	30 μΑ	3 μΑ	300 nA	45 nA
200 μΑ	150 μΑ	30 μΑ	3 μΑ	300 nA	300 nA
2 mA	150 μΑ	30 μΑ	3 μΑ	3 μΑ	3 μΑ
20 mA	150 μΑ	30 μΑ	30 μΑ	30 μΑ	30 μΑ
100 mA	150 μΑ	150 μΑ	150 μΑ	150 μΑ	150 μΑ

Table 38. 0.2 Vrms < test signal voltage ≤ 2 Vrms (measurement time mode = MED, LONG)

DC bias	Impedance range $[\Omega]$				
current range	< 100	100, 300	1k, 3 k	10k, 30 k	100 k
20 μΑ	150 μΑ	30 μΑ	3 μΑ	300 nA	45 nA
200 μΑ	150 μΑ	30 μΑ	3 μΑ	300 nA	300 nA
2 mA	150 μΑ	30 μΑ	3 μΑ	3 μΑ	3 μΑ
20 mA	150 μΑ	30 μΑ	30 μΑ	30 μΑ	30 μΑ
100 mA	150 μΑ	150 μΑ	150 μΑ	150 μΑ	150 μΑ

Table 39. Test signal voltage > 2 Vrms (measurement time mode = MED, LONG)

DC bias $\qquad \qquad \text{Impedance range } [\Omega]$				
current range	≤ 300	1 k, 3 k	10k, 30 k	100 k
20 μΑ	150 μΑ	30 μΑ	3 μΑ	300 nA
200 μΑ	150 μΑ	30 μΑ	3 μΑ	300 nA
2 mA	150 μΑ	30 μΑ	3 μΑ	3 μΑ
20 mA	150 μΑ	30 μΑ	30 μΑ	30 μΑ
100 mA	150 μΑ	150 μΑ	150 μΑ	150 μΑ

Table 40. Input impedance (nominal)

Input impedance	Conditions
0 Ω	Other than conditions below.
20 Ω	Test signal voltage ≤ 0.2 Vrms, Impedance range ≥ 3 k $\Omega,$ DC bias current range $\leq 200~\mu\text{A}$
	Test signal voltage ≤ 2 Vrms, Impedance range ≥ 10 k Ω , DC bias current range ≤ 200 μA
	Test signal voltage > 2 Vrms, Impedance range = 100 k Ω , DC bias current range \leq 200 μ A

DC source signal

Table 41. Test signal voltage

Range	–10 V to 10 V
Resolution	1 mV
Accuracy	0.1% + 3 mV (23°C ±5°C) (0.1% + 3 mV) x 4 (0 to 18°C or 28 to 55°C)

Table 42. Test signal current

Range	-45 mA to 45 mA (nominal)	

Output impedance 100Ω (nominal)



DCR measurement specification

DC resistance (Rdc) measurement function is available when either E4980A-001/200 or E4980AL-032/052/102 is installed.

DC resistance (Rdc) accuracy

Absolute measurement accuracy Aa

Absolute measurement accuracy Aa is given as

Equation 15. Aa = Ae + Acal

Aa Absolute accuracy (% of reading value)
Ae Relative accuracy (% of reading value)

Acal Calibration accuracy

Relative measurement accuracy Ae

Relative measurement accuracy Ae is given as

Equation 16. Ae =
$$[Ab + (Rs / |Rm| + Go \times |Rm|) \times 100] \times Kt$$

Rm Measurement value

Ab Basic accuracy Rs Short offset $[\Omega]$ Go Open offset [S]

Kt Temperature coefficient

Calibration accuracy Acal

Calibration accuracy Acal is 0.03%.

Basic accuracy Ab

Table 43. Basic accuracy Ab is given below.

Measurement time mode	Test signal voltage		
	≤ 2 Vrms	> 2 Vrms	
SHORT	1.00%	2.00%	
MED	0.30%	0.60%	

Open offset Go

Table 44. Open offset Go is given below.

Measurement time mode	Test signal voltage	
	≤ 2 Vrms	> 2 Vrms
SHORT	50 nS	500 nS
MED	10 nS	100 nS

Short offset Rs

Table 45. Short offset Rs is given below.

Measurement time mode	Test signal voltage		
	≤ 2 Vrms	> 2 Vrms	
SHORT	$25~\text{m}\Omega$	250 mΩ	
MED	5 mΩ	50 mΩ	



Effect of cable length (Short offset)

Table 46. The following value is added to Rs when the cable is extended.

Cable length		
1 m	2 m	4 m
0.25 mΩ	0.5 mΩ	1 mΩ

Temperature coefficient Kt

Table 47. Temperature coefficient Kt is given below.

Temperature [°C]	Kt
0 ≤ - < 18	4
18 ≤ - ≤ 28	1
28 < - ≤ 55	4

General Specifications

T 1 1	10		
lahle	48	POWER	source

Voltage	90 VAC - 264 VAC
Frequency	47 Hz - 63 Hz
Power consumption	Max. 150 VA
Table 49. Operating environment	
Temperature	0 - 55°C
Humidity (≤ 40°C, no condensation)	15% - 85% RH
Altitude	0 m - 2000 m
Table 50. Storage environment	
Temperature	–20 - 70°C
Humidity (≤ 60°C, no condensation)	0% - 90% RH
Altitude	0 m - 4572 m

Outer dimensions: 375 (width) x 105 (height) × 390 (depth) mm (nominal)





Figure 2. Dimensions (front view, with handle and bumper, in millimeters, nominal)



 $\label{thm:continuous} \mbox{Figure 3. Dimensions (front view, without handle and bumper, in millimeters, nominal) }$




Figure 4. Dimensions (rear view, with handle and bumper, in millimeters, nominal)




Figure 5. Dimensions (rear view, with handle and bumper, in millimeters, nominal)




Figure 6. Dimensions (side view, with handle and bumper, in millimeters, nominal)




Figure 7. Dimensions (side view, without handle and bumper, in millimeters, nominal)

Weight: 5.3 kg (nominal)

Display: LCD, 320 × 240 (pixels), RGB color

The following items can be displayed:

- Measurement value
- Measurement conditions
- Limit value and judgment result of comparator
- List sweep table
- Self-test message

Note

Effective pixels are more than 99.99%. There may be 0.01% (approx. 7 pixels) or smaller missing pixels or constantly lit pixels, but this is not a malfunction.

Description Supplemental Information **EMC** European Council Directive 2004/108/EC IEC 61326-1:2012 EN 61326-1:2013 CISPR 11:2009 +A1:2010 EN 55011: 2009 +A1:2010 Group 1, Class A IEC 61000-4-2:2008 EN 61000-4-2:2009 4 kV CD / 8 kV AD IEC 61000-4-3:2006 +A1:2007 +A2:2010 EN 61000-4-3:2006 +A1:2008 +A2:2010 3 V/m, 80-1000 MHz, 1.4 - 2.0 GHz / 1 V/m, 2.0 - 2.7 GHz, 80% AMIEC 61000-4-4:2004 +A1:2010 EN 61000-4-4:2004 +A1:2010 1 kV power lines / 0.5 kV signal lines IEC 61000-4-5:2005 EN 61000-4-5:2006 0.5 kV line-line / 1 kV line-ground IEC 61000-4-6:2008 EN 61000-4-6:2009 3 V, 0.15-80 MHz, 80% AM IEC 61000-4-8:2009 EN 61000-4-8:2010 30A/m, 50/60Hz IEC 61000-4-11:2004 EN 61000-4-11:2004 0.5-300 cycle, 0% / 70% Note: When tested at 3 V/m according to EN61000-4-3, the measurement accuracy will be within specifications over the full immunity test frequency range except when the meter frequency is identical to the transmitted interference signal test frequency (the frequencies around the carrier frequency and frequencies around the modulation frequency). ICES/NMB-001 ICES-001:2006 Group 1, Class A AS/NZS CISPR11:2004 Group 1, Class A KN11, KN61000-6-1 and KN61000-6-2 Group 1, Class A MSIP-REM-Kst-WNMODSF36 Safety European Council Directive 2006/95/EC IEC 61010-1:2001/EN 61010-1:2001 Measurement Category I, Pollution Degree 2, Indoor Use IEC60825-1:1994 Class 1 LED CAN/CSA C22.2 61010-1-04 Measurement Category I, Pollution Degree 2, Indoor Use

Environment



This product complies with the WEEE Directive (2002/96/EC) marking requirements. The afixed label indicates that you must not discard this electrical/electronic product in domestic house hold waste.

Product Category: With reference to the equipment types in the WEEE Directive Annex I, this product is classed as a "Monitoring and Control instrumentation" product.



Supplemental Information

Settling time

Table 51. Test frequency setting time

Test frequency setting time	Test frequency (Fm)
5 ms	Fm ≥ 1 kHz
12 ms	1 kHz > Fm ≥ 250 Hz
22 ms	250 Hz > Fm ≥ 60 Hz
42 ms	60 Hz > Fm

Table 52. Test signal voltage setting time

Test signal voltage setting time	Test frequency (Fm)
11 ms	Fm ≥ 1 kHz
18 ms	1 kHz > Fm ≥ 250 Hz
26 ms	250 Hz > Fm ≥ 60 Hz
48 ms	60 Hz > Fm

Switching of the impedance range is as follows:

≤ 5 ms/ range switching

Measurement circuit protection

The maximum discharge withstand voltage, where the internal circuit remains protected if a charged capacitor is connected to the UNKNOWN terminal, is given below.

Table 53. Maximum discharge withstand voltage

Maximum discharge withstand voltage	Range of capacitance value C of DUT
1000 V	C < 2 µF
√2/C V	2 μF ≤ C

1200 1000 800 Voltage [V] 600 400 200 0 1.E-13 1.E-11 1.E-09 1.E-07 1.E-05 1.E-03 1.E-15 Capacitance [F]

Figure 8. Maximum discharge withstand voltage

Note

Discharge capacitors before connecting them to the UNKNOWN terminal or a test fixture to avoid damages to the instrument.

Measurement time

Definition

This is the time between the trigger and the end of measurement (EOM) output on the handler interface.

Conditions

Table 54, 55 and 56 show the measurement time when the following conditions are satisfied:

- Normal impedance measurement other than Ls-Rdc, Lp-Rdc, Vdc-Idc
- Impedance range mode: hold range mode
- DC bias voltage level monitor: OFF
- DC bias current level monitor: OFF
- Trigger delay: 0 sStep delay: 0 sCalibration data: OFFDisplay mode: blank

Table 54. E4980A measurement time [ms](DC bias:OFF)

	Measurement	Test free	luency					
	time mode	20 Hz	100 Hz	1 kHz	10 kHz	100 kHz	1 MHz	2 MHz
1	LONG	480	300	240	230	220	220	220
2	MED	380	180	110	92	89	88	88
3	SHORT	330	100	20	7.7	5.7	5.6	5.6




Figure 9. Measurement time (E4980A, DC bias: OFF)

Note E4980A-005 is obsolete and not available in a new purchase.

Table 55. E4980A-005 measurement time [ms] (DC bias: OFF)

	Measurement	Test frequency							
	time mode	20 Hz	100 Hz	1 kHz	10 kHz	100 kHz	1 MHz	2 MHz	
1	LONG	1190	650	590	580	570	570	570	
2	MED	1150	380	200	180	180	180	180	
3	SHORT	1040	240	37	25	23	23	23	




Figure 10. Measurement time (DC bias: OFF, E4980A-005)

Table 56. E4980AL measurement time [ms]

	Measurement	Test frequency							
	time mode	20 Hz	100 Hz	1 kHz	10 kHz	100 kHz	1 MHz		
1	LONG	729	423	363	353	343	343		
2	MED	650	250	140	122	119	118		
3	SHORT	579	149	26	14	12	12		




Figure 11. Measurement time (E4980AL)



When DC bias is ON, the following time is added:

Table 57. Additional time when DC bias is ON [ms]

Test frequency							
20 Hz	100 Hz	1 kHz	10 kHz	100 kHz	1 MHz	2 MHz	
30	30	10	13	2	0.5	0.5	

When the number of averaging increases, the measurement time is given as

Equation 17. MeasTime + (Ave – 1) × AveTime

MeasTime Measurement time calculated based on Table 54, 55 and 56

Ave Number of averaging AveTime Refer to Table 58

Table 58. Additional time per averaging [ms]

Measurement	Test fred	Test frequency							
time mode	20 Hz	100 Hz	1 kHz	10 kHz	100 kHz	1 MHz	2 MHz		
SHORT	51	11	2.4	2.3	2.3	2.2	2.2		
MED	110	81	88	87	85	84	84		
LONG	210	210	220	220	220	210	210		

Table 59. Measurement time when Vdc-Idc is selected [ms]

Measurement time	Test frequency								
mode	20 Hz	100 Hz	1 kHz	10 kHz	100 kHz	1 MHz	2 MHz		
SHORT	210	46	14	14	14	14	14		
MED	210	170	170	170	170	170	170		
LONG	410	410	410	410	410	410	410		

Add the same measurement time per 1 additional average

Additional Measurement time when the Vdc and Idc monitor function is ON. Add SHORT mode of Table 59. When using only Vdc or Idc, add a half of SHORT mode of Table 59.

Table 60. Measurement time when Ls-Rdc or Lp-Rdc is selected [ms]

Measurement time	Test frequency						
mode	20 Hz	100 Hz	1 kHz	10 kHz	100 kHz	1 MHz	2 MHz
SHORT	910	230	43	24	22	22	22
MED	1100	450	300	280	270	270	270
LONG	1400	820	700	670	660	650	650

Add the three times of Table 58 Additional Time per 1 additional average number



Display time

Except for the case of the DISPLAY BLANK page, the time required to update the display on each page (display time) is as follows. When a screen is changed, drawing time and switching time are added. The measurement display is updated about every 100 ms.

Table 61. Display time

Item	When Vdc, Idc monitor is OFF	When Vdc, Idc monitor is ON
MEAS DISPLAY page drawing time	10 ms	13 ms
MEAS DISPLAY page (large) drawing time	10 ms	13 ms
BIN No. DISPLAY page drawing time	10 ms	13 ms
BIN COUNT DISPLAY page drawing time	10 ms	13 ms
LIST SWEEP DISPLAY page drawing time	40 ms	_
Measurement display switching time	35 ms	_

Measurement data transfer time

This table shows the measurement data transfer time under the following conditions. The measurement data transfer time varies depending on measurement conditions and computers.

Table 62. Measurement transfer time under the following conditions:

Host computer:	HP Z420 Workstation, Xeon CPU ES-1620 0 @3.60 GHz
Display:	OFF
Impedance range mode:	AUTO (The overload has not been generated.)
OPEN/SHORT/LOAD compensation:	OFF
Test signal voltage monitor:	OFF

Table 63. Measurement data transfer time [ms]

Interface	Data transfer format	Using :FETC? command (one point measurement)		Using data buffer memory (list sweep measurement)			
		Comparator ON	Comparator OFF	10 points	51 points	128 points	201 points
GPIB	ASCII	2	2	4	13	28	43
	ASCII Long	2	2	5	15	34	53
	Binary	2	2	4	10	21	36
USB	ASCII	2	2	3	8	16	23
	ASCII Long	2	2	4	9	19	28
	Binary	2	2	3	5	9	13
LAN	ASCII	3	4	5	12	24	36
	ASCII Long	3	3	5	13	29	44
	Binary	3	3	5	9	18	26

DC bias test signal current (1.5 V/2.0 V): Output current: Max. 20 mA

Option 001 (Power and DC Bias enhance):

DC bias voltage: DC bias voltage applied to DUT is given as:

Equation 18. $Vdut = Vb - 100 \times Ib$

Vdut [V] DC bias voltage

Vb [V] DC bias setting voltage

Ib [A] DC bias current

DC bias current: DC bias current applied to DUT is given as:

Equation 19. Idut = Vb/(100 + Rdc)

Idut [A] DC bias current

 $\begin{array}{ll} \mbox{Vb [V]} & \mbox{DC bias setting voltage} \\ \mbox{Rdc } [\Omega] & \mbox{DUT's DC resistance} \end{array}$

Maximum DC bias current

Table 64. Maximum DC bias current when the normal measurement can be performed.

Impedance	Bias current isolat	ion		
range $[\Omega]$	range [Ω] ON	OFF		
		Test signal voltage ≤ 2 Vrms	Test signal voltage > 2 Vrms	
0.1	Auto range mode:	20 mA	100 mA	
1	100 mA	20 mA	100 mA	
10	Hold range mode: its values for the range.	20 mA	100 mA	
100		20 mA	100 mA	
300		2 mA	100 mA	
1 k	_	2 mA	20 mA	
3 k	-	200 μΑ	20 mA	
10 k		200 μΑ	2 mA	
30 k	-	20 μΑ	2 mA	
100 k	-	20 μΑ	200 μΑ	

When DC bias is applied to DUT

When DC bias is applied to the DUT, add the following value to the absolute accuracy Ab.

Table 65. Only when Fm < 10 kHz and |Vdc| > 5 V

SHORT	MED, LONG
$0.05\% \times (100 \text{ mV/Vs}) \times (1 + \sqrt{(100/\text{Fm})})$	$0.01\% \times (100 \text{ mV/Vs}) \times (1 + \sqrt{(100/\text{Fm})})$

Fm [Hz] Test frequency
Vs [V] Test signal voltage



Relative measurement accuracy with bias current isolation

When DC bias Isolation is set to ON, add the following value to the open offset Yo.

Equation 20. Yo_DCI1 × (1 + 1/(Vs)) × $(1 + \sqrt{(500/Fm)})$ + Yo_DCI2

 $\begin{array}{ll} {\rm Zm} \, [\Omega] & {\rm Impedance} \ {\rm of} \ {\rm DUT} \\ {\rm Fm} \, [{\rm Hz}] & {\rm Test} \ {\rm frequency} \\ {\rm Vs} \, [{\rm V}] & {\rm Test} \ {\rm signal} \ {\rm voltage} \end{array}$

Yo_DCI1,2 [S] Calculate this by using Table 66 and 67

Idc [A] DC bias isolation current

Table 66. Yo_DCI1 value

DC bias current range	Measurement time mode		
	SHORT	MED, LONG	
20 μΑ	0 S	0 S	
200 μΑ	0.25 nS	0.05 nS	
2 mA	2.5 nS	0.5 nS	
20 mA	25 nS	5 nS	
100 mA	250 nS	50 nS	

Table 67. Yo_DCI2 value

DC bias current	Measurement time mode			
range	≤ 100 Ω	300 Ω, 1 k Ω	3 k Ω, 10 k Ω	30 k Ω, 100 k Ω
20 μΑ	0 S	0 S	0 S	0 S
200 μΑ	0 S	0 S	0 S	0 S
2 mA	0 S	0 S	0 S	3 nS
20 mA	0 S	0 S	30 nS	30 nS
100 mA	0 S	300 nS	300 nS	300 nS

DC bias settling time

When DC bias is set to ON, add the following value to the settling time:

Table 68. DC bias settling time

	Bias	Settling time
1	Standard	Capacitance of DUT \times 100 \times loge (2/1.8 m) + 3 m
2	Option 001	Capacitance of DUT × 100 × loge (40/1.8 m) + 3 m




Figure 12. DC bias settling time



Web Resources

Visit our Web sites for additional product information and literature.

E4980A Precision LCR Meter www.keysight.com/find/e4980a

E4980AL Precision LCR Meter www.keysight.com/find/e4980al


LCR Meters & Impedance Measurement Products www.keysight.com/find/impedance

RF & MW test accessories www.keysight.com/find/mta

Evolving Since 1939

Our unique combination of hardware, software, services, and people can help you reach your next breakthrough. We are unlocking the future of technology. From Hewlett-Packard to Agilent to Keysight.







myKeysight

myKeysight

www.keysight.com/find/mykeysight

A personalized view into the information most relevant to you.

http://www.keysight.com/find/emt_product_registration

Register your products to get up-to-date product information and find warranty information.

KEYSIGHT SERVICES Accelerate Technology Adoption. Lower costs.

Keysight Services

www.keysight.com/find/service

Keysight Services can help from acquisition to renewal across your instrument's lifecycle. Our comprehensive service offerings—onestop calibration, repair, asset management, technology refresh, consulting, training and more—helps you improve product quality and lower costs.



Keysight Assurance Plans

www.keysight.com/find/AssurancePlans

Up to ten years of protection and no budgetary surprises to ensure your instruments are operating to specification, so you can rely on accurate measurements.

Keysight Channel Partners

www.keysight.com/find/channelpartners

Get the best of both worlds: Keysight's measurement expertise and product breadth, combined with channel partner convenience.

www.keysight.com/find/LCRmeters www.keysight.com/find/E4980A www.keysight.com/find/E4980AL For more information on Keysight Technologies' products, applications or services, please contact your local Keysight office. The complete list is available at: www.keysight.com/find/contactus

Americas

Canada	(877) 894 4414
Brazil	55 11 3351 7010
Mexico	001 800 254 2440
United States	(800) 829 4444

Asia Pacific

Australia	1 800 629 485
China	800 810 0189
Hong Kong	800 938 693
India	1 800 11 2626
Japan	0120 (421) 345
Korea	080 769 0800
Malaysia	1 800 888 848
Singapore	1 800 375 8100
Taiwan	0800 047 866
Other AP Countries	(65) 6375 8100

Europe & Middle East

Austria	0800 001122
Belgium	0800 58580
Finland	0800 523252
France	0805 980333
Germany	0800 6270999
Ireland	1800 832700
Israel	1 809 343051
Italy	800 599100
Luxembourg	+32 800 58580
Netherlands	0800 0233200
Russia	8800 5009286
Spain	800 000154
Sweden	0200 882255
Switzerland	0800 805353
	Opt. 1 (DE)
	Opt. 2 (FR)
	Opt. 3 (IT)
United Kingdom	0800 0260637

For other unlisted countries: www.keysight.com/find/contactus (BP-9-7-17)



www.keysight.com/go/quality

Keysight Technologies, Inc. DEKRA Certified ISO 9001:2015 Quality Management System

This information is subject to change without notice. © Keysight Technologies, 2013, 2014, 2016-2018 Published in USA, April 13, 2018 5989-4435EN www.keysight.com



