

## TAIZHOU HONGWEI GLASSES CO.,LTD.

---

### INSPECTION CERTIFICATE NO.: 26003

To whom it may concern:

This is to certify that the following SUNGLASSES (or lenses) are impact resistant and have been passed our test for impact-Resistant on the base of a statistically significant sampling pursuant to Sunglasses Association of American Standard LT-J (July 12, 1971) and 21 CFR 801. 410 (Meet Current F.D.A. Rulings).

The records of our drop ball test are as our attached sheet:

|                               |                                            |
|-------------------------------|--------------------------------------------|
| EXPORTERS:                    | WENZHOU HUIYU INTERNATIONAL TRADE CO., LTD |
| MAMUFACTORY:                  | TAIZHOU HONGWEI GLASSES CO.,LTD.           |
| QUANTITIES:                   | 2364PCS SUNGLASSES                         |
| DATE OF INSPECTION:           | April. 20th, 2026                          |
| INTENDED USE:                 | Samples                                    |
| DESTINATION:                  | USA                                        |
| F.D.A. REG. NO.:              | 886.5850                                   |
| MEDICAL DEVICE LISTING NO.:   | D461120                                    |
| REGISTERED ESTABLISHMENT NO.: | 3006000233                                 |
| INVOICE NO.:                  |                                            |
| OWNER/OPERATER:               | TAIZHOU HONGWEI GLASSES CO.,LTD.           |
| OWNER/OPERATER NO.:           | 10083746                                   |

# TAIZHOU HONGWEI GLASSES CO.,LTD.

## IMPACT RESISTANT DROP BALL TEST

### MILITARY STANDARD-SAMPLING

Procedure and Tables for Inspection by Attributes

Test Records for Inspection Certificate No.: 26003 Date: April 20, 2026

Invoice Number:

(MIL-STD-105)

Table I-II Samples Size and Acceptance Rejection Criteria

| RATCH OR LOT SIZE |         | NORMAL & TIGHTENED |        |    |           |    |                | REDUCED |    |
|-------------------|---------|--------------------|--------|----|-----------|----|----------------|---------|----|
|                   |         | SAMPLE<br>Size     | NORMAL |    | TIGHTENED |    | SAMPLE<br>SIZE | REDUCED |    |
|                   |         |                    | Ac     | Re | Ac        | Re |                | AC      | RE |
| 51                | 90      | 13                 | 2      | 3  | 1         | 2  | 5              | 1*3     |    |
| 91                | 150     | 20                 | 3      | 4  | 3         | 4  | 8              | 1*4     |    |
| 151               | 280     | 32                 | 4      | 5  | 5         | 6  | 13             | 2*6     |    |
| 281               | 500     | 50                 | 5      | 6  | 8         | 9  | 20             | 3*6     |    |
| 500               | 1,200   | 80                 | 10     | 11 | 12        | 13 | 32             | 5*9     |    |
| 1,201             | 3,200   | 125                | 14     | 15 | 18        | 19 | 50             | 7*10    |    |
| 3,201             | 10,000  | 200                | 21     | 22 | 18        | 19 | 80             | 10*13   |    |
| 10,001            | 35,000  | 315                | 21     | 22 | 18        | 19 | 125            | 10*13   |    |
| 35,001            | 150,000 | 500                | 21     | 22 | 18        | 19 | 200            | 10*13   |    |
| 150,001           | 500,000 | 800                | 21     | 22 | 18        | 19 | 215            | 10*13   |    |
| 500,001 and       | Over    | 1,250              | 21     | 22 | 18        | 19 | 500            | 10*13   |    |

Ac=Acceptance number

\*=If the acceptance number has been exceeded but the rejection number has not been reached, accept the lot, but reinstate inspection.

#### IMPACT RESISTANT TEST RECORDS for the WENZHOU HUIYU INTERNATIONAL TRADE CO.,LTD

Sampling Plan NORMAL

(Normal & Tightened or Reduced)

| LENTYPE       | QUANTITY<br>(PCS) | SAMPLE SIZE<br>(PCS) | UNBROKEN<br>(PCS) | BROKEN<br>(PCS) | RESULT OF TEST<br>Accepted Rejected |
|---------------|-------------------|----------------------|-------------------|-----------------|-------------------------------------|
| Glass         | / /               | / /                  | / /               | / /             | / /                                 |
| Polycarbonate | 2364              | 125                  | 125               | ZERO            | ACCEPTED                            |
| Acrylic       | / /               | / /                  | / /               | / /             | / /                                 |
| POLARIZED     | / /               | / /                  | / /               | / /             | / /                                 |

\*\*\*\*\*



