

INSPECTION CERTIFICATE NO.260507

+This is to certify that the following sunglasses (or lenses) are Impact resistant and have passed our for Impact-resistant on the base of a statistically significant sampling pursuant to Sunglass Association of America Standard LT-1 (July'12, 1971) and 21 CFR 801, 410(Meet Current F.D.A. Rulings). The records of our drop ball test are as our attached sheet:

Exporter & Manufacturer: Wenzhou Polytrust Trading Co., Ltd.

Address: Room 903 Block 3 Daziran Commercial Building, Tangjiaqiao Road,
Wenzhou Zhejiang, 325011, China

Tel:0086-577-88933068

Commodities: SUNGLASSES

BATCH includes all lenses used in the manufacture of the following styles:

JX4002-BLK, JX4002-TOR, JX4004-BLK, JX4004-GRY, JX4004-BRW,
JX4005-BLU, JX4005-OLV, JX4005-BLK, JX4005-ATO, JX4006-GRN,
JX4006-BUR, JX4006-RST, JX4006-BLK, JX4007-BLK, JX4007-RED,
JX4007-TOR, JX4009-GRN, JX4009-TOR, JX4009-BLK, JX4010-GRY,
JX4010-GRN, JX4010-BRW, JX4011-BLK, JX4011-TOR

QUANTITY: 2064 PRS (4128 PCS)

Date of Inspection: May, 07.2026

Destination: U.S.A

SHIPPING MARKS:

F.D.A.Reg.No: 3009494404 OWNER/OPERATOR NO.:10038809

F.D.A. MEDICAL DEVICE LISTING NO.D 135392

MILITARY STANDARD-SAMPLING
IMPACT RESISTANCE

Procedures and Tables for Inspection by Attributes

(MIL-STD-105),

st Records for Inspection Certificate No. 260507 DATE: May.07, 2026

Invoice Number WP26BZ016

Table I-II Sample Size and Acceptance Rejection Criteria

BATCH OR LOT SIZE		NORMAL & TIGHTENED						REDUCE		
		SAMPLE		NORMAL		TIGHTENED		SAMPLE		REDUCED
		Size	Ac	Re	Ac	Re	Size	Ac	Re	
51	to	90	13	2	3	1	2	5	1 *	3
91	to	150	20	3	4	2	3	8	1 *	4
151	to	280	32	5	6	3	4	13	2 *	5
281	to	500	50	7	8	5	6	20	3 *	6
501	to	1,200	80	10	11	8	9	32	5 *	8
1,201	to	3,200	125	14	15	12	13	50	7 *	10
3,201	to	10,000	200	21	22	18	19	80	10 *	13
10,001	to	35,000	315	21	22	18	19	125	10 *	13
35,001	to	150,000	500	21	22	18	19	200	10 *	13
150,001	to	500,000	800	21	22	18	19	315	10 *	13
500,001	and over	1,250		21	22	18	19	500	10 *	13

Ac = Acceptance number

Re = Rejection number

* = If the acceptance number has been exceed but the rejection number has not been reached, accept the lot, but reinstate normal Inspection.

IMPACT RESISTANT TEST RECORDS for the WENZHOU POLYTRUST TRADING CO., LTD.

Sampling Plan NORMAL
(Normal & tightened or Reduced)

LENS TYPE	QUANTITY	SAMPLE SIZE	UNBROKEN	BROKEN	RESULT OF TEST	
	(pcs)	(pcs)	(pcs)	(pcs)	Accepted	Rejected
Glass	//////					
Polarized	4128	200	195	5	ACCEPT	
Polycarbonate	//////					
Sunsensor	//////					
CR 39	//////					