

GR&R; Study Report

Gauge Repeatability & Reproducibility Analysis

Study Date:

2026-05-27

Equipment ID:

opA

Characteristic:

grr_analysis

Method:

Xbar-R

Operators:

N/A

Parts:

N/A

Trials:

N/A

Generated By:


Arad Quality Agent

Component	Value	% Contribution
Repeatability (EV)	0.0009	5.2%
Reproducibility (AV)	0.0000	0.0%
Part Variation (PV)	0.0172	99.9%
Total Variation (TV)	0.0172	100%
GR&R %	5.32%	—
NDC	27	—

ACCEPTABLE

%GR&R; = 5.3% — acceptable
NDC = 27
Equipment: opA

Variance Component Breakdown



AIAG MSA Reference Guidelines

Threshold	Verdict	Guideline
< 10%	Acceptable	Suitable for production
10-30%	Conditional	Engineering review required
> 30%	Not Acceptable	Must improve before use
NDC ≥ 5	Adequate	Can distinguish parts