

# Benzene 10 - 100 ppm

Order-No. 86 10 280



## CAUTION

These instructions for use describe the application scope for Dräger MicroTubes. Any use in connection with the Dräger X-act 7000 requires precise knowledge and observance of the instructions for use of the Dräger X-act 7000.

Use Dräger MicroTubes only in connection with the Dräger X-act 7000.  
Store Dräger MicroTubes in their packaging protected from light and dirt (e.g. dust, water).

---

### Application Scope

The determination of benzene in air.

**Measuring range:** 10 to 100 ppm

**Typical measuring time:** 10 to 75 s

**Temperature:** 0 to 35 °C

**Correction:** <20 °C: +3 % from the measured value per °C

**Humidity:** 1 to 30 mg/L

**Accuracy:** ±20 % (under calibration conditions)