



SUBJECT Microbiology Test

TEST LOCATION TÜV SÜD China

TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME

CLIENT ADDRESS

TEST PERIOD 24-Dec-2019~10-Feb-2020

TEST REQUEST Antimicrobial effectiveness testing - with reference to USP42<51>

Prepared By

Nance Gao

(Nance Gao)
Report Drafter



Authorized Signatory

Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested. (3) The test report shall not be reproduced except in full without the written approval of the laboratory. (4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China





RECEIPT DATE / TEST DATE

24-Dec-2019/ 24-Dec-2019

THE FOLLOWING SAMPLE(S) WAS/WERE SUBMITTED

BY/ ON BEHALF OF THE CLIENT'S AS:

Sample Name: Kindee Organic Bamboo Baby Wipe

Sample Specification: 80pcs/bag

Batch No./Date: DATE:28.11.2019

EXP:27.11.2022

Manufacturer:

SAMPLE NO.	DESCRIPTION	PHOTOGRAPH
721651516	Sample in the bag	

TEST METHOD(S)

Antimicrobial effectiveness testing

- with reference to USP 42<51>

TEST ORGANISMS

Staphylococcus aureus ATCC 6538

Escherichia coli ATCC 8739

Pseudomonas aeruginosa ATCC 9027

Candida Albicans ATCC 10231

Aspergillus brasiliensis ATCC 16404

ACCEPTANCE CRITERIA

- Reference to USP 42<51> Table 3. Criteria for tested microorganisms

For Category 2 Products	
Bacteria	Not less than 2.0log reduction from the initial count at 14days,and no increase from the 14days'count at 28days.
Yeast and molds	No increase* from the initial calculated count at 14and 28days.

Note: No increase* is defined as not more than 0.5 log10 unit higher than the previous value measured.





TEST RESULT(S)

Test Item(s)	Log Reduction On			Result
	Initial bacteria concentration	14 th day	28 th day	
<i>Staphylococcus aureus</i> (ATCC 6538)	5.4	>4.4	>4.4	PASS
<i>Escherichia coli</i> (ATCC 8739)	5.4	>4.4	>4.4	PASS
<i>Pseudomonas aeruginosa</i> (ATCC 9027)	5.4	>4.4	>4.4	PASS
<i>Candida Albicans</i> (ATCC 10231)	5.3	>4.3	>4.3	PASS
<i>Aspergillus brasiliensis</i> (ATCC 16404)	5.2	>4.2	>4.2	PASS

Note: This report is for internal use only by the client

-END OF THE TEST REPORT-

