

Antenatal summary

synth01 | Self-recorded at home | 30 Jun to 28 Jul 2026

RESTING PULSE (OPTICAL SENSOR, QUALITY FILTERED)

Average	86 bpm
Range	71 - 111 bpm
Change over period	91 -> 84 bpm (-7)
Sessions / windows	4 / 1063



per-session mean, 80-91 bpm

OXYGEN SATURATION (UNCALIBRATED, TREND ONLY)

Average	97.0 %
Variability (SD)	1.9 points

ENVIRONMENT (AMBIENT AIR, NOT BODY TEMPERATURE)

Air temperature	26.0 - 31.0 C
Humidity	62 - 74 %

Limitations. Pulse and oxygen saturation come from an uncalibrated optical sensor (MAX30100) worn at home; readings below a signal-quality threshold are excluded rather than reported. Temperature and humidity are ambient air measurements from a DHT11 and are NOT body temperature. Nothing in this summary is a diagnosis. It is intended to support, not replace, clinical assessment at an antenatal visit.