


UROFLOWMETRY ANALYSIS REPORT

Test Date: 2025-09-18 13:50
Session ID: example-healthy-male

Uroflowmetry Analysis Report

Flow Rate and Volume Over Time



Key Metrics

Measurement	Value
Voided Volume	643.4 ml
Peak Flow Rate (Qmax)*	36.1 ml/s
Average Flow Rate (Qave)*	20.7 ml/s
Time to Start	2.5 s
Time to Peak	12.0 s
Flow Time	27.5 s
Total Time	31.5 s

*Flow rate measurements use 8-second moving average smoothing with 2-second sustained rule for Qmax calculation.