SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	9
Dist. factor	20 %
Position	R4.0 A23.3 F1.6
Orientation	S > C-3.6
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	5
Dist. factor	80 %
Position	L0.0 A31.9 F4.8
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 3	
Slices	7
Dist. factor	50 %
Position	R2.4 P21.6 F0.6
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
TR	3000 ms
TE	3.22 ms
Averages	1
Concatenations	21
Filter	None
Coil elements	A32

Contrast

TD	0 ms
Magn. preparation	Slice-sel. IR
TI	1100 ms
Flip angle	6 deg
Fat suppr.	None
Water suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	192
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Ascending
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	270.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	2D
Asymmetric echo	Off
Bandwidth	240 Hz/Px
Flow comp.	No
Echo spacing	6.5 ms
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

\\USER\UserProtocols\Renzo\190905_SAMOWU\VASO_139MAGECSSSI_setup

TA: 17:04   PAT: 3   Voxel size: 0.8×0.8×0.8 mm   Rel. SNR: 1.00   USER: VASO_139

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R4.5 A19.9 H27.5
Orientation	T > C-10.7 > S0.7
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	8.3 %
Slices per slab	96
FoV read	133.0 mm
FoV phase	133.3 %
Slice thickness	0.80 mm
TR	8324.60 ms
TE	24 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Perfusion mode	SS-SI VASO
TI2	650 ms
TI1	50 ms
TI1s	50 ms
Flip angle	4 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	123
Delay in TR	0 ms
Multiple series	Off
Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	650.0 ms
Flow limit	100 cm/s

Resolution

Base resolution	162
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	45
Accel. factor 3D	1
Ref. lines 3D	24
Reference scan mode	Separate
Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	25.0 mm
Thickness	100 mm
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R4.8 A15.5 H27.0
! Orientation	T > C-9.9 > S0.6
! Rotation	90.00 deg
! A >> P	190 mm
! R >> L	140 mm
! F >> H	77 mm

Physio

1st Signal/Mode	None

BOLD

Motion correction	Off
Spatial filter	Off

Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1144 Hz/Px
Free echo spacing	Off
Echo spacing	0.98 ms
EPI factor	216

RF pulse type	Normal
Gradient mode	Normal
Excitation	Slab-sel.
RF spoiling	On

Ampl	100
MAGEC FA	12 in deg
ph.skip 4 Robert (the one)	1
MAGEC SS-SI?	On
Maxwell Correction	Off
log physio files	Off
FFT scale	2.50
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2000 us
RF BWTP	25.0
Renzo: Delta TI	71 ms
EFFECTIVE TR	842504 ms
PatPartitions	104
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	162
FlashRef BW	100 Hz/px
FlashRef TE	7000 us
FlashRef FA	5 deg
use CAIPI	Off