

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Affiliate\Renzo_playground\tested\MP2RAGE944_NIPS_0.75iso

TA: 9:38 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: tfl_wip944_b17stx

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	L0.0 A40.9 F42.8
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	7.1 %
Slices per slab	224
FoV read	240 mm
FoV phase	93.8 %
Slice thickness	0.75 mm
TR	6000 ms
TE	2.99 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(3D)
Coil elements	A32

Contrast

Magn. preparation	Non-sel. IR
TI 1	800 ms
TI 2	2700 ms
Flip angle 1	4 deg
Flip angle 2	5 deg
Fat suppr.	None
Water suppr.	None
2nd Inversion Contrast	On
Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	320
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	32
Accel. factor 3D	1
Reference scan mode	Integrated
Image Filter	Off

Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Single shot
Series	Interleaved
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	L3.0 A20.9 F9.5
! Orientation	T > C-37.3
! Rotation	0.00 deg
! R >> L	172 mm
! A >> P	211 mm
! F >> H	45 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Composing

Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	240 Hz/Px
Flow comp.	No
Echo spacing	7 ms
RF pulse type	Fast
Gradient mode	Fast*
Excitation	Non-sel.
RF spoiling	On

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

FFT Scale Factor	150 %
LIN/PAR Swap	Off
Ext. INV Pulse	On
Flip Angle	700
Phase Filter	0 px
Uniform Image	On
Head Mask on UNI	Off
T1 Map	On
Complex Div. Image	Off
Denoise Weighting	150
FLAWS	Off