## SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

\\USER\NIPS\Fukunaga\20190219_GRE\fl_1024_e1_sl24_50Hz_fc						
TA: 10:29	PAT: Off	Voxel size: 0.2×0.2×1.5 mm	Rel. SNR: 1.00	SIEMENS: gre		

Properties		Saturation mode	Standard
Prio Recon	Off	Special sat.	None
Before measurement	<b></b>		
After measurement		Table position	Н
Load to viewer	On	Table position	
Inline movie	Off	Table position	0 mm
	On	Inline Composing	Off
Auto store images	Off	Tim CT mode	Off
Load to stamp segments		Tim OT mede	311
Load images to graphic	Off	System	
segments	0"	V32	Off
Auto open inline display	Off	A32	On
Start measurement without	On	Deside a la seconda	FIV
further preparation		Positioning mode	FIX
Wait for user to start	Off	MSMA	S-C-T
Start measurements	single	Sagittal	R >> L
Routine		Coronal	A >> P
		Transversal	F >> H
Slice group 1 Slices	24	Save uncombined	Off
	24	Coil Combine Mode	Adaptive Combine
Dist. factor	0 %	AutoAlign	<del></del>
Position	L0.0 A25.0 H36.0	Auto Coil Select	Default
Orientation	Transversal		
Phase enc. dir.	R >> L	Shim mode	Standard
Rotation	90.00 deg	Adjust with body coil	Off
Phase oversampling	0 %	Confirm freq. adjustment	Off
FoV read	205 mm	Assume Silicone	Off
FoV phase	87.5 %	? Ref. amplitude 1H	0.000 V
Slice thickness	1.5 mm	Adjustment Tolerance	Auto
TR	700 ms	Adjust volume	
TE	15.0 ms	! Position	L0.0 A25.0 H0.0
Averages	1	! Orientation	Transversal
Concatenations	1	! Rotation	0.00 deg
Filter	None	! R >> L	240 mm
		! A >> P	240 mm
Coil elements	A32	!F>>H	118 mm
Contrast		!г>>п	110111111
MTC	Off	— Physio	
Magn. preparation	None	1st Signal/Mode	None
Flip angle	40 deg	Segments	1
Fat suppr.	None		
Water suppr.	None	Tagging	None
SWI	Off	Dark blood	Off
		·· Doop control	O#
Averaging mode	Short term	Resp. control	Off
Reconstruction	Magn./Phase	Inline	
Measurements	1	Subtract	Off
Multiple series	Each measurement	Liver registration	Off
•		Std-Dev-Sag	Off
Resolution		Std-Dev-Cor	Off
Base resolution	1024	Std-Dev-Tra	Off
Phase resolution	100 %	Std-Dev-Time	Off
Phase partial Fourier	Off	MIP-Sag	Off
Interpolation	Off	MID Con	Off
DAT mede	Nama	MIP-Tra	Off
PAT mode	None		
Image Filter	Off	MIP-Time	Off
Distortion Corr.	Off	Save original images	On
Prescan Normalize	Off	Wash - In	Off
Normalize	Off	Wash - Out	Off
		TTP	Off
B1 filter	Off Off	PEI	_
Raw filter	Off		Off
Elliptical filter	Off	MIP - time	Off
Geometry			
Multi-slice mode	Interleaved	— MapIt	None
Series	Interleaved	Contrasts	1
	· · · · · · · · · · · · · · · · · · ·		

## SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

## Sequence

•	
Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Bandwidth	50 Hz/Px
Flow comp.	Slice
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On
1	