

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Emily\171208\Localizer DLPFC

TA: 1:05

PAT: Off

Voxel size: 1.0x1.0x5.0 mm

Rel. SNR: 1.00

SIEMENS: tfl

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	9
Dist. factor	200 %
Position	R7.2 A19.0 F6.3
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	5
Dist. factor	80 %
Position	R8.2 A20.7 H10.1
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 3	
Slices	7
Dist. factor	200 %
Position	R8.2 A46.4 F8.3
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	3000 ms
TE	3.17 ms
Averages	1
Concatenations	21
Filter	None
Coil elements	A32

Contrast

TD	0 ms
Magn. preparation	Slice-sel. IR
T1	1100 ms
Flip angle	6 deg
Fat suppr.	None
Water suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	192
Phase resolution	100 %
Phase partial Fourier	Off

Interpolation Off

PAT mode None

Image Filter Off

Distortion Corr. Off

Prescan Normalize Off

Normalize Off

B1 filter Off

Raw filter Off

Elliptical filter Off

Geometry

Multi-slice mode Sequential

Series Ascending

Table position H

Table position 0 mm

Inline Composing Off

System

V32 Off

A32 On

Positioning mode REF

MSMA S - C - T

Sagittal R >> L

Coronal A >> P

Transversal F >> H

Save uncombined Off

Coil Combine Mode Adaptive Combine

AutoAlign ---

Auto Coil Select Default

Shim mode Tune up

Adjust with body coil Off

Confirm freq. adjustment Off

Assume Silicone Off

! Ref. amplitude 1H 220.000 V

Adjustment Tolerance Auto

Adjust volume

Position Isocenter

Orientation Transversal

Rotation 0.00 deg

R >> L 350 mm

A >> P 263 mm

F >> H 350 mm

Physio

1st Signal/Mode None

Dark blood Off

Resp. control Off

Inline

Subtract Off

Std-Dev-Sag Off

Std-Dev-Cor Off

Std-Dev-Tra Off

Std-Dev-Time Off

MIP-Sag Off

MIP-Cor Off

MIP-Tra Off

MIP-Time Off

Save original images On

Sequence

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Introduction	On
Dimension	2D
Asymmetric echo	Off
Bandwidth	240 Hz/Px
Flow comp.	No
Echo spacing	6.4 ms
<hr/>	
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Emily\171208\epi_sms3_ip2_2mm_10_20GLM

TA: 6:18 PAT: 2 Voxel size: 2.0x2.0x2.0 mm Rel. SNR: 1.00 USER: ep2d_bold_sms_mgh_v22

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	57
Dist. factor	0 %
Position	L0.0 A19.0 H6.2
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	195 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1000 ms
TE	23 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

MTC	Off
Flip angle	65 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	360
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	98
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Separate
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R6.3 A21.1 H26.4
! Orientation	T > C3.0
! Rotation	0.00 deg
! R >> L	120 mm
! A >> P	162 mm
! F >> H	50 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	On
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	1
Model transition states	Off
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1890 Hz/Px
Free echo spacing	Off
Echo spacing	0.63 ms
<hr/>	
EPI factor	98
RF pulse type	Normal
Gradient mode	Fast
<hr/>	
Dummy Scans	3
Dummy Scans	4
SMS Factor	3
RF Clip	0
VERSE Factor	1.00
SMS Shift	2
Kernel Size	5x5
Compression Factor	1.00

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Emily\171208\FA4_VASO_122_130mmFOV

TA: 11:04 PAT: 3 Voxel size: 0.9x0.9x1.1 mm Rel. SNR: 1.00 UNKNOWN:

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R6.3 A42.8 H29.1
Orientation	T > C24.1
Phase enc. dir.	P >> A
Rotation	180.00 deg
Phase oversampling	0 %
Slice oversampling	9.1 %
Slices per slab	22
FoV read	150.0 mm
FoV phase	100.0 %
Slice thickness	1.10 mm
TR	2006.00 ms
TE	20 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Perfusion mode	SS-SI VASO
TI2	700 ms
TI1	50 ms
TI1s	50 ms
Flip angle	4 deg
Fat suppr.	Fat sat.
Fat sat. mode	Weak
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	331
Delay in TR	0 ms
Multiple series	Off
Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	700.0 ms
Flow limit	100 cm/s

Resolution

Base resolution	162
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Accel. factor 3D	1
Ref. lines 3D	24
Reference scan mode	Separate

Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	25.0 mm
Thickness	100 mm
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	L0.0 A28.9 H28.3
! Orientation	T > C6.9
! Rotation	0.00 deg
! R >> L	120 mm
! A >> P	162 mm
! F >> H	54 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

Motion correction	Off
Spatial filter	Off

Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1144 Hz/Px
Free echo spacing	Off
Echo spacing	0.98 ms
EPI factor	162

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
<hr/>	
Ampl	100
BWDTH	150 3.1kHz
ph.skip 4 Robert (the one)	30
use Ernst angle	Off
Maxwell Correction	Off
log physio files	Off
FFT scale	1.00
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2000 us
RF BWTP	25.0
Renzo: Delta TI	51 ms
EFFECTIVE TR	48144 ms
PatPartitions	24
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	162
FlashRef BW	141 Hz/px
FlashRef TE	4800 us
FlashRef FA	5 deg
use CAIPI	Off

Table of contents

\\USER

UserProtocols

Emily

171208

Localizer DLPFC

epi_sms3_ip2_2mm_10_20GLM

FA4_VASO_122_130mmFOV