SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\RenzoHuber\Anna_trt\print\Checklist_ok

TA: 1:05 PAT: Off Voxel size: 1.0×1.0×5.0 mm Rel. SNR: 1.00 SIEMENS: tfl

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	9
Dist. factor	80 %
Position	L0.0 A27.3 F14.3
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	5
Dist. factor	100 %
Position	R1.1 A28.9 H44.2
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 3	
Slices	7
Dist. factor	50 %
Position	R1.5 A28.2 F10.0
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	3000 ms
TE	2.24 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Magn. preparation	Slice-sel. IR
TI	1100 ms
Flip angle	6 deg
Fat suppr.	None
Water suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	192
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Single shot
Series	Interleaved
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	L0.0 A23.4 F1.3
! Orientation	Transversal
! Rotation	0.00 deg
! R >> L	350 mm
! A >> P	213 mm
! F >> H	189 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On

Dimension	2D
Asymmetric echo	Allowed
Bandwidth	240 Hz/Px
Flow comp.	No
Echo spacing	5.5 ms
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

\\USER\RenzoHuber\Anna_trt\print\mp2rage_iso0.65_iPAT2_angulated

TA: 9:51 PAT: 2 Voxel size: 0.7×0.7×0.7 mm Rel. SNR: 1.00 USER: mp2rage_wip602B

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	L0.6 P0.0 F9.5
Orientation	S > C-1.5 > T-0.9
Phase enc. dir.	A >> P
Rotation	6.75 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	240
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	0.65 mm
TR	4500 ms
TE	2.66 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Magn. preparation	Non-sel. IR
TI 1	900 ms
TI 2	2750 ms
Flip angle 1	5 deg
Flip angle 2	3 deg
Fat suppr.	Water excit. normal
Water suppr.	None
2nd Inversion-Contrast	On
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	320
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Single shot
Series	Interleaved
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	Head > Brain
Auto Coil Select	Off
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	L0.4 P5.4 H17.9
! Orientation	T > C-6.7
! Rotation	-1.75 deg
! R >> L	139 mm
! A >> P	176 mm
! F >> H	117 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off

Asymmetric echo	Allowed
Contrasts	1
Bandwidth	220 Hz/Px
Flow comp.	No
Echo spacing	7.4 ms
RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
FFT Scale Factor	100 %
Line/Partition Swap	Off
Homodyne Phase Filter	Off
Flat Image	On
T1 Map	On
Division Image	Off
ExtInvPulseOn	On
OffResFreqInv	0
Invflipangle	2000

\\USER\RenzoHuber\Anna_trt\print\mp2rage_iso0.65_iPAT2_angulated_multi_echo

TA: 10:57 PAT: 2 Voxel size: 0.7×0.7×0.7 mm Rel. SNR: 1.00 USER: mp2rage_wip602B

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	L0.6 P0.0 F9.5
Orientation	S > C-1.5 > T-0.9
Phase enc. dir.	A >> P
Rotation	6.75 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	240
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	0.65 mm
TR	5000 ms
TE 1	1.9 ms
TE 2	4.15 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Magn. preparation	Non-sel. IR
TI 1	950 ms
TI 2	2750 ms
Flip angle 1	5 deg
Flip angle 2	3 deg
Fat suppr.	Water excit. normal
Water suppr.	None
2nd Inversion-Contrast	On
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	320
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Single shot
Series	Interleaved
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	Head > Brain
Auto Coil Select	Off
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	L0.4 P5.4 H17.9
! Orientation	T > C-6.7
! Rotation	-1.75 deg
! R >> L	139 mm
! A >> P	176 mm
! F >> H	117 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off

Asymmetric echo	Allowed
Contrasts	2
Bandwidth 1	540 Hz/Px
Bandwidth 2	540 Hz/Px
Flow comp.	No
Readout mode	Bipolar
Echo spacing	7.5 ms
RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
FFT Scale Factor	100 %
Line/Partition Swap	Off
Homodyne Phase Filter	Off
Flat Image	On
T1 Map	On
Division Image	Off
Echo Averaging	On
ExtInvPulseOn	On
OffResFreqInv	0
Invflipangle	2000

\\USER\RenzoHuber\Anna_trt\print\VASO_151_0.8mm_120sl_noPO

TA: 16:35 PAT: 3 Voxel size: 0.8×0.8×0.8 mm Rel. SNR: 1.00 USER: VASO_151

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	L4.8 A29.5 H17.9
Orientation	T > C6.4
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	120
FoV read	130.0 mm
FoV phase	133.3 %
Slice thickness	0.80 mm
TR	8964.00 ms
TE	24 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Perfusion mode	SS-SI VASO
TI2	100 ms
TI1	50 ms
TI1s	50 ms
Flip angle	4 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	111
Delay in TR	0 ms
Multiple series	Off
Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	100.0 ms
Flow limit	100 cm/s

Resolution

Base resolution	162
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	45
Accel. factor 3D	1
Ref. lines 3D	24
Reference scan mode	Separate
Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	25.0 mm
Thickness	100 mm
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	225.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R1.9 A31.5 H18.9
! Orientation	T > C-1.6
! Rotation	90.00 deg
! A >> P	174 mm
! R >> L	130 mm
! F >> H	107 mm

Physio

1st Signal/Mode	None

BOLD

Motion correction	Off
Spatial filter	Off

Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1144 Hz/Px
Free echo spacing	Off
Echo spacing	0.98 ms
EPI factor	216

RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
Ampl	110
MAGEC FA	12 in deg
ph.skip 4 Robert (the one)	1
MAGEC SS-SI?	On
Maxwell Correction	Off
log physio files	On
FFT scale	2.00
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2000 us
RF BWTP	25.0
Renzo: Delta TI	71 ms
EFFECTIVE TR	1075680 ms
PatPartitions	120
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	162
FlashRef BW	100 Hz/px
FlashRef TE	7000 us
FlashRef FA	5 deg
use CAIPI	Off