\\USER\SNC\[XT-ID:93-M-0170]Persichetti\sagital_vaso\rslh_ep3d_T1

TA: 9:19 PM: FIX Voxel size: 0.8×0.8×0.9 mmPAT: 3 Rel. SNR: 1.00 : nih5i

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R29.5 A21.3 F11.1 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Slab Scale	-10 %
Slices per slab	36
FoV read	175 mm
FoV phase	100.0 %
Slice thickness	0.85 mm
TR 1	67.9 ms
TR 2	13673 ms
TE 1	23.40 ms
Averages	2
Filter	None
Coil elements	A32

Contrast - Common

TR 1	67.9 ms
TR 2	13673 ms
TE 1	23.40 ms
Multi-echo spacing	62.2 ms
Magn. preparation	Non-sel. HSN IR
TI 1	1732.2 ms
TI 2	4676.6 ms
Flip angle	5 deg
Fat suppr.	Fat sat.
Magn. Prep. Shots	1

Contrast - Dynamic

Averages	2
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	20
Pause after meas. 1	0.0 s
Pause after meas. 2	0.0 s
Pause after meas. 3	0.0 s
Pause after meas. 4	0.0 s
Pause after meas. 5	0.0 s
Pause after meas. 6	0.0 s
Pause after meas. 7	0.0 s
Pause after meas. 8	0.0 s
Pause after meas. 9	0.0 s
Pause after meas. 10	0.0 s
Pause after meas. 11	0.0 s

Contrast - Dynamic

Pause after meas. 12	0.0 s
Pause after meas. 13	0.0 s
Pause after meas. 14	0.0 s
Pause after meas. 15	0.0 s
Pause after meas. 16	0.0 s
Pause after meas. 17	0.0 s
Pause after meas. 18	0.0 s
Pause after meas. 19	0.0 s

Resolution - Common

FoV read	175 mm
FoV phase	100.0 %
Slice thickness	0.85 mm
Base resolution	206
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	CAIPIRINHA
Acc. factor PE	1
Ref. lines PE	80
Acc. factor 3D	3
Ref. lines 3D	36
CAIPI 3D Shift	1
Reference Scan Mode	EPI/separate
CAIPI Mode (tooltip)	Skipped-CAIPI
Total PAT factor	3

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	R29.5 A21.3 F11.1 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slab Scale	-10 %
Slices per slab	36
FoV read	175 mm
FoV phase	100.0 %
Slice thickness	0.85 mm
TR 1	67.9 ms
TR 2	13673 ms

Geometry - AutoAlign

Slab group	1
Position	R29.5 A21.3 F11.1 mm
Orientation	Sagittal

Geometry - AutoAlign

Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R29.5 A21.3 F11.1
R	29.5 mm
A	21.3 mm
F	11.1 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	Fat sat.

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R33.4 A17.4 F11.1 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	175 mm
! F >> H	175 mm
! R >> L	40 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.204291 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	215.000 V

Sequence - Part 1

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1

Sequence - Part 1

Echo spacing	1.18 ms
Bandwidth	934 Hz/Px

Sequence - Part 2

EPI factor	52
Segmentation	3
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
Turbo factor	36

Sequence - Special

PATRef FA	3 deg
RF duration	2000 us
RF BWT product	8
Ernst T1	1200 ms
PATRef prep. shots	10
Volume dummy shots	0
Dummy Measurements	0
ETL per RTEB	1
Invert PE	Off
Min. TE if PF	On
Echo Time Shift	On
Ramp Sampling	On
NORDIC	Off
SVDPC	Off
Sym VASO	On
Dual-pol. EPI	On
Invert RO	On
Invert 3D	Off
Disable PF reco	Off
Disable PF reco	Off
Save sampling	Off
PE VComp	Off
Water Exc.	-none-
External PC	per Series
Saturation RF	per Shot
FIDNavs	-none-
EPI rise time factor	1.10
Mosaic DICOMs	On
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	50000 10^-6
HSN RF power scale	3.00
Inversion Delay	500 ms
Relaxation Delay	600 ms
Var. FA /MAGEC	0

Sequence - Assistant

Mode	Off