## SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

\\USER\UserProtocols\Renzo\Whole\_slice\_template\MP2RAGE\_slab\_0.5x05x0.5\_T1\_POCS8

TA: 16:14 PAT	: 2 Voxel size: 0.5×0.5×		·
Properties		Image Filter	Off
Prio Recon	Off	Distortion Corr.	Off
	Oli		
Before measurement		Prescan Normalize	Off
After measurement		Raw filter	Off
Load to viewer	On	Elliptical filter	Off
Inline movie	Off	Geometry	
Auto store images	On	Multi-slice mode	Single shot
Load to stamp segments	Off	Series	Interleaved
Load images to graphic	Off	Series	inteneaved
segments			
Auto open inline display	Off	Table position	Н
Start measurement without	On	Table position	0 mm
further preparation		Inline Composing	Off
Wait for user to start	Off	Overtone	
Start measurements	single	System	
Start measurements	Sirigie	V32	Off
outine		A32	On
Slab group 1		Positioning mode	EIV
Slabs	1	Positioning mode	FIX
Dist. factor	50 %	MSMA	S-C-T
Position	R29.7 A24.9 H37.7	Sagittal	R >> L
	T > S34.7 > C0.1	Coronal	A >> P
Orientation		Transversal	F >> H
Phase enc. dir.	A >> P	Save uncombined	Off
Rotation	0.00 deg	Coil Combine Mode	Adaptive Combine
Phase oversampling	0 %	AutoAlign	
Slice oversampling	0.0 %	Auto Coil Select	Default
Slices per slab	72		
FoV read	161 mm	Shim mode	Standard
FoV phase	100.0 %	Adjust with body coil	Off
Slice thickness	0.50 mm	Confirm freq. adjustment	Off
TR	6000 ms	Assume Silicone	Off
TE	4.13 ms	! Ref. amplitude 1H	277.000 V
	1	Adjustment Tolerance	Auto
Averages		Adjust volume	Auto
Concatenations	1		D22.2.440.6.1124.2
Filter	None	! Position	R33.3 A18.6 H34.3
Coil elements	A32	! Orientation	T > C-2.3
ontrast		! Rotation	0.00 deg
	Non-sel. IR	! R >> L	83 mm
Magn. preparation		! A >> P	152 mm
TI 1	900 ms	! F >> H	65 mm
TI 2	2900 ms	Dharaia	
Flip angle 1	6 deg	Physio	
Flip angle 2	7 deg	1st Signal/Mode	None
Fat suppr.	None	Dark blood	Off
Water suppr.	None	Dark blood	
2nd Inversion Contrast	On	Resp. control	Off
		' '	
Averaging mode	Long term	Composing	
Reconstruction	Magn./Phase	Soguence	
Measurements	3	Sequence	
Pause after meas. 1	0.0 s	Introduction	On
Pause after meas. 2	0.0 s	Dimension	3D
Multiple series	Each measurement	Elliptical scanning	Off
•		Asymmetric echo	Off
esolution		Contrasts	1
Base resolution	320	Bandwidth	190 Hz/Px
Phase resolution	100 %	Flow comp.	No
Slice resolution	100 %	Echo spacing	8.4 ms
Phase partial Fourier	Off		U.T IIIU
	6/8	RF pulse type	Fast
Slice partial Fourier	0/0	Gradient mode	Fast
PAT mode	GRAPPA	Excitation	Slab-sel.
Accel. factor PE	2	RF spoiling	On
Ref. lines PE	24		— — — — — — — — — — — — — — — — — — —
		FFT Scale Factor	150 %
Accel. factor 3D Reference scan mode	1 Integrated	Morphometry Analysis	Off
- MARANCA CCAN MAAAA	MICHICA		

Integrated

Reference scan mode

## SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

FID MoCo Logging	Off
FID Coil Phase Corr.	Off
LIN/PAR Swap	On
Ext. INV Pulse	On
Flip Angle	1400
Phase Filter	0 px
Uniform Image	On
Head Mask on UNI	Off
T1 Map	On
Complex Div. Image	On
Denoise Weighting	150
FLAWS	Off