SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\151203_Valsalva\MP2RAGE_0.70iso_800_2700_from_Kanny
TA: 10:08 PAT: 3 Voxel size: 0.7×0.7×0.7 mm Rel. SNR: 1.00 USER: tfl_wip900b17a

Properties		Distortion Corr.	On
Prio Recon	Off	Mode Unfiltered images	3D Off
Before measurement		Prescan Normalize	Off
After measurement		Raw filter	Off
Load to viewer	On	Elliptical filter	Off
Inline movie	Off	1 .	5
Auto store images	On	Geometry	
Load to stamp segments	Off	Multi-slice mode	Single shot
Load images to graphic segments	Off	Series	Interleaved
Auto open inline display	Off	Table position	Н
Start measurement without	On	Table position	0 mm
further preparation		Inline Composing	Off
Wait for user to start	Off		On
Start measurements	single	System	
Routine	-	V32 A32	Off On
Slab group 1			
Slabs	1	Positioning mode	REF
Dist. factor	50 %	MSMA	S - C - T
Position	L0.3 A41.3 F37.9	Sagittal	R >> L
Orientation	Sagittal	Coronal	A >> P
Phase enc. dir.	A >> P	Transversal	F >> H
Rotation	9.60 deg	Save uncombined	Off
Phase oversampling	0 %	Coil Combine Mode	Adaptive Combine
Slice oversampling	7.1 %	AutoAlign Auto Coil Select	Default
Slices per slab	224	Auto Coil Select	Default
FoV read	224 mm	Shim mode	Standard
FoV phase	100.0 %	Adjust with body coil	Off
Slice thickness	0.70 mm	Confirm freq. adjustment	Off
TR	6000 ms	Assume Silicone	Off
TE	3.02 ms	! Ref. amplitude 1H	277.000 V
Averages	1	Adjustment Tolerance	Auto
Concatenations Filter	Distantian Comm (OD)	Adjust volume	
Coil elements	Distortion Corr.(3D) A32	! Position	R1.3 A35.6 H4.8
Con elements	A32	! Orientation	T > C-2.0
Contrast		! Rotation	0.00 deg
Magn. preparation	Non-sel. IR		160 mm
TI 1	800 ms	! A >> P	196 mm
TI 2	2700 ms	! F >> H	49 mm
Flip angle 1	4 deg	Physio	
Flip angle 2	5 deg	1st Signal/Mode	None
Fat suppr.	None	David bland	O#
Water suppr. 2nd Inversion Contrast	None On	Dark blood	Off
		Resp. control	Off
Averaging mode Reconstruction	Long term Magn./Phase	Composing	
Measurements	1	Sequence	
Multiple series	Each measurement	Introduction	On
•		Dimension	3D
Resolution	200	Elliptical scanning	Off
Base resolution	320	Asymmetric echo	Off
Phase resolution Slice resolution	100 % 100 %	Contrasts	1
	100 % 6/8	Bandwidth	240 Hz/Px
Phase partial Fourier Slice partial Fourier	6/8	Flow comp.	No
		Echo spacing	7.1 ms
PAT mode	GRAPPA	RF pulse type	Fast
Accel. factor PE	3	Gradient mode	Fast*
Ref. lines PE	32	Excitation	Non-sel.
Accel. factor 3D	1	RF spoiling	On
Reference scan mode	Integrated		
Image Filter	Off	FFT Scale Factor	150 %
. •		Morphometry Analysis	Off

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

FID MoCo Logging	Off
FID Coil Phase Corr.	Off
LIN/PAR Swap	Off
Ext. INV Pulse	On
Flip Angle	700
Phase Filter	0 px
Uniform Image	On
Head Mask on UNI	Off
T1 Map	On
Complex Div. Image	On
Denoise Weighting	150
FLAWS	Off