

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\High_res_template\use_me_12_slices_0.68_3DVASO_ONE_Hemisphere_GRA

TA: 28:12 PAT: 2 Voxel size: 0.7x0.7x1.5 mm Rel. SNR: 1.00 USER: VASO_109

Properties

| | |
|---|--------|
| Prio Recon | Off |
| Before measurement | |
| After measurement | |
| Load to viewer | On |
| Inline movie | Off |
| Auto store images | On |
| Load to stamp segments | Off |
| Load images to graphic segments | Off |
| Auto open inline display | Off |
| Start measurement without further preparation | On |
| Wait for user to start | Off |
| Start measurements | single |

Routine

| | |
|--------------------|-------------------|
| Slab group 1 | |
| Slabs | 1 |
| Dist. factor | 50 % |
| Position | R35.5 A18.0 H44.7 |
| Orientation | T > S38.3 |
| Phase enc. dir. | R >> L |
| Rotation | 60.00 deg |
| Phase oversampling | 0 % |
| Slice oversampling | 0.0 % |
| Slices per slab | 12 |
| FoV read | 32.7 mm |
| FoV phase | 300.0 % |
| Slice thickness | 1.50 mm |
| TR | 1756.70 ms |
| TE | 25 ms |
| Averages | 1 |
| Concatenations | 1 |
| Filter | None |
| Coil elements | A32 |

Contrast

| | |
|----------------------|---------------|
| Perfusion mode | Picore Q2TIPS |
| TI2 | 800 ms |
| TI1 | 50 ms |
| TI1s | 50 ms |
| Flip angle | 29 deg |
| Fat suppr. | Fat sat. |
| Fat sat. mode | Strong |
| Averaging mode | Long term |
| Reconstruction | Magn./Phase |
| Measurements | 963 |
| Delay in TR | 0 ms |
| Multiple series | Off |
| Perfusion mode | PICORE Q2T |
| Inversion time 1 | 50 ms |
| Saturation stop time | 50 ms |
| Inversion time 2 | 800.0 ms |
| Flow limit | 100.0 cm/s |

Resolution

| | |
|-----------------------|-------|
| Base resolution | 46 |
| Phase resolution | 100 % |
| Slice resolution | 100 % |
| Phase partial Fourier | 6/8 |
| Slice partial Fourier | Off |
| Interpolation | Off |

| | |
|---------------------|----------|
| PAT mode | GRAPPA |
| Accel. factor PE | 2 |
| Ref. lines PE | 24 |
| Accel. factor 3D | 1 |
| Ref. lines 3D | 8 |
| Reference scan mode | Separate |

| | |
|-------------------|-----|
| Prescan Normalize | Off |
| Raw filter | Off |
| Elliptical filter | Off |
| Hamming | Off |

Geometry

| | |
|------------------|-------------|
| Multi-slice mode | Interleaved |
| Series | Ascending |
| Special sat. | Parallel F |
| Gap | 25.0 mm |
| Thickness | 100 mm |
| Table position | H |
| Table position | 0 mm |
| Inline Composing | Off |

System

| | |
|--------------------------|-------------------|
| V32 | Off |
| A32 | On |
| Positioning mode | FIX |
| MSMA | S - C - T |
| Sagittal | R >> L |
| Coronal | A >> P |
| Transversal | F >> H |
| Save uncombined | Off |
| Coil Combine Mode | Sum of Squares |
| AutoAlign | --- |
| Auto Coil Select | Default |
| Shim mode | Standard |
| Adjust with body coil | Off |
| Confirm freq. adjustment | On |
| Assume Silicone | Off |
| ! Ref. amplitude 1H | 220.000 V |
| Adjustment Tolerance | Auto |
| Adjust volume | |
| ! Position | R34.0 A18.8 H32.5 |
| ! Orientation | Transversal |
| ! Rotation | 0.00 deg |
| ! R >> L | 89 mm |
| ! A >> P | 64 mm |
| ! F >> H | 58 mm |

Physio

| | |
|-----------------|------|
| 1st Signal/Mode | None |
|-----------------|------|

BOLD

| | |
|-------------------|-----|
| Motion correction | Off |
| Spatial filter | Off |

Sequence

| | |
|-------------------|------------|
| Introduction | On |
| Dimension | 3D |
| Reordering | Linear |
| Contrasts | 1 |
| Bandwidth | 1026 Hz/Px |
| Free echo spacing | Off |
| Echo spacing | 1.1 ms |
| EPI factor | 138 |

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

| | |
|----------------------|--------------|
| RF pulse type | Normal |
| Gradient mode | Normal |
| Excitation | Slab-sel. |
| RF spoiling | On |
| <hr/> | |
| Ampl | 110 |
| BWDTH | 150 3.1kHz |
| thickness | 30 |
| use Ernst angle | Off |
| Maxwell Correction | Off |
| log physio files | Off |
| FFT scale | 1.00 |
| dummy prepscan time | 3 s |
| z shim | 0.00 mT/m*ms |
| RF duration | 1250 us |
| RF BWTP | 25.0 |
| EFFECTIVE TR | 21080 ms |
| PatPartitions | 12 |
| EPI phase correction | local |
| PAT refscan mode | Flash |
| FlashRef BaseRes | 46 |
| FlashRef BW | 1000 Hz/px |
| FlashRef TE | 4800 us |
| FlashRef FA | 5 deg |
| use CAIPI | Off |