

# SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

\\USER\UserProtocols\Emily\use\_me\_next\_time\Localizer DLPFC

TA: 1:05 PAT: Off Voxel size: 1.0x1.0x5.0 mm Rel. SNR: 1.00 SIEMENS: tfl

## Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	single

## Routine

Slice group 1	
Slices	9
Dist. factor	200 %
Position	R7.2 A19.0 F6.3
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	5
Dist. factor	80 %
Position	R8.2 A20.7 H10.1
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 3	
Slices	7
Dist. factor	200 %
Position	R8.2 A46.4 F8.3
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	3000 ms
TE	3.17 ms
Averages	1
Concatenations	21
Filter	None
Coil elements	A32

## Contrast

TD	0 ms
Magn. preparation	Slice-sel. IR
T1	1100 ms
Flip angle	6 deg
Fat suppr.	None
Water suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

## Resolution

Base resolution	192
Phase resolution	100 %
Phase partial Fourier	Off

Interpolation Off

PAT mode None

Image Filter Off  
Distortion Corr. Off  
Prescan Normalize Off  
Normalize Off  
B1 filter Off  
Raw filter Off  
Elliptical filter Off

## Geometry

Multi-slice mode Sequential  
Series Ascending

Table position H  
Table position 0 mm  
Inline Composing Off

## System

V32 Off  
A32 On  
Positioning mode REF  
MSMA S - C - T  
Sagittal R >> L  
Coronal A >> P  
Transversal F >> H  
Save uncombined Off  
Coil Combine Mode Adaptive Combine  
AutoAlign ---  
Auto Coil Select Default  
Shim mode Tune up  
Adjust with body coil Off  
Confirm freq. adjustment Off  
Assume Silicone Off  
! Ref. amplitude 1H 220.000 V  
Adjustment Tolerance Auto  
Adjust volume  
Position Isocenter  
Orientation Transversal  
Rotation 0.00 deg  
R >> L 350 mm  
A >> P 263 mm  
F >> H 350 mm

## Physio

1st Signal/Mode None  
Dark blood Off  
Resp. control Off

## Inline

Subtract Off  
Std-Dev-Sag Off  
Std-Dev-Cor Off  
Std-Dev-Tra Off  
Std-Dev-Time Off  
MIP-Sag Off  
MIP-Cor Off  
MIP-Tra Off  
MIP-Time Off  
Save original images On

## Sequence

## SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

Introduction	On
Dimension	2D
Asymmetric echo	Off
Bandwidth	240 Hz/Px
Flow comp.	No
Echo spacing	6.4 ms
<hr/>	
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

# SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

\\USER\UserProtocols\Emily\use\_me\_next\_time\ANDREA\_worked\_ok

TA: 11:08 PAT: 3 Voxel size: 0.9x0.9x1.1 mm Rel. SNR: 1.00 UNKNOWN:

## Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

## Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R1.9 A26.7 H33.2
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	180.00 deg
Phase oversampling	0 %
Slice oversampling	9.1 %
Slices per slab	22
FoV read	150.0 mm
FoV phase	100.0 %
Slice thickness	1.10 mm
TR	2006.00 ms
TE	20 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

## Contrast

Perfusion mode	SS-SI VASO
TI2	700 ms
TI1	50 ms
TI1s	50 ms
Flip angle	4 deg
Fat suppr.	Fat sat.
Fat sat. mode	Weak

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	333
Delay in TR	0 ms
Multiple series	Off

Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	700.0 ms
Flow limit	100 cm/s

## Resolution

Base resolution	162
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Accel. factor 3D	1
Ref. lines 3D	24
Reference scan mode	Separate

Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

## Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	25.0 mm
Thickness	100 mm

Table position	H
Table position	0 mm
Inline Composing	Off

## System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	L2.4 A26.9 H26.4
! Orientation	T > C0.2
! Rotation	0.00 deg
! R >> L	120 mm
! A >> P	152 mm
! F >> H	50 mm

## Physio

1st Signal/Mode	None
-----------------	------

## BOLD

Motion correction	Off
Spatial filter	Off

## Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1144 Hz/Px
Free echo spacing	Off
Echo spacing	0.98 ms
EPI factor	162

# SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
<hr/>	
Ampl	100
BWDTH	150 3.1kHz
ph.skip 4 Robert (the one)	30
use Ernst angle	Off
Maxwell Correction	Off
log physio files	Off
FFT scale	1.00
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2000 us
RF BWTP	25.0
Renzo: Delta TI	51 ms
EFFECTIVE TR	48144 ms
PatPartitions	24
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	162
FlashRef BW	141 Hz/px
FlashRef TE	4800 us
FlashRef FA	5 deg
use CAIPI	Off

# SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

\\USER\UserProtocols\Emily\use\_me\_next\_time\MP2RAGE\_slab\_0.4x0.4x1.2\_T1

TA: 2:26 PAT: 2 Voxel size: 0.4x0.4x1.2 mm Rel. SNR: 1.00 USER: tfl\_wip900b17a

## Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

## Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R30.1 P0.3 H28.1
Orientation	T > S30.8 > C-11.5
Phase enc. dir.	R >> L
Rotation	60.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	32
FoV read	140 mm
FoV phase	100.0 %
Slice thickness	1.20 mm
TR	6000 ms
TE	4.23 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

## Contrast

Magn. preparation	Non-sel. IR
TI 1	900 ms
TI 2	2900 ms
Flip angle 1	6 deg
Flip angle 2	7 deg
Fat suppr.	None
Water suppr.	None
2nd Inversion Contrast	On
Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Each measurement

## Resolution

Base resolution	320
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	6/8
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated
Image Filter	Off

Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off

## Geometry

Multi-slice mode	Single shot
Series	Interleaved
Table position	H
Table position	0 mm
Inline Composing	Off

## System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	277.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R26.7 A7.7 H20.0
! Orientation	Transversal
! Rotation	0.00 deg
! R >> L	89 mm
! A >> P	64 mm
! F >> H	58 mm

## Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

## Composing

## Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	190 Hz/Px
Flow comp.	No
Echo spacing	8.6 ms
RF pulse type	Fast
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
FFT Scale Factor	150 %
Morphometry Analysis	Off
FID MoCo Logging	Off
FID Coil Phase Corr.	Off

## SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

LIN/PAR Swap	On
Ext. INV Pulse	On
Flip Angle	1400
Phase Filter	0 px
Uniform Image	On
Head Mask on UNI	Off
T1 Map	On
Complex Div. Image	On
Denoise Weighting	150
FLAWS	Off