

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\190903_SAMOWU\Quin_pilot_250_V1

TA: 1:05 PAT: Off Voxel size: 1.0x1.0x3.0 mm Rel. SNR: 1.00 SIEMENS: tfl

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	9
Dist. factor	20 %
Position	R4.0 A23.3 F1.6
Orientation	S > C-3.6
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	5
Dist. factor	80 %
Position	L0.0 A31.9 F4.8
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 3	
Slices	7
Dist. factor	50 %
Position	R2.4 P21.6 F0.6
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
TR	3000 ms
TE	3.22 ms
Averages	1
Concatenations	21
Filter	None
Coil elements	A32

Contrast

TD	0 ms
Magn. preparation	Slice-sel. IR
TI	1100 ms
Flip angle	6 deg
Fat suppr.	None
Water suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	192
Phase resolution	100 %
Phase partial Fourier	Off

Interpolation Off

PAT mode None

Image Filter Off
 Distortion Corr. Off
 Prescan Normalize Off
 Normalize Off
 B1 filter Off
 Raw filter Off
 Elliptical filter Off

Geometry

Multi-slice mode Sequential
 Series Ascending

Table position H
 Table position 0 mm
 Inline Composing Off

System

V32 Off
 A32 On

Positioning mode REF
 MSMA S - C - T
 Sagittal R >> L
 Coronal A >> P
 Transversal F >> H
 Save uncombined Off
 Coil Combine Mode Adaptive Combine
 AutoAlign ---
 Auto Coil Select Default

Shim mode Tune up
 Adjust with body coil Off
 Confirm freq. adjustment Off
 Assume Silicone Off
 ! Ref. amplitude 1H 270.000 V
 Adjustment Tolerance Auto
 Adjust volume
 Position Isocenter
 Orientation Transversal
 Rotation 0.00 deg
 R >> L 350 mm
 A >> P 263 mm
 F >> H 350 mm

Physio

1st Signal/Mode None
 Dark blood Off
 Resp. control Off

Inline

Subtract Off
 Std-Dev-Sag Off
 Std-Dev-Cor Off
 Std-Dev-Tra Off
 Std-Dev-Time Off
 MIP-Sag Off
 MIP-Cor Off
 MIP-Tra Off
 MIP-Time Off
 Save original images On

Sequence

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Introduction	On
Dimension	2D
Asymmetric echo	Off
Bandwidth	240 Hz/Px
Flow comp.	No
Echo spacing	6.5 ms
<hr/>	
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\190903_SAMOWU\26_slices_assym_Version_129_DNM_TR2412

TA: 17:50 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: VASO_124

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R4.1 A21.5 H40.7
Orientation	T > C-4.1 > S0.3
Phase enc. dir.	P >> A
Rotation	180.00 deg
Phase oversampling	0 %
Slice oversampling	7.7 %
Slices per slab	26
FoV read	133.0 mm
FoV phase	133.3 %
Slice thickness	0.82 mm
TR	2736.60 ms
TE	24 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Perfusion mode	SS-SI VASO
TI2	650 ms
TI1	50 ms
TI1s	50 ms
Flip angle	26 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	391
Delay in TR	0 ms
Multiple series	Off

Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	650.0 ms
Flow limit	100 cm/s

Resolution

Base resolution	162
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	45
Accel. factor 3D	1
Ref. lines 3D	22
Reference scan mode	Separate

Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	25.0 mm
Thickness	100 mm

Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R6.4 A21.3 H40.7
! Orientation	T > C-2.9 > S0.2
! Rotation	90.00 deg
! A >> P	178 mm
! R >> L	133 mm
! F >> H	39 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

Motion correction	Off
Spatial filter	Off

Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1188 Hz/Px
Free echo spacing	Off
Echo spacing	0.97 ms
EPI factor	216

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

RF pulse type	Normal
Gradient mode	Normal
Excitation	Slab-sel.
RF spoiling	On
<hr/>	
Ampl	150
BWDTH	150 3.1kHz
ph.skip 4 Robert (the one)	1
use Ernst angle	Off
Maxwell Correction	Off
log physio files	Off
FFT scale	3.00
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2000 us
RF BWTP	25.0
Renzo: Delta TI	72 ms
EFFECTIVE TR	76624 ms
PatPartitions	28
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	162
FlashRef BW	100 Hz/px
FlashRef TE	6500 us
FlashRef FA	5 deg
use CAIPI	Off

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\190903_SAMOWU\RENZO_GRE7_setup_GRAPPA_2

TA: 8:04 PAT: 2 Voxel size: 0.2x0.2x1.0 mm Rel. SNR: 1.00 USER: RENZO_GRE7

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	24
Dist. factor	0 %
Position	R4.1 A21.5 H40.7
Orientation	T > C-4.1 > S0.3
Phase enc. dir.	R >> L
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	192 mm
FoV phase	75.0 %
Slice thickness	1.0 mm
TR	631 ms
TE	15.0 ms
Averages	2
Concatenations	1
Filter	None
Coil elements	A32

Contrast

MTC	Off
Magn. preparation	None
Flip angle	40 deg
Fat suppr.	None
Water suppr.	None
SWI	Off
Averaging mode	Short term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	960
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	45
Reference scan mode	Integrated
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Saturation mode	Standard
Special sat.	None
Table position	H
Table position	0 mm
Inline Composing	Off
Tim CT mode	Off

System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R6.4 A21.3 H40.7
! Orientation	T > C-2.9 > S0.2
! Rotation	90.00 deg
! A >> P	178 mm
! R >> L	133 mm
! F >> H	39 mm

Physio

1st Signal/Mode	None
Segments	1
Tagging	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Liver registration	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On
Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off

Maplt	None
Contrasts	1
Sequence	
Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Bandwidth	70 Hz/Px
Flow comp.	Slice
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\190903_SAMOWU\RENZO_GRE7_setup_noGRAPPA

TA: 7:36 PAT: Off Voxel size: 0.2x0.2x1.0 mm Rel. SNR: 1.00 USER: RENZO_GRE7

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	24
Dist. factor	0 %
Position	R4.1 A21.5 H40.7
Orientation	T > C-4.1 > S0.3
Phase enc. dir.	R >> L
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	192 mm
FoV phase	75.0 %
Slice thickness	1.0 mm
TR	631 ms
TE	15.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

MTC	Off
Magn. preparation	None
Flip angle	40 deg
Fat suppr.	None
Water suppr.	None
SWI	Off
Averaging mode	Short term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	960
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Saturation mode	Standard
Special sat.	None

Table position	H
Table position	0 mm
Inline Composing	Off

Tim CT mode	Off
-------------	-----

System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R6.4 A21.3 H40.7
! Orientation	T > C-2.9 > S0.2
! Rotation	90.00 deg
! A >> P	178 mm
! R >> L	133 mm
! F >> H	39 mm

Physio

1st Signal/Mode	None
Segments	1

Tagging	None
Dark blood	Off

Resp. control	Off
---------------	-----

Inline

Subtract	Off
Liver registration	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off

MapIt	None
Contrasts	1

Sequence

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Bandwidth	70 Hz/Px
Flow comp.	Slice
<hr/>	
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\190903_SAMOWU\26_slices_assym_Version_129_DNM_TR2412

TA: 17:50 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: VASO_124

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R4.1 A21.5 H40.7
Orientation	T > C-4.1 > S0.3
Phase enc. dir.	P >> A
Rotation	180.00 deg
Phase oversampling	0 %
Slice oversampling	7.7 %
Slices per slab	26
FoV read	133.0 mm
FoV phase	133.3 %
Slice thickness	0.82 mm
TR	2736.60 ms
TE	24 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Perfusion mode	SS-SI VASO
TI2	650 ms
TI1	50 ms
TI1s	50 ms
Flip angle	26 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	391
Delay in TR	0 ms
Multiple series	Off

Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	650.0 ms
Flow limit	100 cm/s

Resolution

Base resolution	162
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	45
Accel. factor 3D	1
Ref. lines 3D	22
Reference scan mode	Separate

Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	25.0 mm
Thickness	100 mm
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R6.4 A21.3 H40.7
! Orientation	T > C-2.9 > S0.2
! Rotation	90.00 deg
! A >> P	178 mm
! R >> L	133 mm
! F >> H	39 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

Motion correction	Off
Spatial filter	Off

Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1188 Hz/Px
Free echo spacing	Off
Echo spacing	0.97 ms
EPI factor	216

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

RF pulse type	Normal
Gradient mode	Normal
Excitation	Slab-sel.
RF spoiling	On
<hr/>	
Ampl	150
BWDTH	150 3.1kHz
ph.skip 4 Robert (the one)	1
use Ernst angle	Off
Maxwell Correction	Off
log physio files	Off
FFT scale	3.00
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2000 us
RF BWTP	25.0
Renzo: Delta TI	72 ms
EFFECTIVE TR	76624 ms
PatPartitions	28
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	162
FlashRef BW	100 Hz/px
FlashRef TE	6500 us
FlashRef FA	5 deg
use CAIPI	Off

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\190903_SAMOWU\MP2RAGE_slab_0.5x05x0.5_T1_POCS8

TA: 7:14 PAT: 2 Voxel size: 0.5x0.5x0.5 mm Rel. SNR: 1.00 USER: tfl_wip900b17a

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R4.1 A21.5 H40.7
Orientation	T > C-4.1 > S0.3
Phase enc. dir.	R >> L
Rotation	90.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	48
FoV read	191 mm
FoV phase	100.0 %
Slice thickness	0.50 mm
TR	6000 ms
TE	4.66 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Magn. preparation	Non-sel. IR
TI 1	961 ms
TI 2	2900 ms
Flip angle 1	6 deg
Flip angle 2	7 deg
Fat suppr.	None
Water suppr.	None
2nd Inversion Contrast	On

Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	2
Pause after meas. 1	0.0 s
Multiple series	Each measurement

Resolution

Base resolution	380
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	6/8
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Single shot
Series	Interleaved
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	277.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R6.4 A21.3 H40.7
! Orientation	T > C-2.9 > S0.2
! Rotation	90.00 deg
! A >> P	178 mm
! R >> L	133 mm
! F >> H	39 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Composing

Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	190 Hz/Px
Flow comp.	No
Echo spacing	9.5 ms
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
FFT Scale Factor	150 %
Morphometry Analysis	Off
FID MoCo Logging	Off

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

FID Coil Phase Corr.	Off
LIN/PAR Swap	On
Ext. INV Pulse	On
Flip Angle	1400
Phase Filter	0 px
Uniform Image	On
Head Mask on UNI	Off
T1 Map	On
Complex Div. Image	On
Denoise Weighting	150
FLAWS	Off

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Renzo\190903_SAMOWU\26_slices_assym_Version_129_DNM_TR2412

TA: 17:50 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: VASO_124

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R4.1 A21.5 H40.7
Orientation	T > C-4.1 > S0.3
Phase enc. dir.	P >> A
Rotation	180.00 deg
Phase oversampling	0 %
Slice oversampling	7.7 %
Slices per slab	26
FoV read	133.0 mm
FoV phase	133.3 %
Slice thickness	0.82 mm
TR	2736.60 ms
TE	24 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Perfusion mode	SS-SI VASO
TI2	650 ms
TI1	50 ms
TI1s	50 ms
Flip angle	26 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	391
Delay in TR	0 ms
Multiple series	Off
Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	650.0 ms
Flow limit	100 cm/s

Resolution

Base resolution	162
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	45
Accel. factor 3D	1
Ref. lines 3D	22
Reference scan mode	Separate

Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	25.0 mm
Thickness	100 mm
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R6.4 A21.3 H40.7
! Orientation	T > C-2.9 > S0.2
! Rotation	90.00 deg
! A >> P	178 mm
! R >> L	133 mm
! F >> H	39 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

Motion correction	Off
Spatial filter	Off

Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1188 Hz/Px
Free echo spacing	Off
Echo spacing	0.97 ms
EPI factor	216

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

RF pulse type	Normal
Gradient mode	Normal
Excitation	Slab-sel.
RF spoiling	On
<hr/>	
Ampl	150
BWDTH	150 3.1kHz
ph.skip 4 Robert (the one)	1
use Ernst angle	Off
Maxwell Correction	Off
log physio files	Off
FFT scale	3.00
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2000 us
RF BWTP	25.0
Renzo: Delta TI	72 ms
EFFECTIVE TR	76624 ms
PatPartitions	28
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	162
FlashRef BW	100 Hz/px
FlashRef TE	6500 us
FlashRef FA	5 deg
use CAIPI	Off