

\\MARTINOS DEVELOPER\HUBER\highres_mesoveins_andTOF\20251218_ALL_MESOVEIN\mesovein_WE
_RO1_use_me_princess

TA: 6:27 min Coil Selection: Manual Voxel Size: 0.4x0.4x0.4 mm³ Acc:: 6 Rel. SNR: 1.00

Properties

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	R0.2 A12.1 F8.9 mm
Orientation	T > C-31.0
Phase Encoding Dir.	A >> P
Slices per Slab	352
Phase Oversampling	2.9 %
Slice Oversampling	0.0 %
FOV Read	165 mm
FOV Phase	112.7 %
Slice Thickness	0.35 mm
TR	60.0 ms
Vol. TR	380.16 s
TE 1	11.00 ms
TE 2	27.60 ms
TE 3	44.20 ms
Averages	1
Multi-echo Shots	1
AutoAlign	---
Coil Elements	AC

Contrast - Common

TR	60.0 ms
Vol. TR	380.16 s
TE 1	11.00 ms
TE 2	27.60 ms
TE 3	44.20 ms
Multi-echo spacing	16.60 ms
MTC	Off
Flip Angle	15 deg
Fat-Water Contrast	Water Excitation
Magn. Prep. Shots	1
Contrasts	3

Contrast - Common

Reconstruction	Magn./Phase
----------------	-------------

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Reordering	Linear

Resolution - Common

FOV Read	165 mm
FOV Phase	112.7 %
Slice Thickness	0.35 mm
Base Resolution	466
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	CAIPIRINHA
Reference Scans	GRE/Separate
Acceleration Factor PE	3
Reference Lines PE	24
Acceleration Factor 3D	2
Reference Lines 3D	24
Reordering Shift 3D	1
Phase Partial Fourier	Off
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	R0.2 A12.1 F8.9 mm
Orientation	T > C-31.0
Phase Encoding Dir.	A >> P
Slices per Slab	352
Phase Oversampling	2.9 %
Slice Oversampling	0.0 %
FOV Read	165 mm
FOV Phase	112.7 %
Slice Thickness	0.35 mm
TR	60.0 ms
Vol. TR	380.16 s
Multi-echo Shots	1

Geometry - AutoAlign

Slab Group	1
Position	R0.2 A12.1 F8.9 mm
Orientation	T > C-31.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R0.2 A12.1 F8.9
R	0.2 mm
A	12.1 mm
F	8.9 mm
Initial Orientation	T > C
T > C	-31.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard
-----------------	----------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Manual
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Brain
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

! Position	L1.6 A19.4 F5.0 mm
! Orientation	T > C-31.8 > S0.7
! Rotation	90.37 deg
! R >> L	123 mm
! A >> P	163 mm
! F >> H	87 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	297.117813 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Sequence - Part 1

Sequence Name	vaso a2d6453
Dimension	3D
Excitation	Slab-sel.
RF Pulse Type	Normal
Gradient Mode	Normal
Reordering	Linear
Bandwidth	466 Hz/Px
Echo Spacing	2.83 ms
Turbo Factor	6336
Segmentation	36
EPI Factor	5

Sequence - Part 2

Introduction	On
RF Spoiling	On

Sequence - Special

RF duration	340 us
RF time x BW	12
PAT ref. FA	3 deg
Fat sat. FA	110 deg
T1 (Ernst FA)	1200 ms
Invert PE	Off
Min. TE w/ PF	On
Ramp Sampling	On
Trigger per shot	Off
Noise image	Off
Relax spoilers	Off
Round up Vol. TR	Off
MT flip angle	500 deg
MT off-res.	2000 Hz
MT RF duration	10240 us
Custom Water Exc.	-none-
Phase Correction	per Series
Saturation RF	per Shot
EPI rise time factor	1.20
G. spoil dephasing[1]	0.0 pi
G. spoil dephasing[2]	4.0 pi
G. spoil dephasing[3]	2.0 pi
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	10 / 10^6
Slab Scale	-10 %
RF spoil scheme	Conventional
Read polarity	Regular RO

Sequence - Assistant

SAR Assistant	Off
---------------	-----