Table of contents

\\WIPs\WIP925B Sparse TFL\MP2RAGE\head\AAHead_Scout_32ch-head-coil

TA: 0:14 PM: REF Voxel size: 1.6×1.6×1.6 mmPAT: 3 Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 A30.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.25 ms
TE	1.53 ms
Averages	1
Concatenations	1
Filter	B1 filter
Coil elements	AC

Contrast - Common

TR	3.25 ms
TE	1.53 ms
Flip angle	16 deg

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1

Resolution - iPAT

Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 A30.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.25 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L0.0 A30.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off

System - Miscellaneous

Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	297.222429 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Flip angle	16 deg
Measurements	1
Time to center	6.3 s

Inline - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	16 deg
Measurements	1
Contrasts	1

Inline - MapIt

TR	3.25 ms
TE	1.53 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off

\\WIPs\WIP925B Sparse TFL\MP2RAGE\head\cstfl_wip925b_08mm_iso_sag

TA: 5:58 PM: REF Voxel size: 0.8×0.8×0.8 mmPAT: 1.0 Rel. SNR: 1.00 : WIP_cmp

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	224
FoV read	240 mm
FoV phase	94.4 %
Slice thickness	0.83 mm
TR	6000.0 ms
TE	1.94 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(3D)
Coil elements	AC

Contrast - Common

TR	6000.0 ms
TE	1.94 ms
Magn. preparation	Non-sel. IR
TI 1	800 ms
TI 2	2700 ms
Flip angle 1	4 deg
Flip angle 2	5 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	240 mm
FoV phase	94.4 %
Slice thickness	0.83 mm
Base resolution	288
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off

Resolution - Common

Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	224
FoV read	240 mm
FoV phase	94.4 %
Slice thickness	0.83 mm
TR	6000.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF

System - Miscellaneous

Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	227 mm
F >> H	240 mm
R >> L	186 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	297.222429 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6000.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
TI 1	800 ms
TI 2	2700 ms
Fat suppr.	None
Dark blood	Off
FoV read	240 mm
FoV phase	94.4 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1

Inline - Common

StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle 1	4 deg
Flip angle 2	5 deg
Measurements	1
Contrasts	1
TR	6000.0 ms
TE	1.94 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Single shot
Echo spacing	5.7 ms
Bandwidth	240 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Incr. Gradient spoiling	Off
Turbo factor	224

Sequence - pTX Pulses

Sequence - Special

Sparse Sampling	On
US	4.0 x
Samples/TR	200
Density	0.50
Jitter Radius	1.2
Reference Scan	External
No. Ref-Lines	32
Centric	Off
Virtual Coils	Off
Shift Inv Pulse	0 Hz
No. Iterations	15
CSM RO Resolution	0
Regularisation INV1	0.00100
Regularisation INV2	0.00100

Sequence - Special

Uniform	On
Denoised UNI	On
FLAWS	Off
FLAWS-hc	Off
FLAWS-hc inv.	Off
Division image	Off
Synthetic TI 0	0 ms
Synthetic TI 1	0 ms
Denoise Lambda	1
Scaling	0 10^
Echo Averaging	Off
FID Monitoring	Off

Sequence - Assistant

Mode	Off

\\WIPs\WIP925B Sparse TFL\MP2RAGE\head\cstfl_wip925b_08mm_iso_sag_pTX

TA: 5:58 PM: REF Voxel size: 0.8×0.8×0.8 mmPAT: 1.0 Rel. SNR: 1.00 : WIP_cmp

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	224
FoV read	240 mm
FoV phase	94.4 %
Slice thickness	0.83 mm
TR	6000.0 ms
TE	2.84 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(3D)
Coil elements	AC

Contrast - Common

TR	6000.0 ms
TE	2.84 ms
Magn. preparation	Non-sel. IR
TI 1	800 ms
TI 2	2700 ms
Flip angle 1	4 deg
Flip angle 2	5 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	240 mm
FoV phase	94.4 %
Slice thickness	0.83 mm
Base resolution	288
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off

Resolution - Common

Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	224
FoV read	240 mm
FoV phase	94.4 %
Slice thickness	0.83 mm
TR	6000.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF

System - Miscellaneous

Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	227 mm
F >> H	240 mm
R >> L	186 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	297.222429 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6000.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
TI 1	800 ms
TI 2	2700 ms
Fat suppr.	None
Dark blood	Off
FoV read	240 mm
FoV phase	94.4 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1

Inline - Common

StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle 1	4 deg
Flip angle 2	5 deg
Measurements	1
Contrasts	1
TR	6000.0 ms
TE	2.84 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Single shot
Echo spacing	7.6 ms
Bandwidth	240 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Incr. Gradient spoiling	Off
Turbo factor	224

Sequence - pTX Pulses

pTX Pulse	1
Pulse type	Excitation
pTX Pulse	2
Pulse type	Inversion

Sequence - Special

Sparse Sampling	On
US	4.0 x
Samples/TR	200
Density	0.50
Jitter Radius	1.2
Reference Scan	External
No. Ref-Lines	32
Centric	Off
Virtual Coils	Off
Shift Inv Pulse	0 Hz

Sequence - Special

No. Iterations	15
CSM RO Resolution	0
Regularisation INV1	0.00100
Regularisation INV2	0.00100
Uniform	On
Denoised UNI	On
FLAWS	Off
FLAWS-hc	Off
FLAWS-hc inv.	Off
Division image	Off
Synthetic TI 0	0 ms
Synthetic TI 1	0 ms
Denoise Lambda	1
Scaling	0 10^
Echo Averaging	Off
FID Monitoring	Off

Sequence - Assistant

Mode	Off