

Table of contents			
\\MARTINOS DEVELOPER			
RENZO			
fast_fMRI_Stefano			
first test_loading_protocols_20251126			
localizer			
stefano_protocol			
stefano_protocol_10SI_TR27			
2D_EPI_1sl_			

\\MARTINOS DEVELOPER\RENZO\fast_fmri_Stefano\first test_loading_protocols_20251126\localizer

TA: 10 sec Coil Selection: Auto Voxel Size: 0.5×0.5×7.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	On
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	On
Graphic segment	All Segments
Inline Movie	Off

Routine

Slice Group	1
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	250 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
TR	6.0 ms
TE	2.16 ms
Averages	2
Concatenations	3
AutoAlign	---

Contrast - Common

TR	6.0 ms
TE	2.16 ms
TD	0.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	20 deg

Contrast - Common

Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	1
SWI	Off
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FOV Read	250 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
Base Resolution	256
Phase Resolution	91 %
Interpolation	On

Resolution - Acceleration

Acceleration Mode	None
Deep Resolve	Off
Phase Partial Fourier	Off
Asymmetric Echo	Allowed

Resolution - Filter

Raw Filter	Off
Elliptical Filter	On
Distortion Correction	2D
Normalize	Image Based
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal

Geometry - Common

Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	250 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
TR	6.0 ms
Multi-Slice Mode	Sequential
Series	Interleaved
Concatenations	3

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard
Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up

System - Adjustments

B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slice-sel.
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.181670 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6.0 ms
Segments	1
Concatenations	3

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	250 mm
FOV Phase	100.0 %
Phase Resolution	91 %

Physio - PACE

Resp. Control	Off
Concatenations	3

Inline - Liver

Liver Registration	Off
Save Original Images	On

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Soft Tissue

Wash-in	Off
Wash-out	Off
TTP	Off
PEI	Off
MIP Time	Off
Measurements	1

Inline - Composing**Inline - MapIt**

MapIt	None
Flip Angle	20 deg
Measurements	1
Contrasts	1
TE	2.16 ms
TR	6.0 ms
Save Original Images	On

Sequence - Part 1

Sequence Name	fl
Dimension	2D
Excitation	Slice-sel.
RF Pulse Type	Fast
Gradient Mode	Fast
Flow Compensation	None
Bandwidth	340 Hz/Px
Asymmetric Echo	Allowed
Segments	1

Sequence - Part 2

Introduction	On
RF Spoiling	On
Acoustic noise reduction	Off

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	0 s

\\MARTINOS DEVELOPER\RENZO\fast_fmri_Stefano\first test_loading_protocols_20251126\stefano_protocol

TA: 2 sec Coil Selection: Auto Voxel Size: 3.0×3.0×3.0 mm³ Acc:: 6 Rel. SNR: 1.00

Properties

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	L0.0 P19.7 F23.3 mm
Orientation	T > C-19.7
Phase Encoding Dir.	A >> P
Slices per Slab	8
Phase Oversampling	0.0 %
Slice Oversampling	50.0 %
FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	3.00 mm
TR	21.8 ms
Vol. TR	43.6 ms
TE 1	11.00 ms
Averages	1
Multi-echo Shots	1
AutoAlign	---

Contrast - Common

TR	21.8 ms
Vol. TR	43.6 ms
TE 1	11.00 ms
Multi-echo spacing	20.10 ms
MTC	Off
Flip Angle	15 deg
Fat-Water Contrast	Standard
Magn. Prep. Shots	1
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	4

Contrast - Dynamic

Reordering	Linear
------------	--------

Resolution - Common

FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	3.00 mm
Base Resolution	60
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	CAIPIRINHA
Reference Scans	GRE/Separate
Acceleration Factor PE	1
Reference Lines PE	24
Acceleration Factor 3D	6
Reference Lines 3D	8
Reordering Shift 3D	0
Phase Partial Fourier	Off
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	L0.0 P19.7 F23.3 mm
Orientation	T > C-19.7
Phase Encoding Dir.	A >> P
Slices per Slab	8
Phase Oversampling	0.0 %
Slice Oversampling	50.0 %
FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	3.00 mm
TR	21.8 ms
Vol. TR	43.6 ms
Multi-echo Shots	1

Geometry - AutoAlign

Slab Group	1
Position	L0.0 P19.7 F23.3 mm
Orientation	T > C-19.7

Geometry - AutoAlign

Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 P19.7 F23.3
L	0.0 mm
P	19.7 mm
F	23.3 mm
Initial Orientation	T > C
T > C	-19.70
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard
-----------------	----------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Brain
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 P19.7 F23.3 mm
Orientation	T > C-19.7
Rotation	0.00 deg
A >> P	180 mm
R >> L	180 mm
F >> H	24 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	123.181670 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Sequence - Part 1

Sequence Name	vaso a2d6453
Dimension	3D
Excitation	Slab-sel.
RF Pulse Type	Normal
Gradient Mode	Performance
Reordering	Linear
Bandwidth	5952 Hz/Px
Echo Spacing	0.33 ms
Turbo Factor	2
Segmentation	1
EPI Factor	60

Sequence - Part 2

Introduction	On
RF Spoiling	On

Sequence - Special

RF duration	1000 us
RF time x BW	20
PAT ref. FA	3 deg
Fat sat. FA	110 deg
T1 (Ernst FA)	1200 ms
Invert PE	Off
Min. TE w/ PF	On
Ramp Sampling	On
Trigger per shot	Off
Noise image	Off
Relax spoilers	Off
Round up Vol. TR	Off
MT flip angle	500 deg
MT off-res.	2000 Hz
MT RF duration	10240 us
Custom Water Exc.	-none-
Phase Correction	per Series
Saturation RF	per Shot
EPI rise time factor	1.00
G. spoil dephasing[1]	0.0 pi
G. spoil dephasing[2]	4.0 pi
G. spoil dephasing[3]	2.0 pi
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	5000 /10^6
Slab Scale	0 %
RF spoil scheme	Conventional
Read polarity	Regular RO

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\MARTINOS DEVELOPER\RENZO\fast_fmri_Stefano\first test_loading_protocols_20251126\stefano_protocol_10SI_TR27

TA: 2 sec Coil Selection: Auto Voxel Size: 2.8x2.8x3.0 mm³ Acc:: 5 Rel. SNR: 1.00

Properties

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	L0.6 P10.2 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slices per Slab	10
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	3.00 mm
TR	13.9 ms
Vol. TR	27.8 ms
TE 1	9.00 ms
Averages	1
Multi-echo Shots	1
AutoAlign	---

Contrast - Common

TR	13.9 ms
Vol. TR	27.8 ms
TE 1	9.00 ms
Multi-echo spacing	12.30 ms
MTC	Off
Flip Angle	15 deg
Fat-Water Contrast	Standard
Magn. Prep. Shots	1
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	4

Contrast - Dynamic

Reordering	Linear
------------	--------

Resolution - Common

FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	3.00 mm
Base Resolution	64
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	CAIPIRINHA
Reference Scans	GRE/Separate
Acceleration Factor PE	1
Reference Lines PE	36
Acceleration Factor 3D	5
Reference Lines 3D	8
Reordering Shift 3D	0
Phase Partial Fourier	6/8
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	L0.6 P10.2 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slices per Slab	10
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	3.00 mm
TR	13.9 ms
Vol. TR	27.8 ms
Multi-echo Shots	1

Geometry - AutoAlign

Slab Group	1
Position	L0.6 P10.2 F7.8 mm
Orientation	Transversal

Geometry - AutoAlign

Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.6 P10.2 F7.8
L	0.6 mm
P	10.2 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard
-----------------	----------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Brain
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.6 P10.2 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	180 mm
R >> L	180 mm
F >> H	30 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	123.181670 MHz
? Ref. Amplitude 1H	0.000 V

System - Tx/Rx

Reset	Off
Image Scaling	1.000

Sequence - Part 1

Sequence Name	vaso a2d6453
Dimension	3D
Excitation	Slab-sel.
RF Pulse Type	Normal
Gradient Mode	Performance
Reordering	Linear
Bandwidth	4882 Hz/Px
Echo Spacing	0.25 ms
Turbo Factor	2
Segmentation	1
EPI Factor	48

Sequence - Part 2

Introduction	On
RF Spoiling	On

Sequence - Special

RF duration	1000 us
RF time x BW	20
PAT ref. FA	3 deg
Fat sat. FA	110 deg
T1 (Ernst FA)	1200 ms
Invert PE	Off
Min. TE w/ PF	Off
Ramp Sampling	On
Trigger per shot	Off
Noise image	Off
Relax spoilers	Off
Round up Vol. TR	Off
MT flip angle	500 deg
MT off-res.	2000 Hz
MT RF duration	10240 us
Custom Water Exc.	-none-
Phase Correction	per Series
Saturation RF	per Shot
EPI rise time factor	1.00
G. spoil dephasing[1]	0.0 pi
G. spoil dephasing[2]	4.0 pi
G. spoil dephasing[3]	2.0 pi
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	5000 /10^6
Slab Scale	0 %
RF spoil scheme	Conventional
Read polarity	Regular RO

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\MARTINOS DEVELOPER\RENZO\fast_fmri_Stefano\first test_loading_protocols_20251126\2D_EPI_1sl_
TA: 2:27 min Coil Selection: Auto Voxel Size: 3.0x3.0x3.0 mm ³ Acc:: 3 Rel. SNR: 1.00

Properties

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	1
Distance Factor	50 %
Position	R1.8 P12.6 F26.3 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm
TR	35.0 ms
TE	10.00 ms
Averages	1
Concatenations	1
AutoAlign	---

Contrast - Common

TR	35.0 ms
TE	10.00 ms
MTC	Off
Flip Angle	15 deg
Fat-Water Contrast	Standard
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	4094
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm

Resolution - Common

Base Resolution	64
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	GRE/Separate
Acceleration Factor PE	3
Reference Lines PE	30
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	2D
Static Field Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	1
Distance Factor	50 %
Position	R1.8 P12.6 F26.3 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm
TR	35.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice Group	1
Position	R1.8 P12.6 F26.3 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R1.8 P12.6 F26.3
R	1.8 mm
P	12.6 mm
F	26.3 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R1.8 P12.6 F26.3 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	3 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181670 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	35.0 ms
Log Signals	Off
Concatenations	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0

BOLD

Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	4094
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	6010 Hz/Px
Echo Spacing	0.25 ms
Free Echo Spacing	Off
EPI Factor	64

Sequence - Part 2

Introduction	Off
Ghost Reduction	Off

Sequence - Special

Version	1
Coil Combine Method	EVD PSNPC
Readout oversampling factor	2.00
Kernel Size	3x3

Sequence - Assistant

SAR Assistant	Off
---------------	-----