\\MARTINOS DEVELOPER\HUBER\Olfactory\Princess_20260610\t2_tse_fs_cor_princes_GRAPPA

TA: 4:43 min Coil Selection: Manual Voxel Size: 0.2×0.2×1.0 mm³ Acc:: 2 Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	On
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	34
Distance Factor	10 %
Position	L0.0 A88.6 F30.9 mm
Orientation	C > T2.2
Phase Encoding Dir.	R >> L
Phase Oversampling	20 %
FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	1.0 mm
TR	4540.0 ms
TE	59.00 ms
Averages	2
Concatenations	2
AutoAlign	---
Coil Elements	AC

Contrast - Common

TR	4540.0 ms
TE	59.00 ms
TD	0.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle Mode	Constant
Flip Angle	110 deg
Fat-Water Contrast	SPAIR
Fat Saturation	Strong
Dark Blood	Off
Contrasts	1
Wrap-up Magn.	None
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1

Contrast - Dynamic

Multiple Series	Each Measurement

Resolution - Common

FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	1.0 mm
Base Resolution	384
Phase Resolution	80 %
Interpolation	On

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Self-calibration
Acceleration Factor PE	2
Reference Lines PE	32
Deep Resolve	Off
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	On
Elliptical Filter	Off
Distortion Correction	2D
Normalize	B1 Filter
Image Filter	On

Geometry - Common

Slice Group	1
Slices	34
Distance Factor	10 %
Position	L0.0 A88.6 F30.9 mm
Orientation	C > T2.2
Phase Encoding Dir.	R >> L
Phase Oversampling	20 %
FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	1.0 mm
TR	4540.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A88.6 F30.9 mm
Orientation	C > T2.2
Phase Encoding Dir.	R >> L
AutoAlign	---
Initial Position	L0.0 A88.6 F30.9
L	0.0 mm
A	88.6 mm

Geometry - AutoAlign

F	30.9 mm
Initial Orientation	C > T
C > T	2.20
> S	0.00
Initial Rotation	0.00 deg

Geometry - Navigator

Geometry - Saturation

Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Manual
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Performance

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A88.6 F30.9 mm
Orientation	C > T2.2
Rotation	0.00 deg
R >> L	180 mm
F >> H	180 mm
A >> P	38 mm
Reset	Off

System - pTx

B1 Shim	TrueForm

System - Tx/Rx

Frequency 1H	297.136831 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	4540.0 ms
Concatenations	2

Physio - Cardiac

Fat-Water Contrast	SPAIR
Magn. Preparation	None
Dark Blood	Off
FOV Read	180 mm
FOV Phase	100.0 %
Phase Resolution	80 %
Motion Correction	None

Physio - PACE

Resp. Control	Off
Concatenations	2

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Motion Correction	None
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Sequence - Part 1

Sequence Name	tse
Dimension	2D
RF Pulse Type	Normal
Gradient Mode	Boost*
Flow Compensation	None
Bandwidth	395 Hz/Px
Echo Spacing	8.46 ms
Free Echo Spacing	Off
Define	Turbo Factor
Turbo Factor	13
Echo Trains per Slice	15

Sequence - Part 2

Introduction	On
Incr. Gradient Spoiling	Off
Phase Correction	Off

Sequence - Part 2

Compensate T2 Decay	Off
Fast Mode	Off
WARP	Off
Acoustic noise reduction	Off
Reduce Motion Sens.	On
Motion Correction	None

Sequence - Assistant

SAR Assistant	Flip Angle > TR > RF Pulse Type
Min Flip Angle	110 deg
Max. TR	7000.0 ms
Allowed Delay	30 s
Boost Assistant	On