\\MARTINOS DEVELOPER\HUBER\Shahin_Kaisu\Shahin_20260708_1mm_2mm\1mm_run1

TA: 4:43 min Coil Selection: Manual Voxel Size: 1.0×1.0×1.0 mm³ Acc:: 4 Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	L0.1 P57.7 H1.5 mm
Orientation	C > T-13.9
Phase Encoding Dir.	H >> F
Slices per Slab	60
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
TR	50.0 ms
Vol. TR	3000 ms
TE 1	20.00 ms
Averages	1
Multi-echo Shots	1
AutoAlign	---
Coil Elements	AC

Contrast - Common

TR	50.0 ms
Vol. TR	3000 ms
TE 1	20.00 ms
Multi-echo spacing	42.40 ms
MTC	Off
Flip Angle	18 deg
Fat-Water Contrast	Fast Fat Saturation
Magn. Prep. Shots	1
Contrasts	1
Reconstruction	Magn./Phase

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	88
Reordering	Linear

Resolution - Common

FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
Base Resolution	180
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	CAIPIRINHA
Reference Scans	GRE/Separate
Acceleration Factor PE	2
Reference Lines PE	128
Acceleration Factor 3D	2
Reference Lines 3D	40
Reordering Shift 3D	1
Phase Partial Fourier	7/8
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	L0.1 P57.7 H1.5 mm
Orientation	C > T-13.9
Phase Encoding Dir.	H >> F
Slices per Slab	60
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	180 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
TR	50.0 ms
Vol. TR	3000 ms
Multi-echo Shots	1

Geometry - AutoAlign

Slab Group	1
Position	L0.1 P57.7 H1.5 mm
Orientation	C > T-13.9
Phase Encoding Dir.	H >> F
AutoAlign	---
Initial Position	L0.1 P57.7 H1.5
L	0.1 mm

Geometry - AutoAlign

P	57.7 mm
H	1.5 mm
Initial Orientation	C > T
C > T	-13.90
> S	0.00
Initial Rotation	-90.00 deg

Geometry - Saturation

Saturation Mode	Quick

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Manual
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Brain
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

! Position	L0.0 P59.6 H4.3 mm
! Orientation	C > T-14.8 > S-0.4
! Rotation	-89.92 deg
! F >> H	153 mm
! R >> L	133 mm
! A >> P	60 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	297.136279 MHz
! Ref. Amplitude 1H	250.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Sequence - Part 1

Sequence Name	vaso a2d6453
Dimension	3D
Excitation	Slab-sel.
RF Pulse Type	Normal
Gradient Mode	Normal
Reordering	Linear
Bandwidth	1112 Hz/Px
Echo Spacing	1.04 ms
Turbo Factor	60
Segmentation	2
EPI Factor	40

Sequence - Part 2

Introduction	Off
RF Spoiling	On

Sequence - Special

RF duration	800 us
RF time x BW	10
PAT ref. FA	3 deg
Fat sat. FA	110 deg
T1 (Ernst FA)	1200 ms
Invert PE	Off
Min. TE w/ PF	On
Ramp Sampling	On
Trigger per shot	Off
Noise image	Off
Relax spoilers	On
Round up Vol. TR	Off
MT flip angle	500 deg
MT off-res.	2000 Hz
MT RF duration	10240 us
Custom Water Exc.	-none-
Phase Correction	per Series
Saturation RF	per Shot
EPI rise time factor	1.20
G. spoil dephasing[1]	0.0 pi
G. spoil dephasing[2]	4.0 pi
G. spoil dephasing[3]	2.0 pi
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	100 /10^6
Slab Scale	-10 %
RF spoil scheme	Conventional
Read polarity	Inverted RO

Sequence - Assistant

SAR Assistant	Off