On Approval

Tableau comparatif des résultats avant et après modification A1078						
	Historique RVX0183_QP1	**Historique RVX0183_QP2**	**Historique RVX0183_QP3**	**Après modification A1078 QP1 02787AE-4Ar**	**Après modification A1078 QP2 02787AE-5Ar**	**Après modification A1078 QP3 02787AE-6Ar**
HCT (µg/Article)	52	45	48	<33*	<33*	<33*
COT (µg/Article)	30	42	58	<33*	<33*	<33*

* inférieur à la limite de détection

On Approval

Ce document contient des informations confidentielles qui sont la propriété de B. Braun Medical. Toute diffusion, distribution ou copie de ce document est strictement interdite sans le consentement écrit préalable de B. Braun Medical. (raugpa16)

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: David, Anthony (davianfr)
Title: SA-FR02-M-Industrialization Manager
Date: Tuesday, 06 October 2026, 07:47 W. Europe Daylight Time
Meaning: Document signed as Author
==========

UserName: Mergault, Coralie (mergcofr)
Title:
Date: Tuesday, 06 October 2026, 09:11 W. Europe Daylight Time
Meaning: Approve Document
==========

UserName: Bourgeois, Clement (bourclfr)
Title: SA-FR02-M-SMI Manager
Date: Wednesday, 07 October 2026, 17:34 W. Europe Daylight Time
Meaning: Approve Document
==========

On Approval

On Approval