B|BRAUN

NOTE TO THE FILE

Date: *01/09/2026*

PURPOSE: **Cleaning process controls description**

Author: Ophélie TRAINSON

Title: IN420_Cleaning validation pieces_Re-Cleaning _Rationale

1- INTRODUCTION


We plan to conduct a final cleaning process qualification (OQ & PQ) campaign for the following cleaning processes:

Reference	Designation	Intermediate cleaning	Final cleaning
A1101	COUPELLE STANDARD	**PQ cleaning** by suppliers	**OQ or PQ cleaning** by B.Braun medical SAS according to FP205_Ed12_I_02, 03 or 04
A1110	COUPELLE BABYPORT		
A1266	SORTIE G 14,5 MM		
A1525	TUNNELISEUR CERTON / BABYPORT		

<ak>AK002</ak>
<abstract>
The large quantity of pieces required for the qualification process justifies evaluating the possibility of cleaning the parts used in the operational qualification (OQ) and performance qualification (PQ, *i.e.* routine parameters) processes according to the current routine parameters for inclusion in kits intended for sale.

To this end, we plan to clean the available components after the OQ and PQ tests using the routine process conforming to FP205_Annexe 6_Ed14. </abstract>

The pieces returned to production will have undergone the cleaning presented in the following scheme:



It is essential to ensure that the sequence of the cleaning cycles will not have any impact on the safety and performance of the pieces.

The goal of this note is to assess the risks associated with the reuse of pieces and to decide on the reuse of pieces used within the framework of the OQ and PQ.

As part of the pieces cleaning process validation, the present note states the feasibility to use previously cleaned pieces for routine production after a routine re-cleaning based on the risk assessment described below.

2- METHOD

Scope: