Certificate Of Processing

Prepared for **GE HEALTHCARE**

STERIS

Gamma Process Run ID **129708B**

Product Code	Product Lot Number	Quantity	UOM
Cytiva (25-40 kGy)	17242818	19	CS
Cust Item ID: 29-1846-12			
Cytiva (25-40 kGy)	17269079	1	CS
Cust Item ID: CB0050L10-21			
Cytiva (25-40 kGy)	17271868	2	CS
Cust Item ID: CB0001L722-01-05PK			
Cytiva (25-40 kGy)	17271872	5	CS
Cust Item ID: CB0002L11-01			
Cytiva (25-40 kGy)	17271875	4	CS
Cust Item ID: CB0002L10MDY01-2			
Cytiva (25-40 kGy)	29JAN2021	3	CS
Cust Item ID: 699968LT			
Cytiva (25-40 kGy)	29JANUARY2021	1	CS
Cust Item ID: 699968SF			
Cytiva (25-40 kGy)	MBO28JAN21CGT	6	CS
Cust Item ID: RDMBOMRT			
Cytiva (25-40 kGy)	MBO29JAN21CGT	10	CS

Processing Run Start Date/Time: 04-Feb-2021 12:00:00 am

Processing Run End Date/Time: 04-Feb-2021 01:27:00 am

Approx. Downtime (hours): 0.00

Minimum Specified Dose (kGy): **25.0**

Minimum Delivered Dose (kGy): **26.4**

Maximum Specified Dose (kGy): **40.0**

Maximum Delivered Dose (kGy): **36.7**

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By

Camara, Mariame (QS/RC Technician)

Document Content Revision: 1

Signed On 2/4/2021 at 9:48 AM

UTC / GMT Offset (hh:mm): -5:00

Processing Location:

STERIS
435 Whitney Street
Northborough, MA 01532
Phone: 508-393-9323
Fax: 844-698-9776

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with the applicable standard, EN ANSI/AAMI/ISO 11137 or EN ANSI/AAMI/ISO 11135. For items processed with gamma irradiation, STERIS certifies that these items received the indicated doses within the precision and accuracy of the dosimetry system used.

090177e19abdfceb\Approved\Approved On: 13-Jul-2022 20:47 (GMT)