

業日誌を作成することを要しない。

- (viii) The following work logs (limited to those for which a column for describing responses to accidents and abnormalities is provided) are prepared in accordance with the procedures for entry in work logs provided in the standard operation manual listed in the left column of Appended Table 1-3. However, in the case of a place which carries out serum separation only, it is not required to prepare the work logs set forth in (c) and (f), and in the case of a place which does not carry out serum separation, it is not required to prepare the work log set forth in (d).

イ 検体受領作業日誌

(a) The work log for specimen reception

ロ 検体搬送作業日誌

(b) The work log for specimen transportation

ハ 検体受付及び仕分作業日誌

(c) The work log for acceptance and sorting of specimens

ニ 血清分離作業日誌

(d) The work log for serum separation

ホ 検査機器保守管理作業日誌

(e) The work log for maintenance and management of examination equipment

ヘ 測定作業日誌

(f) The work log for measurement

- 九 別表第一の三の上欄に掲げる標準作業書に記載された台帳の記入要領に従い、次に掲げる台帳が作成されていること。ただし、血清分離のみを行う場所にあつては、ロからトまで及びヌに掲げる台帳を作成することを要しない。

- (ix) The following ledgers are prepared in accordance with the procedures for entry in ledgers provided in the standard operation manual listed in the left column of Appended Table 1-3. However, in the case of a place which carries out serum separation only, it is not required to prepare the ledgers set forth in (b) through (g) and (j).

イ 委託検査管理台帳

(a) The ledger for entrusted examination management

ロ 試薬管理台帳

(b) The reagent management ledger

ハ 温度・設備管理台帳

(c) The ledger for temperature and equipment management

ニ 統計学的精度管理台帳

(d) The statistical quality control ledger

ホ 外部精度管理台帳

(e) The external quality control ledger

ヘ 検体保管・返却・廃棄処理台帳