

#### 4. OPTIONS

<i>Item</i>	<i>Requirement</i>	Compliance of the Equipment with the concerned requirement(s) Yes/No?	Comments and/or additional information regarding the compliance of the Equipment with the concerned requirement(s)
<b>I</b>	Ability of the equipment to produce de-ionised water at three resistivities namely higher than 0.1, 1 and 10 MΩ cm <sup>-1</sup> .		
<b>II</b>	Ability of the equipment to test cells and stack at 50 bar pressure.		
<b>III</b>	Ability of the equipment to test AAEM cell/stack with diluted caustic solution (e.g. 0.1 M KOH) including pH control in the recirculation loop.		
<b>IV</b>	On-site operator training (on JRC request during stage 2) – 3 days		
<b>V</b>	On-site operator training (on JRC request during stage 2) - 2 days.		

**Options are qualitative or quantitative extras, ancillary to the main purchase, and which are optional for the contracting authority - it has the right to buy them or not – but not to the tenderers, who have to include them in their technical and financial tender. Failure to do so will lead to tender rejection.**

For specifications, standards, brand names, processes or formats named in these specifications, unless otherwise specified, equivalent objects shall be accepted. It shall be for the contractor to justify this equivalence.