



The Application of Esophageal Pressure Measurement in Patients with Respiratory Failure

Evangelia Akoumianaki¹, Salvatore M. Maggiore², Franco Valenza³, Giacomo Bellani⁴, Amal Jubran⁵, Stephen H. Loring⁶, Paolo Pelosi⁷, Daniel Talmor⁶, Salvatore Grasso⁸, Davide Chiumello⁹, Claude Guérin¹⁰, Nicolo Patroniti⁴, V. Marco Ranieri¹¹, Luciano Gattinoni¹², Stefano Nava¹³, Pietro-Paolo Terragni¹¹, Antonio Pesenti⁴, Martin Tobin⁵, Jordi Mancebo¹⁴, and Laurent Brochard¹⁵

¹Department of Intensive Care Medicine, University Hospital of Heraklion, Heraklion, Crete, Greece; ²Department of Anesthesiology and Intensive Care, Agostino Gemelli Hospital, Università Cattolica del Sacro Cuore, Rome, Italy; ³Department of Pathophysiology and Transplantation, Università degli Studi di Milano, Fondazione Istituto di Ricovero e Cura a Carattere Scientifico (IRCCS) Ca' Granda-Ospedale Maggiore Policlinico, Milan, Italy; ⁴Department of Health Science, University of Milan-Bicocca and Department of Emergency Medicine and Intensive Care, San Gerardo Hospital, Monza, Italy; ⁵Division of Pulmonary and Critical Care Medicine, Edward Hines, Jr. Veterans Administration Hospital, Hines and Loyola University of Chicago Stritch School of Medicine, Maywood, Illinois; ⁶Department of Anesthesia, Critical Care, and Pain Medicine, Beth Israel Deaconess Medical Center and Harvard Medical School, Boston, Massachusetts; ⁷Department of Surgical Sciences and Integrated Diagnostics, IRCCS San Martino Istituto Nazionale per la Ricerca sul Cancro, University of Genoa, Genoa, Italy; ⁸Dipartimento dell'Emergenza e dei Trapianti di Organi, Sezione di Anestesiologia e Rianimazione, Università degli Studi di Bari Aldo Moro, Bari, Italy; ⁹Dipartimento di Anestesia, Rianimazione, e Terapia del Dolore, Ospedale Maggiore Policlinico, Milan, Italy; ¹⁰Hôpital de la Croix-Rousse, Hospices Civils de Lyon, Université de Lyon, and Institut National de la Santé et de la Recherche Médicale 1044, Lyon, France; ¹¹Department of Anesthesia and Intensive Care Medicine, University of Turin, and S. Giovanni Battista Molinette Hospital, Turin, Italy; ¹²Dipartimento di Anestesia, Rianimazione, e Emergenza Urgenza, Fondazione IRCCS Ca' Granda-Ospedale Maggiore Policlinico and Dipartimento di Fisiopatologia Medico-Chirurgica e dei Trapianti, Università degli Studi di Milano, Milan, Italy; ¹³Respiratory and Critical Care, Sant'Orsola-Malpighi Hospital, Alma Mater Studiorum University of Bologna and Department of Specialist, Diagnostic, and Experimental Medicine, Bologna, Italy; ¹⁴Servei de Medicina Intensiva, Hospital de Sant Pau, Barcelona, Spain; ¹⁵Critical Care Department, St. Michael's Hospital, Keenan Research Centre and Interdepartmental Division of Critical Care, University of Toronto, Toronto, Ontario, Canada

Abstract

This report summarizes current physiological and technical knowledge on esophageal pressure (Pes) measurements in patients receiving mechanical ventilation. The respiratory changes in Pes are representative of changes in pleural pressure. The difference between airway pressure (Paw) and Pes is a valid estimate of transpulmonary pressure. Pes helps determine what fraction of Paw is applied to overcome lung and chest wall elastance. Pes is usually measured via a catheter with an air-filled thin-walled latex balloon inserted nasally or orally. To validate Pes measurement, a dynamic occlusion test measures the ratio of change in Pes to change in Paw during inspiratory efforts against a closed airway. A ratio close to unity indicates that the system provides a valid measurement. Provided transpulmonary pressure is the lung-distending pressure, and that chest wall elastance may vary among individuals, a physiologically based ventilator strategy should take the

transpulmonary pressure into account. For monitoring purposes, clinicians rely mostly on Paw and flow waveforms. However, these measurements may mask profound patient-ventilator asynchrony and do not allow respiratory muscle effort assessment. Pes also permits the measurement of transmural vascular pressures during both passive and active breathing. Pes measurements have enhanced our understanding of the pathophysiology of acute lung injury, patient-ventilator interaction, and weaning failure. The use of Pes for positive end-expiratory pressure titration may help improve oxygenation and compliance. Pes measurements make it feasible to individualize the level of muscle effort during mechanical ventilation and weaning. The time is now right to apply the knowledge obtained with Pes to improve the management of critically ill and ventilator-dependent patients.

Keywords: pleural pressure; respiratory mechanics; mechanical ventilation

In 1949, Buytendijk first showed that it was possible to use esophageal pressure as a surrogate for pleural pressure (1). In 1952, Dornhorst and Leathart showed that changes

in pleural and esophageal pressures were similar and useful to understand respiratory mechanics (2). Soon thereafter, Cherniack and colleagues confirmed that changes in pleural

pressure were similar to changes in esophageal pressure, although the absolute values of pressures in the pleural space were often more negative than in the esophagus (3). These

(Received in original form December 14, 2013; accepted in final form January 16, 2014)

Correspondence and requests for reprints should be addressed to Laurent Brochard, M.D., Saint Michael's Hospital, 30 Bond Street, Toronto, ON, M5B 1W8 Canada. E-mail: brochardl@smh.ca

This article has an online supplement, which is accessible from this issue's table of contents online at www.atsjournals.org

CME will be available for this article at <http://www.atsjournals.org>

Am J Respir Crit Care Med Vol 189, Iss 5, pp 520–531, Mar 1, 2014

Copyright © 2014 by the American Thoracic Society

Originally Published in Press as DOI: 10.1164/rccm.201312-2193CI on January 27, 2014

Internet address: www.atsjournals.org

findings indicated that measurements of esophageal pressure could provide an estimate of pleural pressure; such measurements have enhanced substantially our knowledge regarding the mechanical properties of the lungs, the chest wall, and the whole respiratory system. In addition, esophageal pressure measurements (and derived parameters such as work of breathing) have advanced our understanding of the pathophysiological mechanisms of acute respiratory failure and ventilator dependency. Surprisingly, however, these measurements have been used mostly in the field of research. Recent studies showing the usefulness of esophageal pressure measurements in ventilator management during acute lung injury, optimizing patient-ventilator interaction and ventilator weaning, have drawn the attention of clinicians and researchers to this “old” technique (4–7).

Despite data showing its usefulness in critically ill patients, esophageal pressure is still hardly used in the clinical arena. This is partially due to technical issues, such as the insertion and proper placement of an esophageal catheter, the feasibility of obtaining accurate measurements, and the interpretation of the measurements. For these reasons, a working group called PLUG (Pleural Pressure Working Group; see the online supplement) gathered together for a 1-day conference to summarize current knowledge on esophageal pressure measurements and to suggest ways in which these measurements could be used in critically ill patients. This review is focused on three main areas: (1) the physiological background of esophageal pressure measurement, (2) the clinical indications, and (3) the description of the technique.

Physiological Background

The Forces

Mechanical work is performed when a force moves its point of application through a distance. In pulmonary physiology, work is done when a pressure (expressed in cm H₂O) changes the volume (expressed in liters) of the system. The driving force for breathing is the intrathoracic pressure generated by the contraction of the respiratory muscles (active conditions), the ventilator substituting for respiratory muscles (passive conditions) or by both the ventilator and the respiratory muscles (assisted ventilation). The various trans-structural pressures involved in overcoming

the loads imposed by the different respiratory structures (lungs, chest wall, and respiratory system) are listed in Table 1 and graphically shown in Figure 1.

During mechanical ventilation, the total pressure applied to the respiratory system (P_{total}) is the sum of the pressure provided by the ventilator (P_{aw}) and the pressure generated by the patient's inspiratory muscles (P_{mus}):

$$P_{total} = P_{aw} + P_{mus} \quad (1)$$

The total pressure applied to the respiratory system must overcome the opposing forces produced by the elastic and resistive properties of the respiratory system. This relationship is described in the equation of motion:

$$P_{total} = P_{aw} + P_{mus} = P_0 + E_{rs} \cdot V + R_{rs} \cdot \dot{V} \quad (2)$$

where P_0 is the value of P_{aw} at the beginning of the breath (zero or a positive value of end-expiratory pressure), E_{rs} is the respiratory system elastance, R_{rs} is the respiratory system resistance, V is the volume difference between the instantaneous volume and the relaxation volume of the respiratory system, and \dot{V} is airflow. At the end of a controlled breath using constant flow insufflation, Equation 2 can be applied using delivered tidal volume for V and set inspiratory (peak) flow for \dot{V} . The values of R_{rs} and E_{rs} are readily measured during passive mechanical ventilation using the end-inspiratory (and end-expiratory) occlusion technique (8).

Of note, the complete equation of motion also includes a third component describing pressure changes in phase with acceleration. This part can be reasonably neglected at conventional frequencies and gas densities, but becomes important under conditions of high-frequency ventilation. Last, in applying Equation 2, one must remember that the system is not linear, and at the extremes of lung volume, lung stresses may exceed those estimated from elastance.

Esophageal Pressure as a Surrogate of Pleural Pressure

In upright subjects, pleural pressure (P_{pl}) has been estimated by measuring esophageal pressure (P_{es}), using an esophageal balloon-catheter system. The respiratory changes in P_{es} are representative of changes in P_{pl} applied to the lung surface (9). The difference between P_{aw} and P_{es} is a valid estimate of transpulmonary pressure

(P_L) in the region surrounding the balloon catheter (9). Absolute values of P_{es} can be influenced by respiratory mechanics, lung volume, weight of the mediastinum, the abdomen, posture, reactivity of the esophageal smooth muscle wall, and mechanical properties of the balloon. Our knowledge about the impact of position, asymmetry of lung disease, lung and chest wall distortion (10), increased abdominal pressure, and large pleural effusion (11) on the observed P_{es} and its respiratory variation is limited. Some data suggest, however, that, even under these conditions, P_{es} remains an acceptable effective average P_{pl} (10).

As a consequence a debate exists about whether the absolute values of P_{es} can be interpreted as reliable absolute values of P_{pl} (12, 13). Data suggest that the use of absolute values of P_{es} is possible in the clinical setting (6, 14), but this approach needs further validation.

Pleural pressure varies within the pleural space because of both gravitational gradients and regional inhomogeneities. Furthermore, diseases that obstruct flow, increase the density of lung tissue, or stiffen the lungs could increase the interregional differences in P_{pl} . For the sake of simplicity, in this review we are assuming that P_{pl} is uniform throughout the pleural space. P_{es} is the most convenient way to measure P_{pl} .

Table 1: Trans-structural Pressures Related to Ventilation

Pressure difference across the respiratory system (P_{rs}):

$$P_{rs} = P_{aw} - P_{bs}$$

Pressure difference across the lung (transpulmonary pressure, P_L):

$$P_L = P_{aw} - P_{pl}$$

Pressure difference across the chest wall (P_{cw}):

$$P_{cw} = P_{pl} - P_{bs}$$

Alternatively:

$$\begin{aligned} P_{rs} &= P_L + P_{cw} \\ &= P_{aw} - P_{pl} + P_{pl} - P_{bs} \\ &= P_{aw} - P_{bs} \end{aligned}$$

Definition of abbreviations: P_{aw} = airway pressure, measured by the ventilator at the proximal end of the airways; P_{bs} = body surface pressure; P_{pl} = pleural pressure.




Figure 1. Schematic representation of the relevant pressures for the respiratory system. Pbs= pressure at body surface; P_{CW} = pressure difference across the chest wall; P_L = transpulmonary pressure; Prs = pressure difference across the respiratory system.

Passive Inflation of the Thorax

Under passive conditions, the pressure applied to move gas into the lung is delivered by the ventilator and is equal to P_{aw}. Because P_{mus} is zero, the equation of motion can be rewritten as follows:

$$P_{aw} = P_0 + (E_{rs} \cdot V) + (R_{rs} \cdot \dot{V}) \quad (3)$$

As E_{rs} is the sum of E_{cw} and E_L:

$$P_{aw} = P_0 + (E_{cw} \cdot V) + (E_L \cdot V) + (R_{rs} \cdot \dot{V}) \quad (4)$$

P_{aw}, V, and \dot{V} are continuously measured by the ventilator. Under conditions of no flow, such as during end-inspiratory or end-expiratory occlusion maneuvers, the resistive pressure component ($R_{rs} \cdot \dot{V}$) of Equations 3 and 4 vanishes and the unknown variables of Equations 3 and 4 are E_{cw} and E_L (P_{cw} and P_L change per unit of volume). The measurement of P_{es} is the only way to distinguish what fraction of P_{aw} is applied to overcome lung and chest wall elastance. It should be clarified that P_L incorporates the pressure difference across the airways and the alveoli. The difference between alveolar pressure and pleural pressure is the transalveolar pressure. In the absence of flow (e.g., during an end-inspiratory occlusion maneuver to obtain the plateau pressure or an end-expiratory occlusion to measure total and auto- or intrinsic positive end-expiratory pressure [15]), and provided that there is no significant airway closure, the P_{aw} measured by the ventilator is equal to the pressure inside the alveoli.

Active Inflation of the Thorax

We refer to “active” conditions when the respiratory muscles are working,

irrespective of the work done by the ventilator. In patients with spontaneous breathing efforts, P_{mus} becomes a significant component of the equation of motion (Equation 2).

In this condition, the P_{aw} displayed by the ventilator poorly reflects the total distending pressures of the lungs, and the measurement of pleural pressure or P_{es} is needed for an accurate P_L estimation. One may consider that the pressure applied to distend the lungs is the sum of a visible component (displayed ventilator pressure, i.e., airway pressure) and an invisible component (P_{es}). Figure 2 illustrates the possible differences between volume control and pressure control regarding the effects of spontaneous breathing on P_L.

In the presence of spontaneous breathing efforts while receiving mechanical ventilation, direct measurement of the level of effort may help the clinician to better adjust the ventilator settings and/or the sedation level. Respiratory muscle effort can be assessed by calculating work of breathing (WOB) and the pressure–time product (PTP) of the esophageal pressure (PTP_{es}), reflecting the effort done by all of the respiratory muscles, or the pressure–time product of the transdiaphragmatic pressure (PTP_{di}), reflecting mostly the effort done by the diaphragm. Transdiaphragmatic pressure (P_{di}) is calculated as the difference between the gastric pressure and P_{es}.

Measuring WOB or PTP is a useful approach to estimate the energy dissipated or consumed by the respiratory muscles (16). Work is expressed as force × displacement. In physiology, the work performed during each respiratory cycle

(from the beginning of inspiratory flow, t₀, to the end of inspiration, T_i, is expressed as the area enclosed in a pressure (P)–volume (V) diagram:

$$WOB = \int_{t_0}^{T_i} P V dt$$

In a spontaneously breathing patient, work measurement requires an estimate of P_{pl}, and P_{es} provides an accurate estimate. P_{es} can be viewed as the static recoil pressure of the relaxed chest wall (P_{cw,rel}) minus the inspiratory pressure developed by the inspiratory muscles in expanding the chest wall (P_{mus}). As lung volume increases, P_{cw,rel} and the respiratory muscles generate a negative pressure relative to this value. P_{mus} can be expressed as

$$P_{mus} = P_{cw,rel} - P_{es} \quad (5)$$

The work performed by respiratory muscles (W_{mus}) equals the integral of the product of P_{mus} and the change in volume:

$$W_{mus} = \int P_{mus} \cdot dV \quad (6)$$

By combining Equations 5 and 6, the following is obtained:

$$W_{mus} = \int (P_{cw,rel} - P_{es}) \cdot dV \quad (7)$$

The dynamic relation between P_{mus} and lung volume during breathing can be expressed graphically using the Campbell diagram (17).

Respiratory muscle activity can also be quantified using PTP_{es}. Like WOB, it is based on the estimation of P_{mus} but it refers to the integral of pressure over time and not over volume. PTP_{es} is therefore the product of the pressure developed by the respiratory muscles multiplied by the time of muscle contraction, expressed in units of cm H₂O × second. It can be used whether or not volume is generated. When volume is generated, WOB and PTP_{es} are usually tightly correlated.

The WOB per breathing cycle is normally expressed in joules. Work per minute is calculated by multiplying the WOB per cycle by the corresponding respiratory frequency. Work per liter is calculated by dividing work per minute by minute ventilation. One joule is the work needed to move 1 L of air across a 10–cm H₂O pressure difference (i.e., the surface enclosed in a rectangle with a base of 10 cm H₂O and a height of 1 L).




Figure 2. Comparison of volume-assist control and pressure-assist control ventilation with recordings of tidal volume (V_T) and airway and esophageal pressure (P_{aw} and P_{es} , respectively) waveforms. The red arrows illustrate the difference between airway and esophageal pressure, that is, transpulmonary pressure or P_L . P_L increases with the patient's effort only during pressure control.

Excellent correlations between WOB and oxygen consumed by, or blood flow to, the respiratory muscles have been shown under experimental and clinical conditions (18–20). Measurements of WOB, however, can sometimes underestimate oxygen consumption by the respiratory muscles. In particular, measurement of mechanical work is totally insensitive to energy expenditure during isometric contraction (21). In addition, mechanical work does not account well for the duration of muscular contraction. PTPes potentially circumvents these problems. It has been shown that PTPdi, under specific experimental conditions, is more closely related to respiratory muscle oxygen consumption than WOB (22).

Few studies have been conducted in healthy subjects to determine the normal range of WOB and muscular effort. During spontaneous breathing, WOB ranged from 2.4 to 7.5 J · minute⁻¹ and from 0.20 to 0.9 J · L⁻¹ (23–25). Normal values of PTPes range from 50 to 150 cm H₂O · second · minute⁻¹ (average, 86 ± 21 cm H₂O · s · min⁻¹) (25). The respiratory work performed by patients receiving partial ventilator assistance can become considerably higher than normal. For example, when patients receive an intermittent mandatory ventilation rate of 10 breaths/minute or pressure support of 7 cm H₂O, common settings in clinical practice, inspiratory PTPes can exceed 200 cm H₂O · second · minute⁻¹, which was double the value recorded in healthy subjects (26).

In the clinical setting, measurement of inspiratory effort may be monitored as the simple changes of P_{es} during inspiration, that is, not taking into account the static recoil pressure of the relaxed chest wall. It is much less precise than measurements of P_{mus} or P_{di} but it may be used as a bedside monitoring tool, as done in some sleep studies or during a weaning trial (5, 27).

Clinical Use of Esophageal Pressure: Passive Conditions

Interpretation of Pressures Displayed by the Ventilator

In sedated and paralyzed subjects, positive pressure ventilation is titrated on the basis of P_{aw} , with the expectation that it closely approximates P_L . In patients with normal E_{cw} , P_{aw} is a reasonable surrogate for P_L . When E_{cw} is high, however, P_{aw} may be significantly higher than P_L . Indeed, a portion of P_{aw} is dissipated in distending the chest wall. As the chest wall becomes stiffer, the proportion of P_{aw} that distends the lung (P_L) decreases progressively. E_{cw} can be elevated in patients with acute respiratory failure for various reasons (28). Increases in E_{cw} and/or in P_{pl} can occur as a result of intraabdominal hypertension, pleural effusion, massive ascites, thoracic trauma, and edema of the intrathoracic and intraabdominal tissues as a result of fluid resuscitation (29–32). In pigs, Mutoh and colleagues have shown that intravascular volume infusion produces abdominal distension, lung volume restriction, and

stiffening of the chest wall, which, in turn, resulted in increases in pleural pressure, E_{cw} , and E_L (33). In a swine model of acute respiratory distress syndrome (ARDS) caused by oleic acid infusion, Quintel and colleagues showed that increasing abdominal pressure by insufflation of air into the peritoneal cavity increased P_{es} , decreased lung volume, and markedly increased edema of intrathoracic tissues (34). These studies suggest that similar mechanisms (intravascular volume infusion, insufflation of air into the peritoneal cavity, tissue edema, etc.) may lead to increased E_{cw} and increased P_{pl} in patients, especially in the context of ARDS. Moreover, several investigators have reported the influence of increased E_{cw} on the pressure–volume curve of the respiratory system both in terms of elastance value and of shape of the curve (30, 35, 36).

Because E_{cw} may vary greatly among individuals, adjusting the ventilator settings only on the basis of P_{aw} may not be a satisfactory strategy when ventilating subjects with ARDS. In fact, a positive pressure breath may injure the lung if it results in end-inspiratory alveolar hyperinflation and/or cyclical alveolar opening and collapse (37). Provided P_L is the real “lung-distending” pressure, that is, the main force that promotes alveolar recruitment and lung inflation, a lung-protective ventilator strategy should take this concept (P_L) into account. Esophageal pressure measurement makes this possible in clinical practice. Figure 3 illustrates three

different clinical situations with different contributions of the chest wall.

Esophageal Pressure to Guide Therapy in ARDS

The usefulness of P_{es} in guiding therapy in ARDS has been shown in the Esophageal Pressure-Directed Ventilation (EPVent) study (6). Because of reduced chest wall compliance, edema, or abdominal distension, P_{es} is often elevated in patients with ARDS, and the calculated P_L can be negative at end expiration. This may indicate closed airways, or flooded or atelectatic lung. Positive end-expiratory pressure (PEEP) could thus be increased until P_L becomes positive at end expiration to keep airways open (with the caveat that positive values do not ensure open alveoli in the zones distal to the sampling catheter). In their single-center, randomized controlled trial, the EPVent investigators compared mechanical ventilation guided by P_{es} measurements (experimental arm) with ventilation based on the protocol of the U.S. National Health Institutes-sponsored ARDSNetwork (control arm) (38). Patients in the control arm were treated with tidal volume set at 6 ml/kg of predicted body weight and PEEP based on the patient's partial pressure of arterial oxygen (P_{aO_2}) and inspired fraction of oxygen (F_{iO_2}). In the experimental arm, PEEP levels were set to achieve a P_L between 0 and 10 cm H₂O at end expiration, according to a sliding scale based on the P_{aO_2}/F_{iO_2} ratio. They also limited tidal volume to keep P_L at less than 25 cm H₂O at end inspiration. At 72 hours, PEEP was on average 18 ± 5 cm H₂O in the experimental arm and 12 ± 5 cm H₂O in the control arm. The study was terminated early, after enrolling 61 patients, for an overwhelming effect of the P_{es} strategy on blood oxygenation. At 72 hours, the P_{aO_2}/F_{iO_2} ratio was 280 ± 126 mm Hg in the P_{es} arm and 191 ± 71 in the control arm ($P = 0.001$). Respiratory system compliance was also significantly improved in the P_{es} group ($P = 0.005$), probably as a consequence of improved recruitment. Although this trial showed a trend toward reduced 28-day mortality (17 vs. 35%; $P = 0.055$), it was not sufficiently powered to show significant change in any outcome variable such as ventilator-free days, length of stay, duration of ventilation, or long-term clinical status. Nevertheless, this research may be considered proof of




Figure 3. Three different clinical situations with different contributions of the chest wall to the pressure generated by the ventilator (P_{aw}) during passive inflation. Despite different levels of P_{aw} , the end-inspiratory P_L is almost similar in the three patients. From top to bottom: tracings of flow, volume, and esophageal, airway, and transpulmonary pressure. ARDS = acute respiratory distress syndrome; P_{aw} = airway pressure; P_{es} = esophageal pressure; P_L = transpulmonary pressure.

concept for the usefulness of Pes measurements in ARDS.

Other investigators have used an elastance-derived method to estimate Ppl, which neglects the absolute values and relies on the tidal Pes swings (ΔP_{es}) to calculate Ecw, that is, $Ecw = \Delta P_{cw}/V_T$ (tidal volume) (39). This method takes into account the lung-distending pressure applied by positive pressure inflation during mechanical ventilation. Because any positive pressure applied at the airway opening acts on two elastic structures connected in series (the lung and the chest wall), Paw is distributed between chest wall and lung elastance. The ratio of lung to respiratory system elastance can be used to better interpret the effect of Paw. Accordingly,

$$P_L = P_{aw} \cdot (E_L/E_{rs})$$

in which $E_L = \Delta P_L/V_T$ and $E_{rs} = \Delta P_{aw}/V_T$ under static conditions (zero flow) between the beginning and the end of a breathing cycle.

The latter method for partitioning lung and chest wall elastance has been used to guide a transpulmonary “open lung” approach in a cohort of patients with severe ARDS related to influenza A-H1N1 (4). This assessment helped clinicians to decide, in severely hypoxemic patients requiring high Paw pressures, whether it was appropriate to further increase pressures on the ventilator or whether an extracorporeal oxygenation technique was preferable.

Although further studies are needed to test alternative methods of calculating P_L , the results from such studies support the use of Pes measurement in sedated and paralyzed subjects in titrating ventilator settings in ARDS.

The Esophageal Catheter in the Operating Room

Data suggest that the ventilatory settings during surgical procedures can have important clinical consequences on postoperative complications (40, 41). General anesthesia or selected surgical procedures such as peritoneal insufflation of gas or positioning can affect the mechanics of the chest wall (42). Obese patients or patients with increased abdominal pressure, scoliosis, spondylitis, fibrothorax, or pleural effusion also have altered chest wall mechanics (43). In all of these circumstances, monitoring of Paw during or after surgery may not be representative

of the distending pressure applied on lung tissue (i.e., the stress) and the resulting deformation (i.e., the strain). Although not proven, monitoring of Pes could be helpful to better customize ventilator settings in all these circumstances.

Understanding the Filling Pressures of the Heart

Because the heart is in the chest, the correct interpretation of intracardiac pressures as filling pressures for the ventricles needs to refer the values of absolute intravascular pressures to the changes in extracardiac pressure or pericardial pressure. This is particularly relevant during changes in lung volume under positive airway pressure. Mean Pes is the more convenient technique to estimate extramural pressure and therefore the transmural filling pressures, that is, the intravascular minus the surrounding extravascular pressure (44, 45). A correct interpretation of the cardiac function curve, also known as the Frank–Starling relation, must take into account the transmural filling pressure, not the intravascular pressure (46). Consequently, for similar intravascular pressures, the transmural filling pressures are usually lower when patients are ventilated with positive pressure ventilation as compared with unassisted spontaneous breathing.

Clinical Use of Esophageal Pressure: Assisted, or Patient-triggered, Mechanical Ventilation

General Issues with Assisted, or Patient-triggered, Ventilation

During spontaneous breathing Pes can be used to assess respiratory muscle effort and WOB generated by the patient. We now summarize some of the clinical scenarios in which this measure may be useful.

There is a complex interplay between the assistance delivered by the ventilator and the motor activity of the respiratory system. For monitoring patient–ventilator interactions, clinicians rely mostly on pulse oximetry, arterial blood gas values, and Paw–flow waveforms that are available on most ventilators. Several clinical research studies have demonstrated how difficult it is to determine the amount of effort in standard ventilatory modes (18, 47–50) and

the presence of asynchrony (26, 51, 52). Relying on measurements of Paw and flow may mask profound asynchrony between the patient and the ventilator (52–54).

Although unloading respiratory muscle activity and avoiding excessive respiratory muscle work constitute a major objective of mechanical ventilation, this has surprisingly not been monitored until recently (55). Monitoring Pes activity offers the potential to monitor patient–ventilator interactions.

Monitoring of Respiratory Muscle Activity and Synchrony during Assisted Ventilation

- When the ventilator does not detect any “patient triggering” and the patient is highly sedated, the presence of respiratory muscle contractions triggered by the ventilator has been described in critically ill patients (56). This phenomenon, known as respiratory entrainment (57), was called *reverse triggering* because the insufflation triggers the respiratory muscle contraction (56). These efforts become evident if respiratory muscle activity is monitored through the measurement of Pes (see Figure 4). This may have important clinical consequences such as double inspiration, tidal volume increase, or erroneous plateau pressure measurements.
- Pressure-preset or pressure-targeted ventilation modes work with the ability to synchronize pressure delivery with patient inspiratory efforts. When lung-protective ventilation is desirable, clinicians should be aware that inspiratory synchronization may amplify patient effort. Such a synchronization may be potentially harmful by increasing P_L and V_T values (58). When lung protection is considered a priority but a certain degree of spontaneous breathing activity is maintained, monitoring of Pes may be the best way to ensure proper delivery of inspiratory assistance.
- During assisted ventilation, the combination of slightly excessive pressures and volumes to assist the patient, as well as excessive ventilator inspiratory time relative to a patient’s neural inspiratory time and to some degree of airway obstruction, may lead to the occurrence of ineffective or missed efforts (26, 51, 52, 59). During the presence of these wasted efforts, the real respiratory rate of the patient can be




Figure 4. Effects of intermittent spontaneous breathing efforts indicated by negative esophageal pressure swings during a bilevel mode of ventilation. The airway pressure tracing alone does not allow one to understand what the patient is doing. From top to bottom: tracings of flow, volume, esophageal pressure, and airway pressure. Paw = airway pressure; Pes = esophageal pressure; V_T = tidal volume.

twice the rate displayed on the ventilator (7, 60). Monitoring Pes or the electrical activity of the diaphragm is helpful in recognizing and treating the cause of ineffective efforts (7). Because this asynchrony is associated with prolonged duration of mechanical ventilation, this may potentially impact the duration of mechanical ventilation. In injured lungs, relatively strong negative Ppl generated by diaphragm contraction may have regional effects in dependent regions that are not uniformly transmitted, causing pendelluft (61). In such cases, the risks associated with large P_l swings may even be underestimated by Pes.

- Short cycles result from a shorter mechanical inspiratory time than the patient's neural inspiratory time (52, 62). The combination of high respiratory drive, high flow rate, and low tidal volumes makes this problem frequent. A real-time online monitoring of respiratory muscle activity, such as Pes or diaphragmatic electromyographic activity, is important to detect the degree of synchronization between the patient's

inspiratory efforts and the insufflation time of the ventilator.

Measurement of Auto- or Intrinsic PEEP

For lung volume to increase in a patient with intrinsic PEEP (PEEPi), the inspiratory muscles must contract and generate an amount of pressure equal to the dynamic component of total PEEP, also referred to as PEEPi, before any volume is displaced. The most accurate method to quantify PEEPi is to measure the drop in esophageal pressure at end expiration at the point of the contraction of the inspiratory muscles until inspiratory flow starts (63). Although expiration normally occurs passively, the coexistence of PEEPi and active expiration is common, especially in patients with chronic obstructive pulmonary disease (64). Positive expiratory swings in gastric pressure are observed during active expiration as a consequence of abdominal muscle recruitment. In this scenario, when the patient starts contracting the inspiratory muscles, the expiratory muscles also start to relax. The drop in

esophageal pressure used to estimate PEEPi is therefore due to the relaxation of the expiratory muscles. To avoid overestimating the value of PEEPi, the abdominal pressure swing resulting from the active expiration should be subtracted from the initial drop in esophageal pressure as first proposed by Lessard and colleagues (64).

Weaning from Mechanical Ventilation

Measurement of WOB can be a useful monitoring tool during a weaning trial (65). Research into weaning pathophysiology has revealed that respiratory effort changes progressively as patients fail a weaning trial (66–69). Over the course of a spontaneous breathing trial, PTPes remained unchanged in weaning success patients (66). In contrast, weaning failure patients developed marked and progressive increase in PTPes as a result of an increase in the mechanical load on the respiratory muscles. By the end of the trial, the weaning failure patients increased their PTPes to more than four times the normal value (66). Over the course of a failed weaning trial, swings in

Pes showed larger changes than did the rapid shallow breathing index (66); accordingly, Pes measurements may provide a simple method for monitoring changes in patient effort. Jubran and colleagues (5) showed that looking at trends in Pes (the Pes trend index) over a weaning trial might be more helpful than spot measurements in weaning prediction. If confirmed, measurement of the Pes trend index could provide a useful clinical tool for patient assessment during weaning. In addition, increases in Pes swings during the trial could alert a physician to search for possible causes and to institute therapy, such as bronchodilators, inotropes, vasodilators, or diuretics.

In addition to its monitoring role, measuring Pes during spontaneous breathing helps in our understanding of the hemodynamic changes that occur during difficult weaning (67, 70). Specifically, the large negative swings in Pes that occur

during difficult weaning account for the increases in venous return into the pulmonary circulation and in left ventricular afterload caused by the increased transmural intrathoracic pressures (71).

Technique

The Balloon Catheter

Pes can be measured using catheters with air-filled or liquid-filled balloons (mainly in neonates) (2), or with small transducers placed in the esophagus (72). The most common technique is to use a catheter with a thin-walled latex balloon sealed at its distal port and filled with air (3). The Pes signal is transmitted through the catheter and is measured at its proximal end by a pressure transducer.

Each type of esophageal balloon (depending on its diameter and length) requires the injection of a specific amount

of air (ranging from 0.5 to 4 ml) to measure Pes correctly (73). The accuracy of Pes monitoring with an esophageal balloon depends on the volume injected (74, 75). For the same amount of air inside the balloon, the measured Pes is significantly higher with balloons of narrower diameter and shorter length than with larger and longer balloons (74, 75). A longer balloon is preferable because it can record pressure from a larger area of the esophageal wall, thus better reflecting Ppl along its length (74, 76). To optimize signal transmission to pressure transducers, the balloon should have high compliance. Balloons that are commonly used are 5 to 10 cm long, 0.01 to 0.18 mm thick, and 3.2 to 4.8 cm in diameter (74, 77). It has been shown that the presence of the nasogastric tube does not invalidate the accuracy of Pes measurements provided that the catheter is correctly positioned (78). Naso- or orogastric tubes




Figure 5. Pressure waveforms during insertion of an esophageal catheter. Paw = airway pressure; Pes = esophageal pressure; Pg = gastric pressure; PS = pressure support.

equipped with an esophageal balloon are now available (79): these devices are recommended if measurements of P_{es} are required for a long period. These tubes can also be equipped with two balloons for measuring esophageal and gastric pressures.

The Procedure for Catheter Placement

After placing the patient in a semirecumbent position and anesthetizing the nose and oropharynx, the catheter is inserted through the nostril. The empty balloon catheter is advanced into the stomach, at which time the balloon is inflated, usually with 0.5 ml of air (this volume can, however, vary with the characteristics of the system [80, 81]). The distal part of the catheter is connected to a pressure transducer, which, in turn, can be connected to a dedicated acquisition system, a patient-monitoring system or an auxiliary pressure port of the ventilator. The presence of a positive pressure deflection during a spontaneous inspiration generally indicates that the balloon is in the stomach, provided that there is no diaphragmatic paralysis. Subsequently the catheter is slowly withdrawn until a negative pressure deflection replaces the positive deflection, indicating that the balloon is in the lower third of esophagus (82, 83) (Figure 5). The dynamic occlusion

test is then performed (*see below*) (77). Reassessment of the correct volume of air in the balloon and the control of its proper positioning are particularly important to ensure reliable measurements of P_{es} over a prolonged period of time.

Validation of Esophageal Pressure Measurement: The Dynamic Occlusion Test

In a spontaneously breathing patient, the classic method to validate the P_{es} measurement is the dynamic occlusion test. It consists of measuring the ratio of change in esophageal pressure to the change in airway opening pressure ($\Delta P_{es}/\Delta P_{aw}$ ratio) during three to five spontaneous respiratory efforts against a closed airway. A $\Delta P_{es}/\Delta P_{aw}$ ratio close to unity indicates that the balloon provides a valid measure of P_{pl} changes (77) (Figure 6). This test does not require patient cooperation. The occlusion test has been validated in normal adults and pediatric patients; it has also been applied in paralyzed subjects (84–87). In sedated and paralyzed patients, the occlusion test is performed by applying manual compression on the chest during airway occlusion (84–86) (Figure 6).

Factors influencing the $\Delta P_{es}/\Delta P_{aw}$ ratio during the occlusion test include the position of the balloon, the amount of air

injected into the balloon, the patient's position, and lung volume. These factors should be checked periodically to ensure the best concordance between swings in P_{es} and in P_{aw} during the occlusion test. The acceptable range of $\Delta P_{es}/\Delta P_{aw}$ ratio during the occlusion test is 10–20% (i.e., from 0.8 to 1.2) (85, 86, 88). Cardiac contractions can distort the P_{es} signal. Patient position, balloon position, and lung volume may influence the amplitude of P_{es} changes due to the cardiac artifact (89). Esophageal contraction due to peristalsis is sometimes present and is easily detected (as a large increase in pressure that bears no relationship with respiratory cycles): in this case, P_{es} measurement should be interrupted until P_{es} returns to its baseline value.

Conclusions

With rapid advances in technology, monitoring of P_{es} can be safely, satisfactorily, and easily performed at the bedside in the intensive care unit. Despite voluminous data showing the usefulness of P_{es} measurements in critically ill patients, the introduction of P_{es} measurements in the intensive care unit has been disappointingly slow.

P_{es} measurements allow the partitioning of respiratory system mechanics into pulmonary and chest wall




Figure 6. Occlusion test in a spontaneously breathing patient (*left*) and in a paralyzed patient (*right*). In the former case, the airway has been occluded during patient effort; in the latter case, the airway has been occluded while applying an external chest compression. P_{aw} axis has been shifted to achieve overlap of the two signals. P_{aw} = airway pressure; P_{es} = esophageal pressure.

components. As such, Pes measurements have enhanced our understanding of the pathophysiology of acute lung injury, patient-ventilator interaction, and weaning failure. By providing a practical means of quantifying respiratory effort, Pes measurements may make it feasible to individualize the level of muscle unloading during mechanical ventilation and may provide a useful clinical tool for assessing patients during a weaning trial. The use of Pes measurements in PEEP titration may help improve oxygenation and compliance in patients with ARDS.

Improved outcome has been demonstrated when physiological principles (low tidal ventilation [38], high PEEP [90], prone position [91], or neuromuscular blocking agents [92]) are applied in ventilator management of patients with ARDS. The time is now right to take advantage of the body of knowledge obtained with Pes and apply it to improve the management of critically ill and ventilator-dependent patients. ■

Author disclosures are available with the text of this document at www.atsjournals.org.

Collaborators of the PLUG working group (Acute Respiratory Failure Section of the European Society of Intensive Care Medicine): Fekri Abroug (Monastir, Tunisia), Evangelia Akoumianaki (Heraklion, Greece), Jean-Michel Arnal (Toulon, France), Antonio Artigas (Sabadell, Spain), Dimitrios Bampalis (Larissa, Greece), Alessandro Beda (Belo Horizonte, Brazil), Gaetan Beduneau (Rouen, France), Thomas Bein (Regensburg, Germany), Giacomo Bellani (Milan, Italy), Francois Beloncle (Angers, France), Luis Blanch (Sabadell, Spain), Laurent Brochard (Toronto, Canada), Luigi Camporota (London, UK), Gilles Capellier (Besancon, France), Alysson R. Carvalho (Rio de Janeiro, Brazil), Davide Chiumello (Milan, Italy), Ricardo Cordioli (Sao Paulo, Brazil), Alexandre Demoule (Paris, France), Jean-Luc Diehl (Paris, France), Eddy Fan (Toronto, Canada), Niall Ferguson (Toronto, Canada), Luciano Gattinoni (Milan, Italy), Dimitris Georgopoulos (Heraklion, Greece), Ewan Goligher (Toronto, Canada), Salvatore Grasso (Bari, Italy), Claude Guerin (Lyon, France), Christophe Guerville

(Marseille, France), Leo M. A. Heunks (Nijmegen, The Netherlands), Robert Huhle (Dresden, Germany), Sébastien Jochmans (Melun, France), Bjorn Jonson (Lund, Sweden), Amal Jubran (Chicago, IL), Brian Kavanagh (Toronto, Canada), Joseph Keenan (St. Paul, MN), John Laffey (Toronto, Canada), Franco Laghi (Chicago, IL), Stephen Loring (Boston, MA), Umberto Lucangelo (Trieste, Italy), Aissam Lyazidi (Rabat, Morocco), Salvatore M. Maggiore (Rome, Italy), Jordi Mancebo (Barcelona, Spain), Tommaso Mauri (Milan, Italy), Alain Mercat (Angers, France), Francesco Mojoli (Pavia, Italy), Stefano Nava (Bologna, Italy), Paolo Pelosi (Genoa, Italy), Gaetano Perchiazzi (Bari, Italy), Antonio Pesenti (Monza, Italy), Lise Piquilloud (Lausanne, Switzerland), Michael Quintel (Göttingen, Germany), Marco Ranieri (Turin, Italy), Jean-Christophe M. Richard (Annecy, France), Nuttapol Rittayamai (Bangkok, Thailand), Arthur Slutsky (Toronto, Canada), Lorenzo Del Sorbo (Toronto, Canada), Daniel Talmor (Boston, MA), Martin Tobin (Chicago, IL), Franco Valenza (Milan, Italy), Norbert Weiler (Kiehl, Germany), and Takeshi Yoshida (Osaka, Japan)

References

1. Buytendijk JH. Intraesophageal pressure and lung elasticity [thesis]. Groningen, the Netherlands: University of Groningen; 1949.
2. Dornhorst AC, Leathart GL. A method of assessing the mechanical properties of lungs and air-passages. *Lancet* 1952;2:109–111.
3. Cherniack RM, Farhi LE, Armstrong BW, Proctor DF. A comparison of esophageal and intrapleural pressure in man. *J Appl Physiol* 1955;8: 203–211.
4. Grasso S, Terragni P, Birocco A, Urbino R, Del Sorbo L, Filippini C, Mascia L, Pesenti A, Zangrillo A, Gattinoni L, et al. ECMO criteria for influenza A (H1N1)-associated ARDS: role of transpulmonary pressure. *Intensive Care Med* 2012;38:395–403.
5. Jubran A, Grant BJ, Laghi F, Parthasarathy S, Tobin MJ. Weaning prediction: esophageal pressure monitoring complements readiness testing. *Am J Respir Crit Care Med* 2005;171:1252–1259.
6. Talmor D, Sarge T, Malhotra A, O'Donnell CR, Ritz R, Lisbon A, Novack V, Loring SH. Mechanical ventilation guided by esophageal pressure in acute lung injury. *N Engl J Med* 2008;359: 2095–2104.
7. Thille AW, Cabello B, Galia F, Lyazidi A, Brochard L. Reduction of patient-ventilator asynchrony by reducing tidal volume during pressure-support ventilation. *Intensive Care Med* 2008;34: 1477–1486.
8. Tobin MJ. Monitoring respiratory mechanics in spontaneously breathing patients. In: Tobin MJ, editor. Principles and practice of intensive care monitoring. New York: McGraw-Hill; 1998, pp. 617–653.
9. Agostoni E, Hyatt R. Static behavior of the respiratory system. In: Fishman AP, Macklem PT, Mead J, editors. Handbook of physiology, Section 3: The respiratory system. Bethesda, MD: American Physiological Society; 1986, pp. 113–130.
10. Pecchiari M, Loring SH, D'Angelo E. Esophageal pressure as an estimate of average pleural pressure with lung or chest distortion in rats. *Respir Physiol Neurobiol* 2013;186:229–235.
11. Kreil WS, Rodarte JR. Effects of acute pleural effusion on respiratory system mechanics in dogs. *J Appl Physiol* 1985;59:1458–1463.
12. Guérin C, Richard JC. Comparison of 2 correction methods for absolute values of esophageal pressure in subjects with acute hypoxemic respiratory failure, mechanically ventilated in the ICU. *Respir Care* 2012;57:2045–2051.
13. Gulati G, Novero A, Loring SH, Talmor D. Pleural pressure and optimal positive end-expiratory pressure based on esophageal pressure versus chest wall elastance: incompatible results. *Crit Care Med* 2013;41:1951–1957.
14. Loring SH, O'Donnell CR, Behazin N, Malhotra A, Sarge T, Ritz R, Novack V, Talmor D. Esophageal pressures in acute lung injury: do they represent artifact or useful information about transpulmonary pressure, chest wall mechanics, and lung stress? *J Appl Physiol* 2010;108:515–522.
15. Pepe PE, Marini JJ. Occult positive end-expiratory pressure in mechanically ventilated patients with airflow obstruction: the auto-PEEP effect. *Am Rev Respir Dis* 1982;126:166–170.
16. Roussos C. Structure and function of the thorax: Energetics. In: Roussos C, Macklem PT, editors. The thorax. New York: Marcel Dekker; 1985, pp. 437–492.
17. Mead J, Loring SJ. Volume displacements of the chest wall and their mechanical significance. In: Roussos C, Macklem PT, editors. The thorax. New York: Marcel Dekker; 1985, pp. 369–392.
18. Brochard L, Harf A, Lorino H, Lemaire F. Inspiratory pressure support prevents diaphragmatic fatigue during weaning from mechanical ventilation. *Am Rev Respir Dis* 1989;139:513–521.
19. Robertson CH Jr, Foster GH, Johnson RL Jr. The relationship of respiratory failure to the oxygen consumption of, lactate production by, and distribution of blood flow among respiratory muscles during increasing inspiratory resistance. *J Clin Invest* 1977;59:31–42.
20. Robertson CH Jr, Pagel MA, Johnson RL Jr. The distribution of blood flow, oxygen consumption, and work output among the respiratory muscles during unobstructed hyperventilation. *J Clin Invest* 1977;59: 43–50.
21. Tobin MJ, Laghi F, Jubran A. Ventilatory failure, ventilator support, and ventilator weaning. *Compr Physiol* 2012;2:2871–2921.
22. Field S, Sanci S, Grassino A. Respiratory muscle oxygen consumption estimated by the diaphragm pressure-time index. *J Appl Physiol* 1984;57:44–51.

23. Ballantine TV, Proctor HJ, Broussard ND, Litt BD. The work of breathing: potential for clinical application and the results of studies performed on 100 normal males. *Ann Surg* 1970;171: 590–594.
24. Delhez L, Petit JM, Mutsers A, Troquet J. [Some ways of function of the thoracopulmonary apparatus in the normal elderly person]. *Acta Tuberc Pneumol Belg* 1961;52:373–388.
25. Mancebo J, Isabey D, Lorino H, Lofaso F, Lemaire F, Brochard L. Comparative effects of pressure support ventilation and intermittent positive pressure breathing (IPPB) in non-intubated healthy subjects. *Eur Respir J* 1995;8:1901–1909.
26. Leung P, Jubran A, Tobin MJ. Comparison of assisted ventilator modes on triggering, patient effort, and dyspnea. *Am J Respir Crit Care Med* 1997;155:1940–1948.
27. Kushida CA, Giacomini A, Lee MK, Guilleminault C, Dement WC. Technical protocol for the use of esophageal manometry in the diagnosis of sleep-related breathing disorders. *Sleep Med* 2002;3:163–173.
28. Jardin F, Genevray B, Brun-Ney D, Bourdarias J-P. Influence of lung and chest wall compliances on transmission of airway pressure to the pleural space in critically ill patients. *Chest* 1985;88:653–658.
29. Malbrain ML, Cheatham ML, Kirkpatrick A, Sugrue M, Parr M, De Waele J, Balogh Z, Leppäniemi A, Olvera C, et al. Results from the International Conference of Experts on Intra-abdominal Hypertension and Abdominal Compartment Syndrome. I. Definitions. *Intensive Care Med* 2006;32:1722–1732.
30. Ranieri VM, Brienza N, Santostasi S, Puntillo F, Mascia L, Vitale N, Giuliani R, Memeo V, Bruno F, Fiore T, et al. Impairment of lung and chest wall mechanics in patients with acute respiratory distress syndrome: role of abdominal distension. *Am J Respir Crit Care Med* 1997;156:1082–1091.
31. Kirkpatrick AW, Roberts DJ, De Waele J, Jaeschke R, Malbrain ML, De Keulenaer B, Duchesne J, Bjorck M, Leppäniemi A, Ejike JC, et al.; Pediatric Guidelines Sub-Committee for the World Society of the Abdominal Compartment Syndrome. Intra-abdominal hypertension and the abdominal compartment syndrome: updated consensus definitions and clinical practice guidelines from the World Society of the Abdominal Compartment Syndrome. *Intensive Care Med* 2013; 39:1190–1206.
32. Terragni PP, Filippini C, Slutsky AS, Birocco A, Tenaglia T, Grasso S, Stripoli T, Pasero D, Urbino R, Fanelli V, et al. Accuracy of plateau pressure and stress index to identify injurious ventilation in patients with acute respiratory distress syndrome. *Anesthesiology* 2013;119: 880–889.
33. Mutoh T, Lamm WJ, Embree LJ, Hildebrandt J, Albert RK. Volume infusion produces abdominal distension, lung compression, and chest wall stiffening in pigs. *J Appl Physiol* 1992;72:575–582.
34. Quintel M, Pelosi P, Caironi P, Meinhardt JP, Luecke T, Herrmann P, Taccone P, Rylander C, Valenza F, Carlesso E, et al. An increase of abdominal pressure increases pulmonary edema in oleic acid-induced lung injury. *Am J Respir Crit Care Med* 2004;169:534–541.
35. Gattinoni L, Pelosi P, Suter PM, Pedoto A, Vercesi P, Lissoni A. Acute respiratory distress syndrome caused by pulmonary and extrapulmonary disease: different syndromes? *Am J Respir Crit Care Med* 1998;158:3–11.
36. Pelosi P, Cereda M, Foti G, Giacomini M, Pesenti A. Alterations of lung and chest wall mechanics in patients with acute lung injury: effects of positive end-expiratory pressure. *Am J Respir Crit Care Med* 1995; 152:531–537.
37. De Prost N, Dreyfuss D. How to prevent ventilator-induced lung injury? *Minerva Anesthesiol* 2012;78:1054–1066.
38. Acute Respiratory Distress Syndrome Network. Ventilation with lower tidal volumes as compared with traditional tidal volumes for acute lung injury and the acute respiratory distress syndrome. *N Engl J Med* 2000;342:1301–1308.
39. Staffieri F, Stripoli T, De Monte V, Crovace A, Sacchi M, De Michele M, Trerotoli P, Terragni P, Ranieri VM, Grasso S. Physiological effects of an open lung ventilatory strategy titrated on elastance-derived end-inspiratory transpulmonary pressure: study in a pig model. *Crit Care Med* 2012;40:2124–2131.
40. Futier E, Constantin JM, Paugam-Burtz C, Pascal J, Eurin M, Neuschwander A, Marret E, Beaussier M, Gutton C, Lefrant JY, et al.; IMPROVE Study Group. A trial of intraoperative low-tidal-volume ventilation in abdominal surgery. *N Engl J Med* 2013;369:428–437.
41. Severgnini P, Selmo G, Lanza C, Chiesa A, Frigerio A, Bacuzzi A, Dionigi G, Novario R, Gregoret C, de Abreu MG, et al. Protective mechanical ventilation during general anesthesia for open abdominal surgery improves postoperative pulmonary function. *Anesthesiology* 2013;118:1307–1321.
42. Pelosi P, Luecke T, Rocco PR. Chest wall mechanics and abdominal pressure during general anaesthesia in normal and obese individuals and in acute lung injury. *Curr Opin Crit Care* 2011;17: 72–79.
43. Leme Silva P, Pelosi P, Rocco PR. Mechanical ventilation in obese patients. *Minerva Anesthesiol* 2012;78:1136–1145.
44. Cassidy SS, Robertson CH Jr, Pierce AK, Johnson RL Jr. Cardiovascular effects of positive end-expiratory pressure in dogs. *J Appl Physiol* 1978;44:743–750.
45. Zarins CK, Virgilio RW, Smith DE, Peters RM. The effect of vascular volume on positive end-expiratory pressure-induced cardiac output depression and wedge-left atrial pressure discrepancy. *J Surg Res* 1977;23:348–360.
46. Magder S. Invasive intravascular hemodynamic monitoring: technical issues. *Crit Care Clin* 2007;23:401–414.
47. Brochard L, Pluskwa F, Lemaire F. Improved efficacy of spontaneous breathing with inspiratory pressure support. *Am Rev Respir Dis* 1987; 136:411–415.
48. Imsand C, Feihl F, Perret C, Fitting JW. Regulation of inspiratory neuromuscular output during synchronized intermittent mechanical ventilation. *Anesthesiology* 1994;80:13–22.
49. Marini JJ, Rodriguez RM, Lamb V. The inspiratory workload of patient-initiated mechanical ventilation. *Am Rev Respir Dis* 1986;134: 902–909.
50. Marini JJ, Smith TC, Lamb VJ. External work output and force generation during synchronized intermittent mechanical ventilation: effect of machine assistance on breathing effort. *Am Rev Respir Dis* 1988;138:1169–1179.
51. Chao DC, Scheinhorn DJ, Stearn-Hassenpflug M. Patient-ventilator trigger asynchrony in prolonged mechanical ventilation. *Chest* 1997; 112:1592–1599.
52. Thille AW, Rodriguez P, Cabello B, Lellouche F, Brochard L. Patient-ventilator asynchrony during assisted mechanical ventilation. *Intensive Care Med* 2006;32:1515–1522.
53. de Wit M, Pedram S, Best AM, Epstein SK. Observational study of patient-ventilator asynchrony and relationship to sedation level. *J Crit Care* 2009;24:74–80.
54. Colombo D, Cammarota G, Alemani M, Carenzo L, Barra FL, Vaschetto R, Slutsky AS, Della Corte F, Navalesi P. Efficacy of ventilator waveforms observation in detecting patient-ventilator asynchrony. *Crit Care Med* 2011;39:2452–2457.
55. Carreaux G, Mancebo J, Mercat A, Dellamonica J, Richard JC, Aguirre-Bermeo H, Kouatchet A, Beduneau G, Thille AW, Brochard L. Bedside adjustment of proportional assist ventilation to target a predefined range of respiratory effort. *Crit Care Med* 2013;41: 2125–2132.
56. Akoumianaki E, Lyazidi A, Rey N, Matamis D, Perez-Martinez N, Giraud R, Mancebo J, Brochard L, Marie Richard JC. Mechanical ventilation-induced reverse-triggered breaths: a frequently unrecognized form of neuromechanical coupling. *Chest* 2013;143: 927–938.
57. Simon PM, Zurob AS, Wies WM, Leiter JC, Hubmayr RD. Entrainment of respiration in humans by periodic lung inflations: effect of state and CO₂. *Am J Respir Crit Care Med* 1999;160:950–960.
58. Richard JC, Lyazidi A, Akoumianaki E, Mortaza S, Cordioli RL, Lefebvre JC, Rey N, Piquilloud L, Sferazza Papa GF, Mercat A, et al. Potentially harmful effects of inspiratory synchronization during pressure preset ventilation. *Intensive Care Med* 2013;39:2003–2010.
59. de Wit M, Miller KB, Green DA, Ostman HE, Gennings C, Epstein SK. Ineffective triggering predicts increased duration of mechanical ventilation. *Crit Care Med* 2009;37:2740–2745.

60. Giannouli E, Webster K, Roberts D, Younes M. Response of ventilator-dependent patients to different levels of pressure support and proportional assist. *Am J Respir Crit Care Med* 1999;159:1716–1725.
61. Yoshida T, Torsani V, Gomes S, De Santis RR, Beraldo MA, Costa EL, Tucci MR, Zin WA, Kavanagh BP, Amato MB. Spontaneous effort causes occult pendelluft during mechanical ventilation. *Am J Respir Crit Care Med* 2013;188:1420–1427.
62. Vignaux L, Vargas F, Roeseler J, Tassaux D, Thille AW, Kossowsky MP, Brochard L, Joliet P. Patient–ventilator asynchrony during non-invasive ventilation for acute respiratory failure: a multicenter study. *Intensive Care Med* 2009;35:840–846.
63. Brochard L. Intrinsic (or auto-) positive end-expiratory pressure during spontaneous or assisted ventilation. *Intensive Care Med* 2002;28:1552–1554.
64. Lessard MR, Lofaso F, Brochard L. Expiratory muscle activity increases intrinsic positive end-expiratory pressure independently of dynamic hyperinflation in mechanically ventilated patients. *Am J Respir Crit Care Med* 1995;151:562–569.
65. Tobin MJ, Jubran A. Weaning from mechanical ventilation. In: Tobin MJ, editor. Principles and practice of mechanical ventilation, 3rd ed. New York: McGraw-Hill; 2012, pp. 1307–1351.
66. Jubran A, Tobin MJ. Pathophysiologic basis of acute respiratory distress in patients who fail a trial of weaning from mechanical ventilation. *Am J Respir Crit Care Med* 1997;155:906–915.
67. Lemaire F, Teboul JL, Cinotti L, Giotto G, Abrouk F, Steg G, Macquin-Mavier I, Zapol WM. Acute left ventricular dysfunction during unsuccessful weaning from mechanical ventilation. *Anesthesiology* 1988;69:171–179.
68. Peters RM, Hilberman M, Hogan JS, Crawford DA. Objective indications for respirator therapy in post-trauma and postoperative patients. *Am J Surg* 1972;124:262–269.
69. Proctor HJ, Woolson R. Prediction of respiratory muscle fatigue by measurements of the work of breathing. *Surg Gynecol Obstet* 1973;136:367–370.
70. Cabello B, Thille AW, Roche-Campo F, Brochard L, Gómez FJ, Mancebo J. Physiological comparison of three spontaneous breathing trials in difficult-to-wean patients. *Intensive Care Med* 2010;36:1171–1179.
71. Summer WR, Permutt S, Sagawa K, Shoukas AA, Bromberger-Barnea B. Effects of spontaneous respiration on canine left ventricular function. *Circ Res* 1979;45:719–728.
72. Gilbert R, Peppi D, Auchincloss JH Jr. Measurement of transdiaphragmatic pressure with a single gastric–esophageal probe. *J Appl Physiol* 1979;47:628–630.
73. Hedenstierna G. Esophageal pressure: benefit and limitations. *Minerva Anesthesiol* 2012;78:959–966.
74. Milic-Emili J, Mead J, Turner JM, Glauser EM. Improved technique for estimating pleural pressure from esophageal balloons. *J Appl Physiol* 1964;19:207–211.
75. Petit JM, Milic-Emili G. Measurement of endoesophageal pressure. *J Appl Physiol* 1958;13:481–485.
76. Schilder DP, Hyatt RE, Fry DL. An improved balloon system for measuring intraesophageal pressure. *J Appl Physiol* 1959;14:1057–1058.
77. Baydur A, Behrakis PK, Zin WA, Jaeger M, Milic-Emili J. A simple method for assessing the validity of the esophageal balloon technique. *Am Rev Respir Dis* 1982;126:788–791.
78. Niknam J, Chandra A, Adams AB, Nahum A, Ravenscraft SA, Marini JJ. Effect of a nasogastric tube on esophageal pressure measurement in normal adults. *Chest* 1994;106:137–141.
79. Chiumello D, Gallazzi E, Marino A, Berto V, Mietto C, Cesana B, Gattinoni L. A validation study of a new nasogastric polyfunctional catheter. *Intensive Care Med* 2011;37:791–795.
80. Mead J, McLroy MB, Selverstone NJ, Kriete BC. Measurement of intraesophageal pressure. *J Appl Physiol* 1955;7:491–495.
81. Mojoli F, Volta A, Pozzi M, Curro I, Milic-Emili J, Torriglia F, Alvisi R, Braschi A. The use of an oesophageal balloon to measure pleural pressure: effects of different inflation volumes. *Intensive Care Med* 2011;37:S194.
82. Araujo Zin W, Milic-Emili J. Esophageal pressure measurement. In: Hamid Q, Shannon J, Martin JG, editors. Physiologic basis of respiratory disease. Hamilton, ON, Canada: BC Decker; 2005, pp. 639–647.
83. Benditt JO. Esophageal and gastric pressure measurements. *Respir Care* 2005;50:68–75; discussion 75–77.
84. D'Angelo E, Robatto FM, Calderini E, Tavola M, Bono D, Torri G, Milic-Emili J. Pulmonary and chest wall mechanics in anesthetized paralyzed humans. *J Appl Physiol* 1991;70:2602–2610.
85. Higgs BD, Behrakis PK, Bevan DR, Milic-Emili J. Measurement of pleural pressure with esophageal balloon in anesthetized humans. *Anesthesiology* 1983;59:340–343.
86. Lanteri CJ, Kano S, Sly PD. Validation of esophageal pressure occlusion test after paralysis. *Pediatr Pulmonol* 1994;17:56–62.
87. Zin WA, Böddener A, Silva PR, Pinto TM, Milic-Emili J. Active and passive respiratory mechanics in anesthetized dogs. *J Appl Physiol* 1986;61:1647–1655.
88. Milner AD, Saunders RA, Hopkin LE. Relationship of intra-oesophageal pressure to mouth pressure during the measurement of thoracic gas volume in the newborn. *Biol Neonate* 1978;33:314–319.
89. Baydur A, Cha EJ, Sassoon CS. Validation of esophageal balloon technique at different lung volumes and postures. *J Appl Physiol* 1987;62:315–321.
90. Briel M, Meade M, Mercat A, Brower RG, Talmor D, Walter SD, Slutsky AS, Pullenayegum E, Zhou Q, Cook D, et al. Higher vs lower positive end-expiratory pressure in patients with acute lung injury and acute respiratory distress syndrome: systematic review and meta-analysis. *JAMA* 2010;303:865–873.
91. Guérin C, Reignier J, Richard JC, Beuret P, Gacouin A, Boulain T, Mercier E, Badet M, Mercat A, Baudin O, et al.; PROSEVA Study Group. Prone positioning in severe acute respiratory distress syndrome. *N Engl J Med* 2013;368:2159–2168.
92. Papazian L, Forel JM, Gacouin A, Penot-Ragon C, Perrin G, Loundou A, Jaber S, Arnal JM, Perez D, Seghboyan JM, et al.; ACURASYS Study Investigators. Neuromuscular blockers in early acute respiratory distress syndrome. *N Engl J Med* 2010;363:1107–1116.