




(a) made vs caught, 10 mW green



(b) efficiency, and counts against the as-built tip



(c) the green spot sets the resolution



Created counts only NVs the green actually reaches, saturated: $R_{\text{exc}} = R_{\text{SAT}} * s / (1 + s)$ with $s = I_{\text{green}} / I_{\text{SAT}}$, integrated over the 80-90 μm layer. Hatched bars (SM, MM) are bare fibers from the characterisation model, a different tracer, so only the two MCF tips are a like-for-like ratio. In (c) a probe on the dashed line is limited by its own excitation spot; below it the collection optics are tighter than the green and the resolution is set by the overlap of the two.