

Protocol-aware task comparison. Green: matched protocol; blue: no matched reference; amber: unmatched protocol or sample; red: unevaluated setting.

Task	Metric	Reference or comparator	Present study	Protocol status
Reconstruction (EndoNeRF NVS)	held-out-view PSNR	EndoGaussian reports NVS PSNR (not re-run here)	not measured on this split	not comparable
Reconstruction (canonical fit)	tissue-masked PSNR	no published number on this mask	36.79 / 34.09 dB pulling / cutting	no matched reference
SCARED portability	tissue-masked PSNR, single keyframe	no matched published number	26.99 / 26.17 dB kf1 / kf2	no matched reference
Tracking ablation	worst-frame loss, same sequence	inertia-only 37.23 (diverges)	constant velocity + rollback 1.55	matched protocol
Mask ablation	tissue-masked PSNR, same frames	tool-mask supervision 11.13 dB	depth-valid tissue mask 36.79 dB	matched protocol
Uniform modulus	relative error, closed loop	no external method on this loop	0.01%	closed-loop verification
FEM background E	median relative error	no matched published result	4.3% median (6 of 60 scenarios)	no matched reference
Phantom contrast	error vs compression test	no published run on this phantom split	3.80x vs 3.56x (6.9%)	no matched reference
Provenance audit	measured / estimated / assumed	no prior method states this tag	explicit tag on every parameter	no direct comparator
Held-out NVS and real-tissue absolute E	held-out PSNR; force-free E	EndoGaussian NVS; no force-free absolute E	not evaluated	unevaluated