

DECLARATION OF CONFORMITY (EU)

No. 20528/2025

1. Product Name and Model:

Electrode Pads (Electrode Patches)

Model No.: EP-1501

2. Importer:

VATSPEEDSL

Calle Antonio Salvador N99.1, 28026 Madrid, Spain

Email: services@vatspeed-eu.com

3. This declaration of conformity is issued under the sole responsibility of the importer:

VATSPEEDSL

Calle Antonio Salvador N99.1, 28026 Madrid, Spain

4. Product Name and Photo:

Electrode Pads



5. Product complies with the directives:

- General Product Safety Regulation (GPSR) EU 2019/1020
- Regulation (EU) 2023/988 on general product safety
- REACH Regulation EC No 1907/2006 (chemical safety)
- RoHS Directive 2011/65/EU (if applicable – not typical for passive electrode pads)

6. Product complies with the standards:

- ISO 9001:2015 – Quality management systems

- ISO 10993-5:2009 – Biological evaluation of medical devices – Cytotoxicity (for skin-contact materials)
- ISO 10993-10:2021 – Biological evaluation – Skin sensitization (irritation and sensitization testing)
- ISO 10993-23:2021 – Biological evaluation – Irritation tests
- EN 71-3:2019+A1:2021 – Migration of certain elements (for material safety of gel/adhesive layers)
- ASTM D3574 – Flexible cellular materials – Testing of polyurethane foam (for foam backing, if applicable)
- IEC 62321 Series – Determination of six regulated substances (RoHS – if electronic connector present)

7. Additional Notes:

These electrode pads have been tested and found to comply with the applicable EU general product safety regulations and relevant chemical safety standards. They are designed for use with EMS (Electrical Muscle Stimulation), TENS (Transcutaneous Electrical Nerve Stimulation), or cosmetic/beauty devices for non-medical, general wellness purposes. The product meets requirements for material safety, skin biocompatibility (non-cytotoxicity, non-sensitization, non-irritation), absence of restricted substances under REACH (including phthalates, heavy metals, and SVHC), and general product safety. Conductive gel adhesive layer provides secure skin contact. Product is intended for single-patient use (or limited reuse as indicated on packaging).

Date: 2025-01-01

Place: Madrid, Spain

Authorized Signatory:

Name: Carlos Fernández

Position: Compliance Officer



Signature:

