



Prio Retino use requirements

The proper functioning of Prio Retino requires as input a fundus image which must be of sufficiently high quality and can be in several possible formats.

Indeed, the good practices for using Prio Retino require

- 1 a non-blurred fundus image for which an eye doctor should be able to make a diagnosis (if asked ultimately)**
- 2 a fundus image in one of the following formats; '.png', '.jpeg' or '.jpg'**

Finally, the fundus image can be generated with any type of fundus camera as long as the above criteria are respected.