

Lab 04 Report

Christian A. Hardy

2021-04-09

Contents


- Experiment 01: Plot Raw Data vs Time
- 01 Questions and Answers
- Experiment 02: Plot Raw Data vs Time
- Experiment 02: Plot LowPass Filtered Data vs Time
- 02 Questions and Answers
- MATLAB Code Repo

Experiment 01: Plot Raw Data vs Time


Using the ismax function with a minimum value trigger of 65 yields mixed results for the heart rate. With the original blue signal, it does not do that much better at getting the actual heart rate versus manually counting portions of the graph.

Having a more dynamic system or a more programatic way of determining a threshold would be good for calculating the proper heart rate from this data. I imagine that is what goes into the medical versions of such devices. Accuracy and decent precision is everything afterall.

I think a better signal processing algorithm would be more suited normalising the graph. When the signal is particularly amplitude modulated this simple method a finding the heart rate fails more often than not.



Chest Lying Down Heart Rate: 72





Hand Lying Down Heart Rate: 62



01 Questions and Answers



Was there variability between the beats? Would you expect the interval between beats to be identical? Why or why not?



There was variability between the beats. This makes sense as the distance from the heart was different. I would expect the variability between beats to be extremely close however. Assuming a healthy blood vessel system, then I couldn't imagine any issues with the

Experiment 02: Plot Raw Data vs Time











Experiment 02: Plot LowPass Filtered Data vs Time

Using the ismax function with a minimum value trigger of 65 yields mixed results for the heartrate. Even with the lowpass filtering applied to the signal, shown in red over the original blue signal, it does not do that much better at getting the actual heartrate when manually counting portions of the graph.


Having a more dynamic system or a more programatic way of determining a threshold would be good for calculating the proper heart rate from this data. I imagine that is what goes into the medical versions of such devices. Accuracy and decent precision is everything afterall.




Chest Lying Down Heart Rate: 31




Chest Running in Place Heart Rate: 40




Chest Sitting and Standing Heart Rate: 21




Chest Squatting Heart Rate: 39




Hand Lying Down Heart Rate: 62



Hand Running in Place Heart Rate: 100



Hand Sitting and Standing Heart Rate: 83



Hand Squatting Heart Rate: 108

02 Questions and Answers

What physiological advantage is there in a slower resting heart rate?

A slower resting heart rate means your heart is more efficient at pumping oxygenated blood around your body. Since it is stronger, it pumps less often, causing less wear on the heart over the lifetime of the person. Also it likely means that the heart can recover from high periods of activity faster, returning to a lower rate and easing the workload much faster than a not as healthy heart.

MATLAB Code Repo:

Matlab code and reference files can be made found at <https://github.com/llyando/EECE4512>, specifically cloning the entire repo and running the code from LAB04.m, first section, to repeat the results in this report.