Quality assurance report for CRP

Quality assurance pipeline

Purpose

This report summarises submitted quality assurance results for the recorded site, instrument, and cycle set. It is intended for method review and report-generation verification only.

Recorded details

- Site: Site {Alpha}\Path
- Device type: Abbott
- Device ID: DEV_1
- Assay: CRP
- Cycles: Cycle-1 and Cycle-2
- Date/time values:
  - 2026-01-01T00:00:00Z
  - 2026-01-01T00:05:00Z
  - 2026-01-02T00:00:00Z
  - 2026-01-02T00:05:00Z
- Units: mg/L

Summary statistics

Cycle-1 mean: 11 mg/L
Cycle-1 allowable range: 10.45–11.55 mg/L
Cycle-2 mean: 12.5 mg/L
Cycle-2 allowable range: 11.875–13.125 mg/L

Cycle values

Cycle-1: 10.5, 11.5
Cycle-2: 12, 13

Tabulated values

Cycle	Value
Cycle-1	10.5
Cycle-2	12
Cycle-1	11.5
Cycle-2	13