## Kyma Analysis Report

Date: 24. feb 2020 Name: 1991b Temperature: 23.0

Filename: 1991b-TPF-UFCM 20200224

Waves analysed : Bz403,Bx403,Izmes39,Ixmes39 Gap [mm] : 200

Phase [mm] : 9 Hall Z [mm]: 1 Peak Field Bx: 0.09038 T RMS Phase Error: 7.50 deg

Taper [mm]: 0 Hall X [mm]: -12 Delta Bx/Bx : 3.3 % Fund. En. (B2E): 3388.3 eV = 0.4 nmPeak Field Bz: 0.23162 T Radiation wavelength (lin. fit): 0.4 nm

Peaks used: 104

Undulator Period: 22.00+-0.06 mm Delta Bz/Bz: 12.7 % e-beam energy: 3 GeV Bx Taper: -25.527 G/m Bz Taper: -21.745 G/m


Krms: 0.361

V Dipole: 17.1 G cm Iz(coil): 14.1 G cm Pred. C.C. currents V Quadrupole: -216.9 G Iz(BzC): 14.2 Gcm Iz(Bz): 29.0 G cm I1: 0.6840 A V Sextupole: 255.7 G/cm Ix(coil): 11.8 G cm I2: -0.4059 A

Ix(BxC): 11.8 Gcm Ix(Bx): 327.7 G cm

H Dipole: -11.2 G cm  $IIx(sBx) : -0.1682 \text{ G m}^2$ I3: -2.5241 A H Quadrupole: 289.9 G IIz(sBz): -0.2816 G m^2 I4: 0.8603 A

H Sextupole: 358.7 G/cm BzTrj RMS dev.: 0.4 um BxTrj RMS dev.: 0.7 um



Kyma Srl S.S. 14 km 163,5 in Area Science Park – IT-34149 Trieste, Italy - VAT no.: IT-01131640326 Voice: +39.040.375.8796 - Fax: +39.040.375.8029 - E-mail: info@kyma-undulators.eu - web: www.kyma-undulators.eu