VAISALA

Certificate report no. H45-15060013

CALIBRATION CERTIFICATE

Instrument Humidity and Temperature Probe HMP155
Serial number L0550630
Manufacturer Vaisala Oyj, Finland
Calibration date 2nd February 2015

The above instrument was calibrated by comparing the readings of the instrument to working standards of the manufacturer. The reference humidity was calculated from dewpoint temperature and temperature readings with the exception of the driest condition that was measured as relative humidity. Dewpoint temperature was measured with a 373 LHX dewpoint meter. Temperature and relative humidity were measured with two factory working standards. At the time of shipment, the instrument described above met its operating specifications.

The 373 LHX dewpoint meter has been calibrated at Centre for metrology and accreditation (MIKES) by using a MIKES working standard traceable to National Institute of Standards and Technology (NIST). The temperature readings of the factory working standards have been calibrated at an ISO/IEC 17025 accredited calibration laboratory (FINAS), Vaisala Measurement Standards Laboratory (MSL) by using MSL working standards traceable to NIST. The relative humidity readings of the factory working standards have been calibrated at the Vaisala factory by using a 373 LHX dewpoint meter.

Humidity calibration results

Reference humidity	Reference temperature	Observed humidity	Observed probe temperature	Additional probe temperature	Humidity difference	Permissible difference
%RH	°C	%RH	°C	°C	%RH	%RH
+ 0.2	+ 21.99	0.0	-	+ 21.96	- 0.2	±1.0
+ 12.7	+ 21.97	+ 12.6	-	+ 21.99	- 0.1	± 1.0
+ 33.6	+ 21.98	+ 33.3	-	+ 21.95	- 0.3	± 1.0
+ 54.7	+ 21.98	+ 54.4	-	+ 21.98	- 0.3	± 1.0
+ 75.4	+ 21.99	+ 75.8	-	+ 22.00	+ 0.4	± 1.0
+ 94.6	+ 22.00	+ 95.6	-	+ 22.01	+ 1.0	± 1.7

Temperature calibration results

Reference temperature	Observed probe temperature	Temperature difference		Additional probe temperature	Temperature difference	Permissible difference
°C	°C	°C		°C	°C	°C
+ 21.99	-	-		+ 22.00	+ 0.01	± 0.10

Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
MBW 373LHX	11-0404	2014-03-20	M-14H023
PTU307 / T	B2850024	2014-06-02	K008-X01911
HMT337 / T	B0950001	2014-06-02	K008-X01121
PTU307 / U	B2850024	2014-11-03	H45-14451001
HMT337 / RH	B0950001	2014-11-03	H45-14451002

Uncertainties (95 % confidence level, k=2)
Humidity ± 0.6%RH @ 0...40%RH, ± 1.0%RH @ 40...97%RH
Temperature ± 0.10 °C.
Ambient conditions / Humidity 31 ± 5%RH, Temperature 21 ± 1 °C, Pressure 982 ± 1 hPa.

Technician

Doc216127-C

Vaisala Oyj | PO Box 26, FI-00421 Helsinki, Finland
Phone +358 9 894 91 | Fax +358 9 8949 2227
Email firstname.lastname@vaisala.com | www.vaisala.com
Domicile Vantaa, Finland | VAT FI01244162 | Business ID 0124416-2

VAISALA

Certificate report no. H81-15060088

CALIBRATION CERTIFICATE

Instrument	Humidity and Temperature Probe HMP155
Order code	A4GB12B3A3B1B2A
Serial number	L0550630
Manufacturer	Vaisala Oyj, Finland
Calibration date	2nd February 2015

The analog outputs of the above instrument were measured by using working standards of the manufacturer. The outputs were forced by digital input signals to three output values. The observed values were determined by measuring the voltage over the output terminals. All results are traceable in terms of voltage to NIST.

Analog output channel 1 calibration results

Output forced to V	Observed output V	Difference V	Permissible difference V
1.000	1.000	0.000	±0.001
5.000	5.000	0.000	±0.001
9.000	9.000	0.000	±0.001

Analog output channel 2 calibration results

Output forced to V	Observed output V	Difference V	Permissible difference V
1.000	1.000	0.000	±0.001
5.000	5.000	0.000	±0.001
9.000	9.001	0.001	±0.001

Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
HP34970A	MY41005150	2014-06-24	1250-307057321

Uncertainty (95 % confidence level, k=2)
Voltage ±0.00075V

Ambient conditions / Humidity 22.00± 5%RH, Temperature22.30 ± 2 °C, Pressure 986.80 ± 20 hPa.

Technician

doc211861b

Vaisala Oyj | PO Box 26, FI-00421 Helsinki, Finland
Phone +358 9 894 91 | Fax +358 9 8949 2227
Email firstname.lastname@vaisala.com | www.vaisala.com
Domicile Vantaa, Finland | VAT FI01244162 | Business ID 0124416-2