



BECTON DICKINSON (THAILAND) CO.,LTD
24TH FL., THE OFFICES AT CENTRAL WORLD 999/9
RAMA I
PATHUMWAN
BANGKOK
THAILAND
10330
TEL +662 646 1800

Service Report

TO:	FOR:
Kanchana Sonjai Phrachomklao Phetchaburi Hospital 53 Rodfai Rd., T. Klongkrachang Phetchaburi, 76000 032-425505	Work Order Number#: WO-03589462 Case Number#: 02422767

WORK ORDER DETAILS			
Top-Level		Order Type	Field Service
Serial Number	E34297300589	Billing Type	Time and Material
Product	342973 - FACSCalibur Cytometer 3 Color Basic IVD	Customer Down	No
Work Order Task Code	T113 Repair / Troubleshooting Visit		
SERVICE REQUESTED / OBSERVED SYMPTOM / ERROR CODE, BY THE CUSTOMER			
Printer can't print out			
IDENTIFIED PROBLEM			
Printer can't print out			
WORK PERFORMED			
Install new printer Epson L1210 S/N X8LC026107 completed check instrument connected with printer pass , check printing working normally, the instrument working normally			
Install new printer , the instrument working normally			

PARTS USED

The customer acknowledges that the services described above have been provided to customer's satisfaction. Unless otherwise agreed in a valid agreement between the customer and BD, the customer agrees to pay BD's invoice for the services described above at BD's standard rates and subject to BD's standard terms and conditions.