



Case number CAS-0025386910

Order number WO-35857115 | Visit Date 10/03/2026

Instrument Details		Customer Details	
Instrument/Module	cobas 6800	Customer Number	0052027574
Serial Number	1376	Customer Name	P'CHOMKLAO HOSPITAL รพ.พระจอมเกล้า
Internal Instrument Name		Street Address	53 ถ.รถไฟ ต.คลองกระเซง
		ZIP	76000 - Phetchaburi
		Contact Name	Supaporn Sae-aeaw
		Contact Phone	+66864684997
Additional Details			
	PO		

Service Activity Code Remote Support

Purpose of Visit

Instrument: cobas 6800 (S/N 1376), P'CHOMKLAO HOSPITAL รพ.พระจอมเกล้า

Issue: The customer has alleged that the cobas HPV test results from the last two batches showed a higher HPV positive rate than usual. Most of the samples were received from the same hospital hub.

Workflow: The customer transferred sample 1000 µL from the primary container to a secondary tube.

Action: Review the raw data in Colibri and found that the HPV target showed a signal, which was reported as Positive.
No-hardware related flag.

Solution: Suggest the customer clean the autopipette and repeat testing on some samples.

Performed Activities

Review the raw data in Colibri and found that the HPV target showed a signal, which was reported as Positive.
No-hardware related flag.

Time Report


Category	Start	End	Hours	Billing Type
05913616001-Service labour time	10/03/2026 10:00	10/03/2026 11:00	1	Free of Charge

Signature or Stamp

Customer
Supaporn Sae-aeaw

Roche Support
Supatsa Dedsatit

Sign here



By signing this document

- Customer representative confirms they have reviewed the work carried out as described and accepts the system for use.
- Roche Support person confirms the above mentioned work and checks have been completed..
- .

Hotline : 1800-0193-99
Tel : 02 7912200
Line Id : @rcscthailand

Roche Customer Support Center