



Case number CAS-0025859593

Order number WO-36576697 | Visit Date 05/11/2026

Instrument Details

Instrument/Module cobas 6800
Serial Number 1376

Internal Instrument Name

Additional Details

PO

Customer Details

Customer Number 0052027574
Customer Name P'CHOMKLAO HOSPITAL
รพ.พระจอมเกล้า
Street Address 53 ถ.รถไฟ ต.คลองกระเซง
ZIP 76000 - Phetchaburi
Contact Name Supaporn Sae-aeaw
Contact Phone +66864684997

Service Activity Code Remote Support

Purpose of Visit

Issue: The customer reported that the instrument changed to "Error" status while initializing after performing periodic maintenance.
Message Code: Status of magazine "C" has changed to "Error"
Action: The customer restarted the instrument.
TellMe Analysis: No related issues were detected.

Solution: After restarting the instrument, then returned to standby mode and worked normally.

Performed Activities

Remotely to the IG Server to download the PR file to analyze the possible root cause via Tellme.

Time Report


Category	Start	End	Hours	Billing Type
05913616001-Service labour time	05/11/2026 10:20 AM	05/11/2026 11:00 AM	0.67	Free of Charge

Signature or Stamp

Customer
Supaporn Sae-aeaw

Roche Support
Nahathai Uttamamul

Sign here



By signing this document
- Customer representative confirms they have reviewed the work carried out as described and accepts the system for use.
- Roche Support person confirms the above mentioned work and checks have been completed..
- .

Hotline : 1800-0193-99
Tel : 02 7912200
Line Id : @rcscthailand

Roche Customer Support Center