



Case number CAS-0023550589

Order number WO-33176253 | Visit Date 27/06/2025

Instrument Details

Instrument/Module cobas 6800

Serial Number 1376

Internal Instrument Name

Additional Details

PO

Customer Details

Customer Number 0052027574

Customer Name P'CHOMKLAO HOSPITAL
รพ.พระจอมเกล้า

Street Address 53 ถ.รถไฟ ต.คลองกระเซง

ZIP 76000 - Phetchaburi

Contact Name Supaporn Sae-aeaw

Contact Phone +66864684997

Service Activity Code Courtesy Visit

Purpose of Visit

Proactive visit Q2

- Get data/ information for the proactive visit checklist.
- Monitor HPV connect > User print label from cobas infinity
Workflow for HPV: Receive specimen with HIS (Hosxp) > Labnumber link to LIS (cobas infinity) > Testing on cobas800 > Data send back to LIS+HIS > Customer submit result to KTB portal and colab manually.
- Monitor U06T issue > Not observe again.
- Recheck step from hibernate to standby step > The instrument working normally.
- Monitoring cobas 6800 issue and discuss previous issue with user.

Performed Activities

Time Report

Category	Start	End	Hours	Billing Type
05913616001-Service labour time	27/06/2025 09:00	27/06/2025 14:00	5	Free of Charge

Signature or Stamp

Customer
Supaporn Sae-aeaw

Roche Support
Supatsa Dedsatit

Sign here



By signing this document

- Customer representative confirms they have reviewed the work carried out as described and accepts the system for use.
- Roche Support person confirms the above mentioned work and checks have been completed..
- .

Hotline : 1800-0193-99
Tel : 02 7912200
Line Id : @rcscthailand

Roche Customer Support Center