July 30, 2024 *DELIVERED VIA ELECTRONIC MAIL*

Kirill Smirnov
CEMEX Lyons Plant
5134 Ute Highway
Longmont, Colorado 80503
kirill.smirnov@cemex.com

Re: CEMS QA/QC Plan Approval(s)
CEMEX Lyons Cement Plant
PM, Flow, Opacity, NOx, CO, O2, SO2, THC and Hg Analyzers/Monitors
AIRS ID: 013-0003 | Permit No. 95OPBO082

Mr. Smirnov,

The Colorado Department of Public Health and Environment, Air Pollution Control Division has received and reviewed the revised CEMS Quality Assurance/Quality Control plans submitted on June 21, 2024. The table below indicates the plan identification, pollutant and/or parameter(s) monitored, monitor information, and the permit and/or regulation requiring the CEMS.

Document Name Date and/or Revision #	Parameter	Monitor Manufacturer/Model	Applicable Permit or Regulation
Continuous Monitoring System Quality Program Revision 5 – May, 2024	PM	SICK SP100	95OPBO082, Condition 18.1 40 CFR 63 Subpart A 40 CFR 60 Appendix F
	Flow	TML ULTRAFLOW 150	
	Opacity	LAND 4500 MKII	
	NO, SO2, CO, CO2, O2, H2O	SICK MCS 100E	
	THC	SICK GMS-810 FIDOR (kiln) SICK GMS-840 FIDOR (dryer)	
	Hg	APEX XC 6000 Sorbent Trap	

The plan identified above meets the requirements of the applicable procedure of 40 CFR Part 60, Appendix F, and is approved by the division (July 30, 2024). Please maintain a copy of this approval with your CEMS records.

Sincerely,

Dario Castro
Emissions Monitoring Unit Inspector
Compliance and Enforcement Program