

## Performance of the LSST Calibration Systems

PATRICK INGRAHAM<sup>1</sup>

<sup>1</sup>*Rubin Observatory Project Office, 950 N. Cherry Ave., Tucson, AZ 85719, USA*

(Dated: April 22, 2020)

### ABSTRACT

This is the abstract in the abstract.tex file

### 1. INTRODUCTION

- Much of LSST science will be limited by systematic errors in the photometry.
- The importance of characterizing and eliminating/compensating for these systematics was recognized early on and systems have been developed at tested to minimized their impact. Preliminary design work in [Ingraham et al. \(2016\)](#).
- Photometric errors also come from atmospheric absorption, but these effects are measured by a separate telescope ([Ingraham \(2020\)](#)) and corrections applied at the table level
- Discuss challenges of creating proper flat fields and why a single flat-field screen is not sufficient. Need to go very blue, materials absorb, small space between dome and telescope, reflective optic very fast etc. Needs to be serviceable. Constraints on alignment, mass, scattered light during ops
- Paper discusses the suite of instruments developed, designed and fabricated to perform these analyses
- Presents current performance and future planned developments

### 2. CALIBRATION SCREEN SYSTEM

- Top level description of the screen system
  - White reflective screen permanently mounted on the rotating enclosure.
  - Inclined to correspond to a telescope elevation of 21deg
  - Flats take into account transmission variations over the FoV and are a function of wavelength
  - Traditional dome flats have limitations, particularly at low-frequency type variations. This puts strong constraints on the system

- Used daily to take high SNR flats to monitor dust evolution
- Daily flats are delta correction to the master flats
- Master flats built using monochromatic illumination, removal of low-frequency illumination variation needs to be taken out using a secondary correction
- Discuss flatness and wavelength requirements which dictate design
  - Flatness and continuity relates directly to intensity variation over the focal plane
  - Aiming for a 1% uniformity results in a flatness requirements of  $\sim 3\text{mm}$  over the entire surface
  - Need wavelength range that spans the entire LSST bandpass, but more required to verify characterize any red or blue leaks
  - Single surface not possible due to assembly and transportation constraints, said multiple panels must be used
  - Means tip/tilt system must be incorporated to align adjacent panels (using laser tracker)
  - Servicing the vent gates means one must be able get a lift behind the screen, necessitates that screen must be actuated and able to switch between service (vertical) and operational (inclined) positions
- Include figure showing the screen in the dome and where the laser is located
- Uniformity measurement verification in the dome very challenging. Instead verified in the Tucson lab using sample panels and all the other real hardware before shipping

## 2.1. *Illumination Sources*

### 2.1.1. *White Light LEDs*

- Daily flats require very bright broadband sources to minimize time required for daytime calibrations.
- Useful to have (very) high flux source for other detector characterization (PTCs etc)
- Following DECam design (cite Jennifer Marshall) selected bright diodes, one per filter
- insert table of lamp, central wavelength, wattage, exposure time to get half well depth
- Monochromatic scans require a tunable laser

### 2.1.2. *Laser*

- Selected Ekspla NT-242 due to wavelength range and 1 kHz pulse rate. Wanted to keep lower pulse energies to minimize alignment error destruction of fibers.
- Must be fed into fiber for transport, but fibers heavily absorb blue light. Selected special broadband fiber (CeramOptec Optran UV/WF)
- Laser must be located on rotating enclosure
- Required air cooling as laser will be in conditions that drop below freezing.
- Laser located in special container that does temperature control and handles cleanliness
- takes  $\sim 3$ h to warm up and  $\sim 1$ h to cool down
- Laser use primarily done during cloudy nights
- Discussion of laser stability testing (wavelength and intensity)
- Plot of integration time versus wavelength to get  $\text{SNR} = 100$  per pixel
- Discussion of laser dropout at half the pump wavelength and how it can be mitigated using different laser design

### 2.2. *Screen Design and Fabrication*

- Tried to maximize panel size and minimize gaps (1mm)
- Each panel requires 3 points to do tip/tilt and be fully constrained
- panel substructure mounted to a rigid super structure
- Made entirely of aluminum for weight reasons
- Bearing is type XYZ
- actuation between service and inclined position controlled to  $X$   $\mu\text{m}$ , (about an arcsec of rotation) and performed by actuator located  $3/4$  of the way up the screen (include photo?)
- panel alignment good to  $X$   $\mu\text{m}$ , performed using laser tracker
- all support structure heat treated but flatness will be monitored to check for drift

### 2.3. *Illumination System*

- Discuss original multi-projector design presented in [Sebag et al. \(2014\)](#), lack of blue throughput resulted in re-design
- Minimizing fiber length meant putting laser on telescope platform (not designed for that) or on dome. Decided against putting laser on floor such that no human is required to connect fiber. Important to keep operation from requiring human intervention.
- Fiber placed at center of screen with NA=0.11, reflector mounted on telescope top-end to redirect light to screen
- fiber intensity matched against BRDF of screen material which governed the shape of the optic
- Optic carved out of pure aluminum. Can tolerate higher scattering compared to glass, easier to shape
- mounted on whippletree to equally loadshare and not over-constrain the optic. Also has flexures against rotation
- Optic needs to be aligned to boresight, so mounted on top end just above camera cable wrap on a manual hexapod
- Optic surrounded by black baffling and has actuators which remove 3 petals exposing the optic for flats, then close to minimize scattered light and protect the optic
- Regular cleaning will be performed, reflectivity degradation tolerable. Might get a clear coat if required
- Center area contains mountings for a small amount of opto-mechanics
- Three nests for retro-reflectors machined into optic for alignment purposes

### 2.4. *Alignment*

- XY positions of reflector, screen, and fiber positioning all done and fed into Spatial Analyzer model
- Telescope relative to screen is done from using alignment system for AOS ([Xin \(2020\)](#)) where laser tracker at center of M1M3 measures positions of 6 retro-reflectors on dome to get position angle (and decenter relative to mirror)
- Dome positioning is too rough ( $\sim 10\text{cm}$ ) so angular correction is performed by moving the telescope
- Telescope then moved to specified elevation to match calibration screen inclination angle.

- Alignment that cannot be performed via laser tracker is tip/tilt of fiber, which is controlled using Zaber stage
- TODO: Describe alignment that isn't even determined yet but will be done in the lab in Tucson

### 3. COLLIMATED BEAM PROJECTOR

- CBP designed to mimic star light, and can put arbitrary shapes on the focal plane, but only using a small portion of the pupil.
- When used in coordination with the telescope, can be used to measure the low-to-mid frequency transmission properties of the optical system.
- also useful for other CCD characterization (brighter fatter, charge transfer efficiencies etc.)
- Used in conjunction with the Tunable laser in 2.1.2 can measure/monitor filter curve transmission.
- Demonstrated in [Coughlin et al. \(2016, 2018\)](#)
- A LSST specific CBP has been built for LSST and is mounted on the dome above the laser.

#### 3.1. *Design and Fabrication*

- Designed and Fabricated by DFM Engineering
- Folded Schmidt design with 4.1 degree FoV. Multi-element corrector gives high IQ over the full field (figure in section below)
- Holds 5 masks, source intensity is monitored using Hamamatsu S2281 photodiode in an integrating sphere.
- Each mask can be rotated for precise alignment
- Mount repeatability and minimum motions are X arcsec. No pointing model used as corrections are incremental and very small.
- Position angles required for LSST and the CBP dance to choose a position. Mention software designed to perform the coordinate transformations.

#### 3.2. *Image Quality and Opto-mechanical Performance*

Have on-axis data from acceptance testing, but need actual test data before writing this in any detail.

- show map of measured IQ as a function of field angle
- compare to theoretical curve

- discuss collimation and internal collimation source
- Discuss light levels (at diffraction limit and for resolved objects, which are all surface brightness calcs)
- Discuss CBP throughput measurements as a fcn of wavelength and how this is measured (hopefully Stubbs' solar cells)

### 3.3. *Mask Designs and Use-Cases*

- Mask designs, depending on use case, must take into account distortion in the CBP and distortion in the LSST optics in order to form a desired image on the detector
- Show distortion matrix for each system and explain how to take a desired focal plane image and what mask shape to generate
- Default masks put one spot per CCD. Have two masks, one with a hole size of  $X \text{ um}$  ( $Y \text{ arcsec}$ ) the other with  $P \text{ um}$  ( $Q \text{ arcsec}$ ), also have Ronchi grating
- Need opaque surface as chrome deposits have small throughput/background. Use holes cut by laser ( $5 \text{ }\mu\text{m}$ ) and/or electric Discharge machining (EDM,  $100 \text{ }\mu\text{m}$  holes).
- Figure showing mask designs

### 3.4. *Relative Throughput Calibration of LSST*

- How to measure LSST transmission function, specifically for filters
- possibility to expand this to absolute measurement, but beyond the scope of this paper I think

### 3.5. *Ghost Tracking*

- Useful for characterizing ghosts, but ghosts from CBP data will appear slightly different due to only a partial pupil illumination.
- Ghosts that are in dome flats are not necessarily the same as ghosts from point sources!
- Better used for qualitative analysis and understanding as quantification would be very challenging

## 4. MASTER FLAT CONSTRUCTION

Needs to be coordinated closely with [Lupton \(2020\)](#). This could go in PSTN-026 but I think it better belongs here. This is actually a very significant research project and will take months to get working... if it works.

- Dome flats are useful for high-frequency (pixel-to-pixel level responses such as hot/cold/non-linear pixels), Not useful for large variations. Also affected by ghosts that are in the flats but not in real data!
- Variability over the flat needs to be measured, this will indicate sampling spacing from CBP
- Full technique still being explored

#### 4.1. *Procedure*

#### 4.2. *Calibrate CBP Relative throughput*

At non-vignetted area of the pupil, measure correction for relative throughput between holes. Do this by moving CBP to put adjacent spots on same pixels as central spot to get relationship, spirals out to get relative mapping.

#### 4.3. *Raster over Pupil*

Explain how to raster over the pupil, using normal incidence for on-axis source. Need index that tracks and relates all photometry to each position.

#### 4.4. *Merging Dome and CBP data*

Need to fit the broad variations in dome flats and remove them, then multiply by the measured variations from the CBP data to give proper illumination correction (including vignetting).

#### 4.5. *Flat Field Analysis*

Discuss variations and artifacts in flats, compare to star flat measured during commissioning. Need to coordinate here with [Lupton \(2020\)](#) as it'll use all the downstream infrastructure. This flat is really just an input to pipelines ([Jenness \(2020\)](#)).

## APPENDIX

### A. REFERENCES

#### REFERENCES

- |                                                                                                                                                                                                                                                    |                                                                                                                                                                                                                                                                                                         |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>Coughlin, M., Abbott, T. M. C.,<br/>Brannon, K., et al. 2016, Society of<br/>Photo-Optical Instrumentation<br/>Engineers (SPIE) Conference Series,<br/>Vol. 9910, A collimated beam projector<br/>for precise telescope calibration, 99100V</p> | <p>Coughlin, M. W., Deustua, S., Guyonnet,<br/>A., et al. 2018, in <a href="#">Society of<br/>Photo-Optical Instrumentation<br/>Engineers (SPIE) Conference Series,<br/>Vol. 10704, Proc. SPIE, 1070420</a><br/>Ingraham, P. 2020, Atmospheric<br/>Properties with the LSST Auxiliary<br/>Telescope</p> |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

- Ingraham, P., Stubbs, C. W., Claver, C., et al. 2016, Society of Photo-Optical Instrumentation Engineers (SPIE) Conference Series, Vol. 9906, The LSST calibration hardware system design and development, 99060O
- Jenness, T. 2020, LSST Data Management Software System
- Lupton, R. 2020, LSST Calibration Strategy and Pipelines
- Sebag, J., Gressler, W. J., Liang, M., et al. 2014, Society of Photo-Optical Instrumentation Engineers (SPIE) Conference Series, Vol. 9145, Calibration systems for LSST, 91453L
- Xin, B. 2020, Performance of the LSST Active Optics System



## B. ACRONYMS

<b>Acronym</b>	<b>Description</b>
AOS	Active Optics System
CBP	Collimated Beam Projector
CCD	Charge-Coupled Device
FoV	Field of View (also denoted FOV)
LSST	Legacy Survey of Space and Time (formerly Large Synoptic Survey Telescope)
M1M3	Primary Mirror Tertiary Mirror
PSTN	Project Science Technical Note
arcsec	arcsecond second of arc (unit of angle)
deg	degree; unit of angle