

AAMI EC57 DS2 Benchmark: Macro & Micro Performance (k = 2 Model)

Evaluated across all 22 independent patient recordings of the AAMI EC57 DS2 partition

1. DS2 22 Patients Macro Metrics Summary (Mean ± Std Dev)

Evaluation Scope	Accuracy (%) (Mean ± Std)	Macro Precision (%) (Mean ± Std)	Macro Recall (%) (Mean ± Std)	Macro F1-Score (%) (Mean ± Std)
All 22 DS2 Patients	90.09 ± 19.54%	66.59 ± 21.54%	66.63 ± 21.83%	62.44 ± 22.93%
21 Patients (Excl. 232 Outlier)	93.29 ± 13.20%	66.84 ± 21.99%	67.41 ± 22.18%	64.51 ± 21.78%
Clean Cohort (20 Records, Excl. 117 & 232)	95.99 ± 5.56%	67.68 ± 22.37%	69.79 ± 21.22%	66.32 ± 21.28%

2. DS2 22 Patients Micro Metrics Summary (With Micro Accuracy)

Micro Metric across 22 Patients	Normal (N) (Mean ± Std)	SVEB (S) (Mean ± Std)	VEB (V) (Mean ± Std)	Fusion (F) (Mean ± Std)	Unknown (Q) (Mean ± Std)	Overall Micro Cohort (Mean ± Std)
Micro Diagnostic Accuracy (%)	90.62 ± 20.28%	94.66 ± 16.05%	98.55 ± 1.85%	96.22 ± 12.88%	99.98 ± 0.04%	90.09 ± 19.54%
Sensitivity / Recall (%)	95.02 ± 13.15%	29.90 ± 32.37%	88.47 ± 14.50%	0.00 ± 0.00%	0.00 ± 0.00%	---
Class Precision (%)	98.84 ± 2.46%	32.12 ± 38.55%	80.53 ± 22.21%	0.00 ± 0.00%	0.00 ± 0.00%	---
Class F1-Score (%)	95.90 ± 10.54%	23.84 ± 27.55%	82.15 ± 15.49%	0.00 ± 0.00%	0.00 ± 0.00%	---

DS2 Individual Patient Breakdown: Macro & Micro Tables

Individual Macro Metrics

Record	Accuracy	Macro Prec.	Macro Rec.	Macro F1
100	99.56%	99.85%	89.90%	93.97%
103	99.90%	49.95%	50.00%	49.98%
105	94.67%	48.73%	53.73%	50.66%
111	99.95%	75.00%	99.98%	83.32%
113	99.83%	85.71%	91.61%	88.43%
117*	39.47%	50.00%	19.75%	28.31%
121	98.23%	44.43%	66.09%	49.70%
123	99.93%	100.00%	99.97%	99.98%
200	97.50%	57.65%	63.92%	59.72%
202	97.75%	56.09%	64.28%	59.55%
210	95.77%	47.70%	47.42%	47.37%
212	100.00%	100.00%	100.00%	100.00%
213	87.56%	40.86%	48.41%	44.28%
214	94.95%	45.38%	44.27%	44.82%
219	92.57%	40.20%	50.98%	43.01%
221	99.55%	100.00%	99.73%	99.86%
222	76.38%	64.96%	79.50%	65.61%
228	95.08%	81.07%	69.60%	74.31%
231	99.87%	66.65%	50.00%	55.54%
232*	22.81%	61.27%	50.29%	18.97%
233	92.39%	49.99%	58.40%	47.42%
234	98.26%	99.42%	68.00%	68.94%

Individual Micro Metrics (With Micro Acc.)

Record	Micro Acc.	Normal (N)	SVEB (S)	VEB (V)	Fusion (F)	Unknown (Q)
100	99.56%	100.00%	69.70%	100.00%	---	---
103	99.90%	100.00%	0.00%	---	---	---
105	94.67%	95.32%	---	65.85%	---	0.00%
111	99.95%	99.95%	---	100.00%	---	---
113	99.83%	99.89%	83.33%	---	---	---
117*	39.47%	39.49%	0.00%	---	---	---
121	98.23%	98.28%	0.00%	100.00%	---	---
123	99.93%	99.93%	---	100.00%	---	---
200	97.50%	98.33%	60.00%	97.33%	0.00%	---
202	97.75%	98.74%	63.64%	94.74%	0.00%	---
210	95.77%	98.02%	9.09%	82.56%	0.00%	---
212	100.00%	100.00%	---	---	---	---
213	87.56%	100.00%	0.00%	93.64%	0.00%	---
214	94.95%	97.00%	---	80.08%	0.00%	0.00%
219	92.57%	92.74%	14.29%	96.88%	0.00%	---
221	99.55%	99.46%	---	100.00%	---	---
222	76.38%	75.75%	83.25%	---	---	---
228	95.08%	99.23%	33.33%	76.24%	---	---
231	99.87%	100.00%	0.00%	50.00%	---	---
232*	22.81%	100.00%	0.58%	---	---	---
233	92.39%	98.25%	57.14%	78.19%	0.00%	---
234	98.26%	100.00%	4.00%	100.00%	---	---