

ECG Arrhythmia Benchmark: Macro and Micro Performance Tables

Comparison of AAMI Independent Test (DS2) and AAMI-kfold Cross-Validation (DS1)

Table 1: Macro Metrics (AAMI DS2 vs. AAMI-kfold DS1)

Evaluation Paradigm / Split	Accuracy (%) (Mean ± Std)	Macro Precision (%) (Mean ± Std)	Macro Recall (%) (Mean ± Std)	Macro F1-Score (%) (Mean ± Std)
AAMI (DS2 Test Benchmark)	93.22 ± 13.88%	67.46 ± 22.16%	68.12 ± 22.35%	64.12 ± 20.61%
AAMI-kfold (DS1 Validation)	91.82 ± 2.09%	45.07 ± 12.92%	44.09 ± 8.60%	43.32 ± 7.72%

Table 2: Micro / Per-Class Metrics across 5 AAMI Labels [N, S, V, F, Q]

Evaluation Paradigm / Split	Normal (N) (Mean ± Std)	SVEB (S) (Mean ± Std)	VEB (V) (Mean ± Std)	Fusion (F) (Mean ± Std)	Unknown (Q) (Mean ± Std)
AAMI Sensitivity / Recall (%) (DS2 Test)	94.69 ± 14.34%	35.68 ± 33.19%	89.66 ± 13.66%	0.00 ± 0.00%	0.00 ± 0.00%
AAMI-kfold Sensitivity / Recall (%) (DS1 Validation)	97.02 ± 1.33%	33.51 ± 28.97%	87.28 ± 8.01%	2.65 ± 2.28%	0.00 ± 0.00%
AAMI F1-Score (%) (DS2 Test)	95.90 ± 10.54%	23.84 ± 27.55%	82.15 ± 15.49%	0.00 ± 0.00%	0.00 ± 0.00%
AAMI-kfold F1-Score (%) (DS1 Validation)	96.70 ± 0.75%	37.06 ± 30.29%	80.56 ± 8.29%	2.27 ± 1.82%	0.00 ± 0.00%