

DS2 Independent Test Benchmark: Patient-Wise Macro Metrics

Evaluated across all 22 independent patient recordings of the AAMI EC57 DS2 partition

Table 1: DS2 Patient-Wise Macro Metrics (All 22 Patients)

Patient Record	Accuracy (%)	Macro Precision (%)	Macro Recall (%)	Macro F1-Score (%)
100	99.56%	99.85%	89.90%	93.97%
103	99.86%	49.95%	49.98%	49.96%
105	94.82%	49.90%	56.18%	52.30%
111	99.95%	75.00%	99.98%	83.32%
113	99.78%	100.00%	66.67%	75.00%
117	33.59%	50.00%	16.81%	25.16%
121	98.33%	45.56%	99.44%	51.87%
123	99.93%	100.00%	99.97%	99.98%
200	97.88%	58.84%	64.17%	60.80%
202	97.84%	56.00%	64.30%	59.41%
210	96.18%	47.77%	48.83%	48.17%
212	100.00%	100.00%	100.00%	100.00%
213	87.65%	40.89%	48.75%	44.43%
214	94.47%	44.27%	44.30%	44.27%
219	92.61%	39.26%	50.99%	42.30%
221	99.63%	100.00%	99.78%	99.89%
222	79.36%	66.04%	76.78%	67.21%
228	95.27%	77.41%	81.13%	77.79%
231	99.87%	66.65%	50.00%	55.54%
232	22.58%	61.23%	50.14%	18.63%
233	92.81%	49.86%	55.25%	47.49%
234	98.22%	99.41%	67.33%	67.67%
Average (Mean)	90.01%	67.18%	67.30%	62.05%
Std Deviation	± 20.23%	± 22.20%	± 22.66%	± 22.68%
Mean ± Std Dev	90.01 ± 20.23%	67.18 ± 22.20%	67.30 ± 22.66%	62.05 ± 22.68%

DS2 Independent Test Benchmark: Micro / Per-Class Sensitivity

Sensitivity (Recall %) across 5 AAMI Heartbeat Classes [N, S, V, F, Q]

Table 2: DS2 Patient-Wise Micro Sensitivity across [N, S, V, F, Q]

Patient Record	Normal (N)	SVEB (S)	VEB (V)	Fusion (F)	Unknown (Q)
100	100.00%	69.70%	100.00%	---	---
103	99.95%	0.00%	---	---	---
105	95.36%	---	73.17%	---	0.00%
111	99.95%	---	100.00%	---	---
113	100.00%	33.33%	---	---	---
117	33.62%	0.00%	---	---	---
121	98.33%	100.00%	100.00%	---	---
123	99.93%	---	100.00%	---	---
200	98.51%	60.00%	98.18%	0.00%	---
202	98.83%	63.64%	94.74%	0.00%	---
210	98.02%	9.09%	88.21%	0.00%	---
212	100.00%	---	---	---	---
213	100.00%	0.00%	95.00%	0.00%	---
214	96.35%	---	80.86%	0.00%	0.00%
219	92.79%	14.29%	96.88%	0.00%	---
221	99.56%	---	100.00%	---	---
222	79.89%	73.68%	---	---	---
228	99.11%	66.67%	77.62%	---	---
231	100.00%	0.00%	50.00%	---	---
232	100.00%	0.29%	---	---	---
233	98.25%	42.86%	79.88%	0.00%	---
234	100.00%	2.00%	100.00%	---	---
Average (Mean)	94.93%	33.47%	89.66%	0.00%	0.00%
Std Deviation	± 14.06%	± 33.26%	± 13.66%	± 0.00%	± 0.00%
Mean ± Std Dev	94.93 ± 14.06%	33.47 ± 33.26%	89.66 ± 13.66%	0.00 ± 0.00%	0.00 ± 0.00%