

## Indications for Use

510(k) Number (if known)  
K212758

Device Name  
Autoplaque 3.0

### Indications for Use (Describe)

Autoplaque is intended to provide an optimized non-invasive application to analyze coronary anatomy and pathology and aid in determining treatment paths from a set of Computed Tomography (CT) Angiographic images.

Autoplaque is a workstation-based post processing application. It is a non-invasive diagnostic reading software intended for use by cardiologists and radiologists as an interactive tool for viewing and analyzing cardiac CT data for determining the presence and extent of coronary plaques and luminal stenoses.

The software is not intended to replace the skill and judgment of a qualified medical practitioner and should only be used by people who have been appropriately trained in the software's functions, capabilities and limitations. Users should be aware that certain views make use of interpolated data. This is data that is created by the software based on the original data set. Interpolated data may give the appearance of healthy tissue in situations where pathology that is near or smaller than the scanning resolution may be present.

Autoplaque must be installed on a suitable commercial computer platform. It is the user's responsibility to ensure the monitor quality and ambient light conditions are consistent with the clinical applications.

Typical users of Autoplaque are trained medical professionals, including but not limited to radiologists, clinicians, technologists, and others.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

#### **\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*