




Certificate of Pre-Amplifier Calibration

Model Number:	AH-2010-025	Input Capacitance:	6.1 pF (at 5 MHz)
Serial Number:	1109	Capacitance Uncertainty:	2.5%
Gain:	20 dB (nominal)	Gain Uncertainty:	0.03 dB
Bandwidth:	25 MHz (nominal)	Output Impedance:	50 Ohms (nominal)
Data File Name:	AH-2010-20-025-1109_240610.txt		



Signature: *Samuel M. Howard* Calibration Date: 10-Jun-2024

1290 Hammerwood Ave, Sunnyvale, CA 94089 T:(408)745-0383 F:(408)745-0956

www.ondacorp.com