


AE 242 Aerospace Measurements Laboratory



Two thermometers are used for measuring temperature. 200 samples are recorded when temperature input is 5 units. From the above two observations, what can be deduced about thermometers?




Two thermometers are used for measuring temperature. 200 samples are recorded when temperature input is 5 units. From the above two observations, what can be deduced about thermometers?



Bias and drift can be estimated by suitable curve fitting (static case).

Accuracy and Precision?



Accuracy: How close observations are with truth.

<u>Precision</u>: How close observations are with respect to each other under same conditions.


Absolute accuracy: When results are compared with truth.


Relative accuracy: When results are compared with observed data.


Accuracy and Precision?


Accuracy: How close observations are with truth.

<u>Precision</u>: How close observations are with respect to each other under same conditions.









Simulation of 100 measurements, true value is 10.

- a) Precise and accurate
- b) Imprecise and accurate
- c) Precise and inaccurate
- d) Imprecise and inaccurate

Accuracy and Precision?

Absolute accuracy: When results are compared with truth.

When the results are presented as absolute value, i.e. deviation from the truth.

Relative accuracy: When results are compared with observed data.

When the results are presented as relative to measured value. i.e. percentage of deviation from the truth divided by truth.

Absolute accuracy error					
= Mesurement - true					

Relative accuracy error					
= 100 * (Mesurement - true)/true					

S. No.	True	Measurement	Error=measure ment - true	Relative accuracy %
1	10	9.5	-0.5	-5.0
2	10	10.1	0.1	1.0
3	10	9.9	-0.1	-1.0
4	10	10.5	0.5	5.0
5	10	9.8	-0.2	-2.0
6	10	10.3	0.3	3.0
7	10	9.3	-0.7	-7.0
8	10	9.7	-0.3	-3.0

What is standard?

Standard forms the basis for consistency. Various measurements can be compared easily using standard. Standard unit of length, weight & time.

International standard: Units of measurement of physical quantities to the highest possible accuracy. Not used for day to day use. Not freely available. Not referenced to other standard.

The meter is the length of the path travelled by light in vacuum during a time interval of 1/299 792 458 of a second.

1 sec is defined as 9,192,631,770 cycles of the atomic resonant frequency of cesium-133.

Bureau International des Poits et Mesures standard http://www.bipm.org/en/about-us/ National Physical Laboratory http://nplindia.org/

What is standard?


Primary standard: These standards are kept at national laboratories. These are calibrated against international standard. Not used frequently. Not freely available.

Secondary standard: These maintained by industrial measurement laboratory for their own calibration of working standard. These are freely available.


Working standard: These are high accuracy devices commercially available and calibrated against secondary or primary standard. These are used for calibrating laboratory equipment or for checking the quality of products etc.

Standard Hierarchy

Hierarchy by metrological level



Hierarchy by geographical location



G.M.S. de Silva, Basic Metrology for ISO 9000 Certification, Butterworth-Heinemann, 2002

What is calibration?

It is a act or result of quantitative comparison between known standard and the output of measuring instrument measuring the same quantity.

Comparing the standard and test instrument simultaneously.

Quantity is varied in ascending and descending order and results are compared.

Standard should be ten times more accurate compared to expected accuracy of the instruments.

Calibration can be done for various environmental conditions.

Types of calibration

<u>Primary Calibration</u>: Device / system calibrated against primary standards. Primary standards are available with national physical laboratories. After primary calibration device is employed as secondary calibration device.

<u>Secondary Calibration</u>: When secondary device is used for calibration then it is called as secondary calibration.

Types of calibration

<u>Direct Calibration</u>: Calibration using a known input source. It is as good as primary calibration. Direct calibrated devices can be used for secondary calibration. E.g. Flow meter is calibrated by accurately measuring the mass of fluid for a given time. In this another accurate flow meter is not required for calibration.


Routine Calibration: Periodically checking the accuracy and proper functioning of an instrument with standard. Very important for reliable functioning of a system.


Types of measurement

Static measurement: Quantities do not change with time or variation is very slow. Input and output can be related using algebraic expressions

Dynamic measurement: When the signal is varying with time. Input and output related by differential equations. More complex relationship as compared to static measurement.

Errors in measurement are introduced due to construction of the sensor, due to specific transducer, due to threshold, due to dead space, due to random behavior etc. Error characteristics may remain constant over time or may change with time. Some errors are introduced only during switching.






Static sensitivity – slope of input-output calibration curve. It could be linear or nonlinear.

Calibration curve also effected by interfering or modifying inputs.

Accuracy is specified as± A percent of readingor ± B percent of full scale




Accuracy is specified as

± A percent of reading

or ± B percent of full scale

For the first case error is zero when input is zero and it is also small for small input, this is not a practical case. There will be always finite error.




Calibration curve also effected by interfering or modifying inputs.

When the instrument is switched on, it may different bias. This bias may have dependency on interfering or modifying inputs.

Example


Fluid pressure exerted on piston results into force. Piston surface is sliding on cylinder surface (friction). Spring is deformed due to application of piston force (could be non-linear). Linear motion of piston is converted to rotary by needle mechanism (mechanical play).

Many of the above factors introduces error.




Measurement Error - Hysteresis

Noncoincidence of loading and unloading curve results in hysteresis. This could be due to material behavior i.e. energy stored during loading is not recovered during unloading or friction force like stiction. It is also observed in ferromagnetic material during cyclic application of magnetic field, resulting into heating.




Backlash



Whenever direction of motion changes, output will be change after some input (amount of backlash). Found in mechanical system. Important to estimate for precise control. It can result to oscillations in the system.


Deadzone




 z_t is deadzone

$$z_2 = z_1 - z_t \text{ for } z_1 \ge z_t$$
 $z_2 = z_1 + z_t \text{ for } z_1 \le z_t$
 $z_2 = 0 \text{ for } -z_t < z_1 < z_t$

Some specific zone in which output will not change even when input is given. It may be intentionally introduced e.g. joystick.




What happens when direction of rotation changes?




Angle can be measured using rotary potentiometer.

What will be output between 270 and 0 degree?




A shaft is rotating (clockwise and anticlockwise) and its true angular position is given in above figure.

What will be the output of sensor connected to shaft which is a) Having backlash b) Having dead zone



Sensor with back lash. Whenever the direction of rotation changes, it will take finite movement before the output changes.



Sensor with dead zone. Whenever the shaft position is in dead zone output will remain constant.



Body temperature measurement





Pressure measurement



What are the common functions in above measurement systems?





$$V = \sqrt{2\Delta p/\rho}$$

Differential pressure

Air Speed

Human needs information which can be comprehended






Differential pressure


$$V = \sqrt{2\Delta P/\rho}$$

Air Speed


Human needs information which can be comprehended




Observer has interest in measured quantity. It may not be directly available, may have intermediate steps.




Primary sensing element: It converts measured quantity (measurand) into some other form. Output of primary sensing element converts into some other convenient variable while preserving the information content of the original signal. Electrical quantity (resistance, capacitance, inductance, voltage and current) is preferred. Easier to process. E.g. Resistance is proportional to temperature. 32




<u>Variable conversion element</u>: This element will convert output of the primary sensing element to another variable suitable for instrumentation. e.g. strain measurement: strain to resistance (strain gage) and resistance to voltage (Wheatstone bridge).




Variable manipulation element: This element will use a predefined mathematical operation. e.g. opamp



<u>Data Transmission element</u>: Primary sensing element and data storage or display may not be co-located. It could be simple wired or wireless.



<u>Data Storage element</u>: Data storage may be analog or digital. Used for analysis and archiving information



<u>Data presentation element</u>: User will typically need information in units relevant to measured medium (temperature, pressure, displacement etc.). Information is reconverted to present in the original units.

In a measurement system it is not necessary that all the elements are present and may not be in same sequence 37

Example - 1

Piston – primary sensing element, variable conversion element (pressure to force)


Piston rod – Data transmission


Spring – Variable conversion element : spring force is converted to linear motion

Linkage: Variable manipulation element: Amplification of the piston movement


Pointer & scale: Data presentation element


All the elements may not be present or may not be in sequence for all the instruments





Example - 2






Wish to measure weight of unknown object?



Pressure gage: Weights are kept in such a way that plunger do not move

Volume will remain constant



Pressure gage: Force is balanced by spring deflection. Plunger movement rotates the indicator

Volume will not remain constant



<u>Deflection type</u>: Measured quantity produces some physical effect that causes some opposing effect in the system and which is balanced. The balance is achieved by deflection. Easier to use.

<u>Null type</u>: Device attempts to maintain deflection at zero by suitable application of an effect opposing that generated by the measured quantity. Measure quantity is inferred from the opposing effect. Requires feedback mechanism to balance. Generally used for precise and accurate measurements.